

PM₁₀/PM_{2.5} comparison exercise in Oslo, Norway

Study in 2015-2016 and 2018

Leif Marsteen and Claudia Hak



NILU report 21/2021 Norwegian Environment Agency M-2108 2021	ISBN: 978-82-425-3059-2 ISSN: 2464-3327	CLASSIFICATION: A – Unclassified (open report)
DATE 30.09.2021	SIGNATURE OF RESPONSIBLE PERSON Ole-Anders Braathen, Deputy director (sign.)	NUMBER OF PAGES 109
TITLE PM10/PM2.5 comparison exercise in Oslo, Norway Study in 2015-2016 and 2018		PROJECT LEADER Leif Marsteen
		NILU PROJECT NO. O-102084
AUTHOR(S) Leif Marsteen and Claudia Hak		QUALITY CONTROLLER Kjersti K. Tørnkvist
REPORT PREPARED FOR Norwegian Environment Agency Contact person: Line Merete Karlsøen		CONTRACT REF. 17088083
ABSTRACT <p>The purpose of the comparison was to test for equivalence and establish calibration functions for automatic PM-analysers commonly used in Norway. The reference laboratory performed a field test at three different locations in Oslo during summer and winter conditions in the periods September 2015 to July 2016 and February to March 2018. Participating analysers were Palas Fidas 200, Grimm EDM 180, TEI TEOM 1405 DF, TEI FH 62 I-R, and R&P TEOM 1400AB.</p> <p>The report proposes a system to carry out ongoing verification of equivalence in the Norwegian monitoring network and how to calibrate analyser data.</p>		
NORWEGIAN TITLE Rapport fra sammenligning av måleinstrumenter for PM ₁₀ /PM _{2.5} i Norge, gjennomført i Oslo 2015-2016 og 2018		
KEYWORDS <div style="display: flex; justify-content: space-around; text-align: center;"> Comparison PM Calibration </div>		
ABSTRACT (in Norwegian) <p>Formålet med sammenligningen var å ekvivalensteste og etablere kalibreringsfaktorer for de vanligste automatiske PM-målere som er i bruk i Norge. For å etablere faktorene utførte Referanselaboratoriet en feltstudie på tre steder i Oslo under sommer- og vinterforhold i periodene september 2015 til juli 2016 og februar til mars 2018. Måleinstrumentene som deltok var Palas Fidas 200, Grimm EDM 180, TEI TEOM 1405 DF, TEI FH 62 I-R, og R&P TEOM 1400AB.</p> <p>Rapporten beskriver et mulig system for kontinuerlig verifikasjon av kalibreringsfaktorene i de norske målenettene og hvordan analysedata skal kalibreres.</p>		
PUBLICATION TYPE: Digital document (pdf)		COVER PICTURE: Source: NILU

© NILU – Norwegian Institute for Air Research

NILU's ISO Certifications: NS-EN ISO 9001 and NS-EN ISO 14001. NILU's Accreditation: NS-EN ISO/IEC 17025.

Preface

This report was written by NILU in the capacity of the Norwegian Reference Laboratory for Air Quality. The Reference Laboratory is financed by the Norwegian Environment Agency.

The reference laboratory wishes to thank The Norwegian Public Roads Administration, Industriell måleteknikk/ GRIMM Aerosol Technik and Tillquist/ Palas for providing instruments to the comparison, and the reference laboratories of Sweden, UK, Denmark and Austria for sharing their experiences. Also thanks to Jan Henrik Wasseng, Reiar Kravik, Dorothea Schulze and Jøran Solnes Skaar, all from NILU, who were responsible for the field work.

Contents

Preface	2
Contents	3
Summary	5
Summary of results from PM ₁₀ comparison	5
Summary of results from PM _{2.5} comparison	6
Summary for possible system to correct PM analyser data	6
1 Introduction.....	7
2 Methods	8
2.1 Reference instruments.....	8
2.2 Candidate instruments.....	8
2.3 Filters and filter conditioning.....	10
2.4 Instruments characterising the site	10
2.5 Data treatment	10
3 Measurement design.....	11
3.1 Measurement locations and measurement periods	11
3.2 Site specific reference instruments	13
3.3 Site specific candidate instruments	14
3.4 Location of inlets on the shelter	15
4 Results.....	17
4.1 Results from instruments characterising the site	17
4.2 Results from PM ₁₀ comparison	20
4.2.1 Results from Leckel SEQ47/50 PM ₁₀ reference samplers	22
4.2.2 Comparison of Palas Fidas 200 PM ₁₀ candidate.....	24
4.2.3 Comparison of Grimm EDM 180 PM ₁₀ candidate	26
4.2.4 Comparison of TEOM 1405DF PM ₁₀ candidate	28
4.2.5 Comparison of TEI FH62 IR PM ₁₀ candidate with SA 246b (USEPA) impactor	30
4.2.6 Comparison of TEI FH62 IR PM ₁₀ candidate with EN12341 impactor.....	32
4.3 Results from PM _{2.5} comparison.....	34
4.3.1 Results from Leckel SEQ47/50 PM _{2.5} reference samplers.....	37
4.3.2 Comparison of Fidas 200 PM _{2.5} candidate	38
4.3.3 Comparison of Grimm 180 PM _{2.5} candidate	40
4.3.4 Comparison of TEOM 1405 DF PM _{2.5} candidate	42
4.3.5 Comparison of TEI FH62 IR PM _{2.5} candidate with EN12341 impactor	44
4.3.6 Comparison of TEOM 1400 AB PM _{2.5} candidate	46
5 Discussion and conclusion, PM comparison.....	48
5.1 PM ₁₀ candidates.....	48
5.2 PM _{2.5} candidates	49
6 Possible system for Norway to calibrate AMS PM data.....	50
6.1 Background	50
6.1.1 Equivalence testing	51
6.1.2 Ongoing verification of equivalence	51
6.1.3 AMS in Norwegian air quality monitoring networks	52

6.2	Status in Norway and some selected countries.....	56
6.3	Suggested system for Norway	61
6.3.1	Selection of sites	61
6.3.2	Calibration method	67
6.3.3	Organisation of ongoing verification in Norway	68
6.3.4	Ongoing verification after the first year	70
	References	71
	Appendix A PM10 performance characteristics tables	72
	Appendix B PM_{2.5} performance characteristics tables	88
	Appendix C Geographic distribution of analyser types over Norway	104

PM₁₀/PM_{2.5} comparison exercise in Oslo, Norway

Study in 2015-2016 and 2018

Summary

The purpose of the comparison was to test for equivalence and establish calibration functions for automatic PM monitors commonly used in Norway.

To establish the calibration functions and estimate the uncertainty of the measurement methods, the national reference laboratory performed a field test at three different locations in Oslo during summer and winter conditions in the periods September 2015 to July 2016 and a second comparison during winter conditions in February 2018 to March 2018. Each field test period lasted approximately 6 weeks. The comparison was performed according to the "Guide to the demonstration of equivalence", [1].

Four Leckel SEQ47/50 sequential samplers were run in parallel as reference instruments, two sampling PM₁₀ and two sampling PM_{2.5}. A total of 7 candidate methods (CM) were tested, out of which 5 methods were tested with 2 instruments in parallel. Meteorology (wind, temperature, relative humidity) was measured at all sites.

Based on the results from the PM-comparison and on the need of performing ongoing verification exercises for PM analysers, a possible system for Norway to introduce ongoing verification of equivalence is suggested in the second part of this report (Section 6)

Summary of results from PM₁₀ comparison

The comparison included 2 reference samplers and 7 candidates. Table 1 lists calibration functions to be applied to measurements of PM₁₀. The expanded relative uncertainty is the uncertainty of the measured values after calibration. When pairs of candidates were tested in parallel, the listed factors originate from one of the paired candidates, see Table 9 for details.

Table 1. PM₁₀ calibration functions.

Candidate	Calibration function		Expanded rel. uncertainty	Comments
	Slope	Intercept		
Palas Fidas 200	0.898	2.24	16.4 %	
Grimm EDM 180	1.026	1.52	26.2 %	
TEI TEOM 1405 DF	1.126	0.00	10.0 %	Roadside winter and urban background
TEI FH 62 I-R	0.819	0.00	9.5 %	Roadside with SA 246b (USEPA) impactor
TEI FH 62 I-R	0.961	0.00	6.8 %	Urban background with EN12341 impactor

Measured data is calibrated according to: $\text{Calibrated value} = \text{Slope} * \text{Measured value} + \text{Intercept}$

All paired candidates (Fidas 200 and EDM 180) had satisfactory between-candidates uncertainties. EDM 180 and FH62IR (with USEPA inlet) failed to pass the comparability test for PM₁₀ with expanded relative uncertainties higher than 25 %. All candidates had significant deviation in the slope, while Fidas 200 and EDM 180 were the only candidates with significant deviation in the offset. After calibration, all candidates, except EDM 180, passed the test for expanded relative uncertainty below 25 %.

Both EDM 180-candidates failed to pass the expanded relative uncertainty test after calibration for PM₁₀ in the first comparison. They failed because they failed to pass the uncertainty test at the Hjortneskaia roadside site during winter season. This site is close to the harbour and sea salt may have

had an effect on the measurements. A second comparison at the roadside site Smestad during winter season was organised to rule out the possible effect of sea salt. Both candidates passed the uncertainty test at Smestad. The expanded relative uncertainty of all data decreased after the second comparison, but candidate Grimm1 still failed to pass the uncertainty test, also after calibration.

Summary of results from PM_{2.5} comparison

The comparison included 2 reference samplers and 9 candidates. Table 2 lists calibration functions to be applied to measurements of PM_{2.5}. The expanded relative uncertainty is the uncertainty of the measured values after calibration. When 2 candidates were tested in parallel, the listed factors originate from one of the paired candidates, see Table 10 for details.

Table 2. PM_{2.5} calibration functions.

Candidate	Calibration functions		Expanded rel. uncertainty	Comments
	Slope	Intercept		
Palas Fidas 200	1.009	0.00	19.4 %	
Grimm EDM 180	0.903	0.00	17.3 %	
TEI TEOM 1405 DF	1.045	0.00	10.6 %	Roadside winter and urban background
TEI FH 62 I-R	0.837	0.00	22.4 %	With EN12341 impactor
R&P TEOM 1400AB	1.451	-3.86	12.2 %	

Measured data is calibrated according to: $\text{Calibrated value} = \text{Slope} * \text{Measured value} + \text{Intercept}$

All paired candidates had satisfactory between-candidates uncertainties except for FH62IR where the second candidate (TEI2) was unstable throughout the whole measurement campaign and all its results were rejected. All candidates, except Fidas 200 and TEOM 1405 DF, failed to pass the comparability test for PM_{2.5} with expanded relative uncertainties higher than 25 %. All candidates except Fidas 200 had significant deviation in the slope, while only TEOM 1400 AB had significant deviation in the offset. After calibration all candidates passed the test for expanded relative uncertainty below 25 %.

Both Fidas 200-candidates passed the expanded relative uncertainty test after calibration in the first comparison, but they failed to pass the uncertainty test for PM_{2.5} at the Hjortneskaia roadside site during winter season. This site is close to the harbour and sea salt may have had an effect on the measurements. A second comparison at the roadside site Smestad during winter season was organised to rule out the possible effect of sea salt. Both candidates passed the uncertainty test at Smestad. The expanded relative uncertainty of all data after calibration did not change.

Summary of possible system to correct PM analyser data

Up to now, only PM₁₀-measurement data from TEOM 1400 and TEI FH 62-IR has been calibrated in Norway. As the comparison exercise carried out in Oslo in the period 2015-2018 shows, discrepancies between the results of different PM-analyser types and the reference method, the need of calibrating PM analysers has been identified.

A possible system to carry out ongoing verification of equivalence in the Norwegian monitoring network and to calibrate analyser data is suggested. The calibration is based on the system which has been successfully applied in Austria for over a decade and involves first calibration of incoming data using a calibration formula established in an earlier comparison exercise (forward calibration) and

finally calibrating the raw data at the end of year using the new calibration formula obtained from the full year (backward calibration).

Section 6 describes how a future system in Norway for ongoing verification of PM-analysers may be set up. This involves both the decision of the number of sites to be tested for ongoing verification (depending on the outcome of the most recent PM-comparison), the selection of these stations, the criteria of how to select stations and how stations in Norway may be grouped. It is also suggested how to calibrate the raw data and how the system for ongoing verification may be organised in Norway.

1 Introduction

According to the European “Directive 2008/50/EC on ambient air quality and cleaner air for Europe” (CAFÉ directive, [2]), the member states are obliged to monitor the concentrations of suspended particulates in ambient air. The directive defines the reference method for sampling PM₁₀ and PM_{2.5} to be the gravimetric method, that is sampling on filter and post-weighing in the laboratory (Annex VI, Sections A, 4 and A, 5). The directive refers to the European standard EN 12341:2014 “Standard gravimetric measurement method for the determination of the PM₁₀ or PM_{2,5} mass concentration of suspended particulate matter”, [3], where the measurement method is described.

All network operators in Norway use automatic PM analysers which do the analysis on-site. Data is read at a relatively high frequency by a data logger and hourly averages are calculated and stored for transfer to the home system. The directive allows the use of any measurement method as long as one can show that the alternative method gives the same results, after calibration if necessary, as the reference method. If this can be shown, the method is called an equivalent method. Most commercially available PM analysers have been tested and declared equivalent methods by specialised test laboratories such as TÜV in Germany. The tests are performed according to chapter 9 in the “Guide to demonstration of equivalence, GDE”, [1], or the European standard EN 16450:2017 “Automated measuring systems for the measurement of the concentration of particulate matter (PM₁₀; PM_{2,5})”, [4]. Both documents give detailed information on test protocol and data analysis.

Equivalence of a candidate instrument is demonstrated by measuring ambient air in parallel with a reference sampler during various meteorological and site conditions that are representative of the future use of the candidate. The relationship between a candidate instrument and the reference method is analysed using orthogonal regression. The slope and intercept of the calibration function must stay within certain limits, [1]. In addition, the expanded relative uncertainty shall be less than 25 %, [1] and [2]. If either criterion is not fulfilled the candidate data may be calibrated against the reference sampler data and the expanded relative uncertainty recalculated. If the new uncertainty is below 25 % the candidate is accepted as an equivalence method provided the measured values are always calibrated before reported.

In order to determine which PM₁₀/PM_{2.5} monitors are most suitable for the Nordic conditions, NILU performed a field test at three different locations in the city of Oslo, Norway, during summer and winter conditions. Each field test period lasted approx. 6 weeks. To cover the different ambient conditions as described in EN12341, the following types of test sites were selected, indicating also the major contributing source:

- Roadside in the city (traffic, exhaust and road dust)
- Urban background in the city (all sources)

The candidate instruments in this comparison exercise included the most commonly used automatic PM monitors in Norway. The measurement methods were TEOM, TEOM FDMS, light scattering and β -gauge. The comparison was performed according to “The Guide to demonstration of equivalence”

with deviations as noted in the text. Similar comparisons have been performed in other Nordic countries, [5], [6]. In addition to documenting the candidates' fulfilment of the data quality objectives, the report documents how the candidates' response varies by site and season. However, since the comparison was not repeated using the same setup at the same sites and season, but another year, it is not possible to conclude that the findings are typical for the given sites and seasons.

2 Methods

2.1 Reference instruments

The reference sampler was the Leckel SEQ47/50 sequential sampler. The sampler has an automatic filter changer containing 15 filters including one blank filter. A cooling unit kept the filters at low temperature to avoid evaporation of semi volatile components. Each filter was exposed for 24 hours and the filter stack was replaced every 14 days. The sampler reported sampled volume at operational conditions.

Table 3 lists the reference instruments used in the comparison. The reference instruments comply with the requirements of EN12341:2014.

Table 3. Reference instruments. "n" in Code indicates instrument number 1, 2, ...

Instrument	Make	Method	Design flow	PM _x	Owner	Code
SEQ47/50 sampler	Sven Leckel, Ingenieurbüro	Gravimetric	2.3 m ³ /h	PM ₁₀	NILU	RMn
SEQ47/50 sampler	Sven Leckel, Ingenieurbüro	Gravimetric	2.3 m ³ /h	PM _{2.5}	NILU	RMn

2.2 Candidate instruments

All candidate instruments were automated measurement systems (AMS).

Table 4 lists the candidate instruments used in the comparison. The guide to demonstration of equivalence (GDE) requires two instruments of each type. This was accomplished for most candidate instruments, see Table 8.

Table 4. Candidate instruments. "n" in Code indicates instrument number 1, 2, ..., a, ...

Instrument	Make	Method	Design flow	PM _x	Sample inlet	Instrument provider	Code
TEOM 1405 DF	Thermo Fisher Sc.	Micro balance	1 m ³ /h	PM ₁₀ PM _{2.5}	Sheath air, ambient temp.	The Norwegian Public Roads Administration	TEOMDF
TEOM 1400 AB	R&P	Micro balance	1 m ³ /h	PM _{2.5}	Heated 50°C	NILU	TEOMn
EDM 180	GRIMM Aerosol Teknik	Light scatter	0.072 m ³ /h	PM ₁₀ PM _{2.5}	Sheath air, nafion drier	Industriell måleteknikk/ GRIMM A. T.	GRIMMn
Fidas 200	Palas	Light scatter	0.288 m ³ /h	PM ₁₀ PM _{2.5}	IADS heating	Tillquist/ Palas	FIDASn
FH62 I-R	Thermo Fisher Sc.	β-gauge	1 m ³ /h	PM _{2.5}	Heated 30°C	NILU	TEIn
FH62 I-R	Thermo Fisher Sc.	β-gauge	1 m ³ /h	PM ₁₀	Heated 15°C above amb. temp.	NILU	TEIn

In the TEOM 1400 AB instrument (Tapered Element Oscillating Microbalance), a small filter is located on top of a hollow pin. The pin is forced to oscillate at its resonance frequency. As mass accumulates on the filter, the resonance frequency decreases. The resonance frequency is inversely proportional to the accumulated mass on the filter. TEOM 1400 AB measures either PM₁₀ or PM_{2.5} depending on the size selective inlet.

TEOM 1405 DF is a TEOM with an FDMS (Filter Dynamics Measurement System) unit. TEOM 1405 DF measures the loss of volatiles on the sampling filter by switching between base (normal) operation and reference (filtered) operation every 6 minutes. During the base period, suspended particles and volatiles are sampled on the filter. During the reference period, all suspended particles and volatiles are removed from the air stream before it passes through the filter and the TEOM should measure zero mass. But some of the volatiles captured during the previous base period may evaporate from the filter during the reference period. This is measured as a loss of mass and added to the mass measured collected during the previous base period. TEOM 1405 DF is a dichotomous analyser, meaning it measures both PM_{2.5} and PM_c (coarse fraction) simultaneously using two TEOM measurement systems in parallel. The inlet air passes through a PM₁₀ size selective inlet followed by a virtual impactor which splits the sampled air into PM_{2.5} and PM_c fractions. PM₁₀ is calculated by adding PM_{2.5} to PM_c (PM₁₀ = PM_c + PM_{2.5}).

The light scattering measurement method is based on a laser illuminating a chamber in which single particles enter. Each particle will scatter the light from the laser. The scattered light is detected at an angle of 90°. The number of light pulses registered equals the number of particles in the air passing through the chamber. The intensity of the scattered light is a measure of the particle size / diameter. Fidas utilises a high intensity white light LED as light source.

In the β-gauge instrument, the exposed filter is bombarded with β-particles and the β-particles penetrating the filter are measured in an ionisation chamber. As mass accumulates on the filter during sampling, fewer β-particles penetrate the filter and are measured. A reference value is continuously monitored by a second ionisation chamber to compensate for temperature and pressure fluctuations. The number of β-particles penetrating the filter is proportional to the particle mass on the filter. The filter is actually a filter tape and the tape is advanced every 24 hours exposing a new fresh area. The FH62-IR monitor measures either PM₁₀ or PM_{2.5} depending on the size selective inlet.

The PM monitors measured the mass concentration continuously and the values were recorded in a data logger as 1 hour averages. The 1 hour averages were later aggregated to 24 hour averages for comparison with the reference method. All candidate monitors reported concentrations at operational conditions.

2.3 Filters and filter conditioning

The reference sampling filters were 47 mm, Pall Zefluor 2 µm. The filters were conditioned 48 hours at 21 °C and 47 % relative humidity before weighing in a clean-room both before and after exposure.

2.4 Instruments characterising the site

Meteorological conditions at the sites were monitored by a Vaisala WXT520 unit. Wind direction, wind speed, temperature, relative humidity and barometric pressure were monitored and recorded as hourly averages. Table 5 lists instruments characterising the site.

Table 5. Instruments characterising the site.

Instrument	Make	Parameters	Owner
WXT520	Vaisala	Wind direction, wind speed, temperature, relative humidity, barometric pressure, indoor temperature	NILU
TM-RS232	Papouch	Temperature, measured indoor	NILU
TM-RS232	Papouch	Temperature, measured in pump room	NILU

2.5 Data treatment

Data from the candidate instruments were logged every 10 seconds and aggregated to 1-hour averages in the data logger. During post processing, data evaluation and discrimination was done on the 1-hour averages. The quality controlled 1-hour averages were aggregated to 24-hour averages for comparison with the reference method. The 24-hour average runs from 00:00 to 23:59.

The reference instruments were programmed to cover a sampling period equal to the 24-hour averaging period of the candidate instruments.

Both candidate and reference instruments reported concentrations and volumes at ambient conditions. All results were used "as measured" without applying any correction. TEOM 1400 AB corrects its measured values internally according to the formula $\text{ReportedValue} = 1.03 * \text{MeasuredValue} + 3$. The corrected values are used in this report.

When more than 2 hours of monitor data were missing, the 24-hour average of that day was invalidated. Values from periods of instrument failures were removed from the data sets. Candidates of same type were evaluated separately. No outliers were removed from the data. Data has been removed due to technical reasons only.

The reference values were based on averages of the two reference samplers. The 24-hour reference value was based on one single instrument when the other instrument did not produce a valid 24 hour measurement. The reference method averages of all data in the tables in App. A are unique to each candidate because the average is calculated only for pairs of reference and candidate values.

The test protocol requires that at least 40 of the total number of daily averages of PM₁₀ and PM_{2.5} are above 30 µg/m³ and 18 µg/m³ respectively. The requirement was often not fulfilled indicating that the PM levels were low and not necessarily that the measurement method did not perform satisfactorily.

Fidas 200 and EDM 180 candidate instruments' characteristics were calculated based on values from both comparisons (Hjortneskaia/Sofienbergparken and Smestad). The paired analysers (eg. the two Fidas 200) used in the second comparison were not the same as the ones used in the first comparison. To avoid bias due to how the analysers in the two comparisons were combined, the paired values of the second comparison were averaged and added to both data series in the first comparison. The between-analyser uncertainty was calculated based on data from the first comparison only.

An inspection of the hourly averages showed that TEOM 1405 DF reported 1.7 % of all hourly PM₁₀ averages and 5.8 % of all hourly PM_{2.5} averages below zero (down to -5 ug/m³). The negative values were included in the daily averages.

The Excel workbook "Equivalence Tool, version 10" (Orthogonal regression and equivalence test utility), [7], developed by RIVM and used by AQUILA members, was used to test the comparability between each candidate and the reference method and to calculate calibration functions for the candidate methods. The relationship between a candidate and the reference method is:

$$y_i = a + b * x_i$$

where a and b are intercept and slope respectively. The criterion for accepting the calculated slope and intercept in the comparability test depends on the calculated uncertainty of the slope and intercept respectively. This punishes good candidates because the smaller the uncertainties in slope and intercept are, the smaller deviation from 1 and 0.0 is allowed in the slope and intercept respectively. In this report the criteria for slope (b) and intercept (a) are set to:

$$0.98 < b < 1.02$$

$$-1.00 < a < 1.00$$

The requirement for comparability is an expanded relative uncertainty of the candidate method below 25 %, calculated at the limit value of 50 ug/m³ for PM₁₀ and 30 ug/m³ for PM_{2.5}, [3]. Detailed results for each candidate are listed in Appendix A and B.

3 Measurement design

3.1 Measurement locations and measurement periods

The measurement sites were selected to represent both common and extreme situations in accordance with the European standard EN12341. Two sites were located in Oslo; one roadside site at E18, the main south-western highway connecting the city with residential areas outside Oslo, and one site in urban background surroundings.

The roadside site at Hjortneskaia is exposed to heavy traffic from the highway that carries about 75 000 vehicles per day. The surrounding area is open with residential buildings to the north. The harbour with a large marina is next to the southern side of the station.

The background site at Sofienberg is in a green park with trees. The surrounding area is mainly residential and consists to a high degree of four to five floor buildings built between 1850 and 1900 for workers. Heating in these buildings is by electricity, but many probably also use stoves during cold conditions. Many buildings in the area have been renovated in the past 10 – 20 years and some have been taken down and replaced by modern residential houses. The park is surrounded by roads carrying local traffic on all four sides.

The second roadside site at Smestad is at Ring 3, a western highway in the north-west of Oslo. The site is exposed to heavy traffic from the highway that carries about 46 000 vehicles per day. The surrounding area consists mainly of residential buildings and is away from the seaside.

Figure 1 shows the measurement site locations in Oslo.



Figure 1. Measurement site locations (<https://norgeskart.no/>)

Table 6 lists the measurement sites and measurement periods.

Table 6. Measurement sites and periods.

Site	Season	Start	End	Code
Roadside, Hjortneskaia	Autumn	2015.09.16	2015.10.28	RSA
	Winter	2015.10.29	2015.12.20	RSW
Urban background, Sofienbergparken	Winter	2016.01.26	2016.03.14	UBW
	Summer	2016.05.10	2016.07.11	UBS
Roadside, Smestad	Winter	2018.02.02	2018.03.22	RSW2

Due to delays in the availability of instruments, the summer season at the roadside site Hjortneskaia had to be shifted to autumn, and the winter season at the same site started just after end of the autumn measurement period.

The test protocol requires that in each comparison period a minimum of 40 valid daily data pairs (a data pair representing at least one result from the reference method and one from the candidate method from the same 24-hour period) shall be obtained, [1].

The instruments ran for approx. six weeks at each site collecting ideally 42 samples from each reference sampler and candidate instrument. With a total of 4 periods this totals 168 daily averages from each instrument. Fidas 200 and EDM 180 participated in the second comparison at Smestad, totalling their number of samples to 210 daily averages.

The instruments were located in two rows on top of a shelter. The roof top was approx. 2.5 m above ground. Sampling inlets were located at between 85 and 110 cm above the roof. The meteorological

tower was fixed to one corner of the shelter and the sensors were located in the tower at approx. 7.5 m above ground. The same shelter was used at all sites.

3.2 Site specific reference instruments

Table 7 lists the reference instruments and their locations.

Table 7. Characterisation of reference instruments and their locations. Number of valid samples are shown in table. Periods with less than the required 40 samples are marked in red.

Instrument	Ser. No.	Head	PM _x	Location					Comment
				RSA	RSW	UBW	UBS	RSW2	
RM1	10/0060	EN12341	PM _{2.5}	15		49	47		Instrument failure at RSA and RSW
RM3	09/0052	EN12341	PM _{2.5}	40	52	49	47		
RM2	11/0051	EN12341	PM ₁₀	42	52	49	47		
RM4	19/0061	EN12341	PM ₁₀	42	52	10	47		Instrument failure at UBW
RM3-5	18/0055	EN12341	PM _{2.5}					45	
RM4-6	18/0054	EN12341	PM ₁₀					48	

Reference sampler RM4 broke down during startup at the urban background winter comparison (UBW). The instrument was brought to the lab for repair and later deployed again.

Reference sampler RM1 was in operation for 15 days from start at the roadside site during the autumn season before it was removed for service. The sampler was repaired and deployed again at the start of the urban background winter comparison.

3.3 Site specific candidate instruments

Table 8 lists the candidate instruments and their locations.

Table 8. Characterisation of candidate instruments and their locations. Number of valid 24 hour averages are shown in table. Periods with less than the required 40 averages are marked in red. Difference in number of PM₁₀ and PM_{2.5} averages on same candidate is due to availability of reference samples.

Instrument	Ser. No.	Sampling inlet	PM _x	Location					Comment
				RSA	RSW	UBW	UBS	RSW2	
TEOMDF	1405 A22 8151404	SA 246b louvered	PM ₁₀ PM _{2.5}		19 18	47 47	47 46		RSA: Tube mix-up from factory
TEOM1	21741	SA 246b	PM _{2.5}	39	49	49	42		30°C inlet heating
TEOM2	29608	SA 246b	PM _{2.5}	40	55	43	45		30°C inlet heating
Grimm1	18A07083	TSP	PM ₁₀ PM _{2.5}	42 42	52 52	49 49	47 46		No inlet heating Nafion drier
Grimm2	18A07031	TSP	PM ₁₀ PM _{2.5}	42 42	52 52				Instrument failure at UBW
Grimm2a	18A09102	TSP	PM ₁₀ PM _{2.5}			42 42	47 46		Replaced Grimm2
Fidas1	6232	Sigma-2	PM ₁₀ PM _{2.5}	42 42	46 46	49 49	34 33		
Fidas2	6416	Sigma-2	PM ₁₀ PM _{2.5}	42 42	50 51	50 49	45 43		
TEI1	600	EN12341	PM _{2.5}	42	52	49	37		30°C inlet heating
TEI2	1185	EN12341	PM _{2.5}	41	48	49	33		Inlet heated 10 - 15°C above amb.
TEI3	0716	SA 246b	PM ₁₀	42	52				Inlet heated 10 - 15°C above amb.
TEI3a	0716	EN12341	PM ₁₀			49	47		TEI3 w/ different impactor
Fidas1-3	6766	Sigma-2	PM ₁₀ PM _{2.5}					47 45	
Fidas2-4	7166	Sigma-2	PM ₁₀ PM _{2.5}					47 45	
Grimm1-3	18A07083	TSP	PM ₁₀ PM _{2.5}					44 41	No inlet heating Nafion drier
Grimm2-4	18A09062	TSP	PM ₁₀ PM _{2.5}					44 41	No inlet heating Nafion drier

Only one TEOM 1405 DF (TEOMDF) participated in the comparison. The instrument was fresh from the factory. During service after the first measurement period at the roadside site (autumn season) it was discovered that the tubes leading sampled air to the coolers were interchanged at the factory so that the PM₁₀ fraction was measured by the PM_{2.5} unit and vice versa. The data was rejected.

The Grimm2 candidate broke down during startup at the urban background site, winter season, and was replaced by Grimm 2a for the remaining comparison. In the data treatment they are both called Grimm2.

TEI3 and TEI3a are the same instrument using different impactors.

Due to limited availability of instruments, TEI1 and TEI2 (both PM_{2.5}) participated with different inlet heating of 20°C and ambient temperature respectively. However, the TEI2 analyser was unstable throughout the whole measurement campaign and all TEI2 results were rejected.

3.4 Location of inlets on the shelter

Figure 2 shows the instrument configuration on top of the shelter at the Hjortneskaia roadside measurement site.




Figure 2. Instrument configuration on shelter roof (not to scale) at Hjortneskaia roadside site.

The road was next to the lower row of instruments in Figure 2. The direction of the traffic in the two lanes closest to the shelter was from right to left.

Figure 3 shows the instrument configuration on top of the shelter at the Sofienbergparken urban background measurement site. Grimm2 was replaced by Grimm2a due to technical problems. TEI3 using USEPA inlet was replaced by TEI3a (same instrument) using EN12341 inlet.




Figure 3. Instrument configuration on shelter roof (not to scale) at the Sofienbergparken urban background site.

Figure 4 shows the instrument configuration on top of the shelter at the Smestad roadside measurement site.




Figure 4. Instrument configuration on shelter roof top (not to scale) at the Smestad roadside site.

The road was next to the lower row of instruments in Figure 4. The direction of the traffic in the two lanes closest to the shelter was from right to left.

4 Results

The first comparison covered autumn and winter seasons at Hjortneskaia roadside site and winter and summer season at Sofienbergparken urban background site. The second comparison covered winter season at Smestad roadside site.

TEOM 1405 DF was not in operation during the autumn season and most of the winter season at the roadside site. TEI FH62IR (PM₁₀) was in operation only at the roadside site. TEI FH62IRa (PM₁₀) was in operation only at the urban background site. The results of these analysers are representative only for the specific sub parts of the comparison.

4.1 Results from instruments characterising the site

Wind speed, wind direction, temperature, relative humidity and precipitation at the individual sites were measured by a Vaisala model WXT520 Weather Transmitter. The wind sensor is based on an array of three ultrasonic transducers. The precipitation sensor is based on a piezoelectrical sensor detecting the impact of individual raindrops on a steel plate. The precipitation sensor is not considered an accurate device but more indicative.

The highest wind speeds, as measured at the sites, see Figure 5, occurred during winter seasons at the Hjortneskaia roadside site (RSW) and Smestad roadside site (RSW). The lowest wind speeds occurred during summer season at the Sofienbergparken urban background site (UBS).

Figure 6 shows 24 hour average temperatures of all sites. The average monthly temperature during the first measurement campaign was comparable to the average monthly normal measured at the official meteorological station at Blindern¹ except for December at the roadside site (RSW) where the average temperature was 2 °C which is 5 °C higher than the normal.

October was very dry at the roadside site (RSA) with only 9 mm monthly precipitation compared to the 84 mm monthly normal at Blindern. May was more wet than normal at the urban background site (UBS) with 74 mm monthly precipitation compared to the 53 mm normal at Blindern. There was little snow during the first measurement campaign except for March where the monthly average measured at Blindern was 5 cm.

The primary wind direction, see Figure 7, during the comparison at the roadside site (RSA and RSW) was from North-East along the south-western highway into Oslo and residential areas. The wind speed was low, mostly below 2 m/s. Some occurrences of higher wind speeds, above 6 m/s, from South-West coming from the fjord and a large ferry terminal were observed during winter season. The primary wind direction during the comparison at the urban background site (UBW and UBS) was from South coming from the residential areas, the city centre and the railway station. The wind speed was low, mostly below 2 m/s but with some occurrences up to 4 m/s also from the South.

In the second measurement campaign at the roadside site Smestad (RSW2) during winter season, the average monthly temperature was similar to the normal temperature at Blindern. The average temperature was around -3 °C which is 5 °C lower than the average temperature at Hjortneskaia roadside site during similar season. There was more precipitation than normal in February and less in March. The average snow depth in both February and March was close to 50 cm and much more than the snow depth at the Hjortneskaia roadside site. The primary wind direction during the comparison was from North-East along the northern highway into Oslo and from residential areas.

¹ Blindern is situated at 90 m above sea level and about 3.5 km from all sites.




Figure 5. Wind speed, 24 hour averages, measured at all sites and seasons as indicated by vertical bars.




Figure 6. Ambient temperature, 24 hour averages, measured at all sites and seasons as indicated by vertical bars.




Figure 7. Wind roses, by site and season.




Figure 8. Relative humidity, 24 hour averages, measured at all sites and seasons as indicated by vertical bars.




Figure 9. Barometric pressure, 24 hour averages, measured at all sites and seasons as indicated by vertical bars.

4.2 Results from PM₁₀ comparison

The performance characteristics of the PM₁₀ candidates based on data from both comparisons are summarized in Table 9. Numbers in red indicate significant deviations from performance criteria. Numbers in blue indicate a deviation in slope and/or intercept that makes calibration necessary. All paired candidates (Fidas 200 and EDM 180) had satisfactory between-candidates uncertainties. EDM 180 and FH62IR (with USEPA inlet) failed to pass the comparability test for PM₁₀ with expanded relative uncertainties higher than 25 %. All candidates had significant deviation in the slope, while Fidas 200 and EDM 180 were the only candidates with significant deviation in the offset. After calibration, all candidates, except EDM 180, passed the test for expanded relative uncertainty below 25 %. The expanded relative uncertainty after calibration was also calculated using the calibration function with the intercept set to 0.0 (zero). The expanded relative uncertainties of TEOM 1405 DF and FH62IRs remained unchanged due to their intercepts being close to zero, indicating that the calibration function may be applied with the intercept set to zero.

Table 9. Summary of performance characteristics of the PM₁₀ candidates. Numbers in red indicate significant deviations from performance criteria. Numbers in blue indicate a deviation in slope and/or intercept that makes calibration necessary. "x" indicates there was only one candidate participating and between CM uncertainty could not be calculated.

Test PM ₁₀	Criteria	Fidas200	EDM180	TEOM DF	FH62IR	FH62IRa
Between CM uncertainty	$u_{bs,CM} < 2.5 \mu\text{g}/\text{m}^3$	0.38	0.94	x	x	X
Comparability						
Number of values		218	234	113	94	96
Data capture	> 90 %	91.2 %	97.9 %	59.2 %	49.2 %	50.3 %
Slope, b	$0.98 < b < 1.02$	1.1141	0.9743	0.8884	1.2212	1.0407
Intercept, a	$-1 < a < 1$	-2.4923	-1.4799	0.1670	-0.6774	0.2092
Expanded rel. uncertainty	25 %	21.8 %	27.1 %	23.1 %	42.8 %	10.5 %
Calibrated data, $RM = a+b*CM$						
Slope, b		0.898	1.026	1.126	0.819	0.961
Intercept, a		2.237	1.519	-0.188	0.555	-0.201
Expanded rel. uncertainty	25 %	16.4 %	26.2 %	10.0 %	9.5 %	6.9 %
Expanded rel. uncertainty, a=0	25 %	18.9 %	26.7 %	10.0 %	9.5 %	6.8 %

Legend: Fidas200: Palas Fidas 200 (Fidas1), EDM180: Grimm EDM 180 (Grimm1)

TEOM DF: TEOM 1405 DF (TEOMDF), FH62IR: TEI FH62IR (TEI3), FH62IRa: TEI FH62IR (TEI3a)

TEOM DF, FH62IR and FH62IRa failed the data coverage criterion because they did not participate in the whole comparison campaign.

Both EDM 180 candidates failed to pass the expanded relative uncertainty test after calibration for PM₁₀ in the first comparison. They also failed to pass the uncertainty test at the Hjortneskaia roadside site during winter season. This site is close to the harbour. A second comparison at the roadside site Smestad during winter season was organised to rule out the possible effect of sea salt. Both candidates passed the uncertainty test at Smestad. The expanded relative uncertainty of all data decreased after the second comparison, but candidate Grimm1 still failed to pass the uncertainty test, also after calibration.

Figure 10 shows the site and season dependency of the slope of the calibration function for all candidates during the comparison. Only EDM 180 (Grimm1) and Fidas 200 (Fidas2) participated at all sites. EDM 180 had the highest spread in slope values. The candidate overestimated the results by 33 % during summer season at the urban background site (UBS) and underestimated the results by 39 % during autumn season at the roadside site (RSA). Still it had a slope close to 1 based on data from all sites and seasons, illustrating the challenges of how to apply the calibration factors. TEOM 1405 DF (TEOMDF) had the smallest spread in the slope varying from 0.963 during summer season at the urban background (UBS) site to 1.111 during winter season at the same site (UBW).




Figure 10. Site and season dependent slope of the calibration function, PM_{10} candidates. Red line indicates 1:1 relationship with reference. Raw candidate data is multiplied by the slope and the intercept is added to the result to get calibrated data. The candidate reports values higher than the reference sampler when slope is less than 1.00 (red line). See Table 9 for legend explanation.

Figure 11 shows the site and season dependency of the intercept of the calibration function for all candidates. There was some spread in the intercepts of Fidas 200 (Fidas1) and EDM 180 (Grimm1) candidates with some intercepts as high as $4.5 \mu\text{g}/\text{m}^3$. Other candidates had intercepts closer to zero.




Figure 11. Site and season dependent intercept of the calibration function, PM_{10} candidates. Red line indicates zero offset from reference sampler. Raw candidate data is multiplied by the slope and the intercept is added to the result to get calibrated data. The intercept compensates for a constant deviation from the reference values. See Table 9 for legend explanation.

4.2.1 Results from Leckel SEQ47/50 PM_{10} reference samplers

The Leckel SEQ47/50 PM_{10} reference samplers RM2 and RM4 participated in the first comparison. Reference sampler RM4 broke down during startup at the urban background winter comparison (UBW). The instrument was repaired and later deployed again. Only one reference sampler, called RM4-6, participated in the second roadside winter measurement campaign (RSW2). The weighing laboratory experienced problems with high humidity in the weighing room during the urban background summer comparison (UBS). This led to 3 pauses in the comparison.

The average of paired CM data was used in the comparison. Valid data from only one sampler represented the average when data from the other sampler was not available. This happened during 41 of 191 days in the first comparison where the reference samplers run in parallel.

Figure 12 shows PM₁₀ reference data from all sites and seasons.




Figure 12. PM₁₀ reference data from RM2 and RM4, all sites and seasons as indicated by vertical bars. Only one reference sampler RM4-6 participated in the second roadside winter (RSW2) measurement campaign.

Figure 13 shows the relationship between the two RMs. There is little spread in the data.




Figure 13. Leckel SEQ47/50 PM₁₀ references, RM4 vs RM2, data from first comparison. Red line indicates 1:1 relationship.

Table 18 in App. A shows the performance characteristics of the RM. Suitability of data was calculated from averages of RM data (paired and not paired). 23 % of the samples were above the 28 µg/m³ criterion for suitability of data. This is above the 20 % minimum criterion and the criterion is fulfilled.

The between RM uncertainty $u_{bs,RM}$ of all data was $0.69 \mu\text{g}/\text{m}^3$. For all data above $30 \mu\text{g}/\text{m}^3$ it was $0.92 \mu\text{g}/\text{m}^3$. This is below the $2.5 \mu\text{g}/\text{m}^3$ maximum criterion indicating good relationship between the references.

4.2.2 Comparison of Palas Fidas 200 PM₁₀ candidate

The Palas Fidas 200 PM₁₀ candidates Fidas1 and Fidas2 participated in the first comparison. Candidates Fidas1-3 and Fidas1-4 participated in the second comparison at the Smestad roadside site (RSW2). All candidate instruments participated without technical problems.

Figure 14 shows candidate (CM) and reference (RM) data from all sites and seasons. The CMs compare well with the RM during the first comparison, measuring some lower concentrations during autumn season at the roadside site (RSA). The CMs measure higher concentrations than the RM during the second comparison at Smestad roadside site RSW2.




Figure 14. Fidas 200 PM₁₀ candidates Fidas1 and Fidas2, data from all sites and seasons as indicated by vertical bars. Candidates Fidas1-3 and Fidas2-4 were in operation at RSW2.

Figure 15 shows the relationship between the CMs and each CM and the RM.




Figure 15. Palas Fidas 200 PM_{10} candidates Fidas2 vs Fidas1, Fidas1 and Fidas2 vs reference method, first comparison. Red line indicates 1:1 relationship.

Table 19 in App. A shows the performance characteristics of the CM. The between-CM uncertainty $u_{bs,CM}$ of all data was $0.38 \mu\text{g}/\text{m}^3$. For all data above $30 \mu\text{g}/\text{m}^3$ it was $0.68 \mu\text{g}/\text{m}^3$. This is below the $2.5 \mu\text{g}/\text{m}^3$ maximum criterion indicating good relationship between the candidates.

The expanded relative uncertainty of the uncorrected Fidas1 and Fidas2 data in the comparability test was 21.8 % and 21.7 % respectively. This is below the maximum criterion of 25 % and both candidates passed the test. The slopes of the comparability functions were 1.1141 and 1.1190, respectively, which is outside the criterion of 1.0 ± 0.2 . Their intercepts were -2.4923 and -2.5496, respectively, which is less than the minimum criterion of -1, requiring calibration. After calibration, the expanded relative uncertainties decreased to 16.4 % and 15.6 % respectively and both candidates passed the uncertainty test at all sites and seasons. Calibration of all data using only slope resulted in expanded relative uncertainties of 18.9 % and 18.3 % respectively. The candidate failed the expanded relative uncertainty test for values above $30 \mu\text{g}/\text{m}^3$.

Figure 16 indicates the slope of the calibration function for each site and season. The slope of the calibration function varied from 0.837, overestimating PM levels by 19 %, during summer season at the urban background site (UBS) to 1.047, underestimating PM levels by 5 %, during winter season at the same site (Fidas1). The intercept of the calibration function varied from $-0.51 \mu\text{g}/\text{m}^3$ during winter season at the Smestad roadside site (RSW2) to $3.625 \mu\text{g}/\text{m}^3$ during winter season at the Hjortneskaia roadside site (RSA). For all data, the slope and intercept were 0.898 and $2.237 \mu\text{g}/\text{m}^3$ respectively. Based on the results, if the slope for all data is applied in the field the results may be reported from 14 % too low to 7 % too high depending on site and season.




Figure 16. Palas Fidas 200 PM₁₀ Fidas1 and Fidas2 candidates, season and site dependent slope of the calibration function. Red line indicates 1:1 relationship with reference. The candidate reports higher values than the reference sampler when slope is less than 1.00 (red line). Candidates Fidas1-3 and Fidas2-4 were in operation at RSW2.

Palas Fidas 200 participated in the second comparison at the Smestad roadside site (RSW2) to investigate why both CMs failed to pass the uncertainty test for PM_{2.5} at the Hjortneskaia roadside site (RSW), both during winter season, see 4.3.2. For PM₁₀, the CMs passed the test at RSW with expanded relative uncertainties of 22 %. The expanded relative uncertainty of the CMs at RSW2 was 16.5 %. The slope of the calibration function changed from approximately 0.93 at RSW to 0.85 at RSW2. The intercept of the calibration function changed from around 3.5 µg/m³ at RSW to close to zero at RSW2.

4.2.3 Comparison of Grimm EDM 180 PM₁₀ candidate

The Grimm EDM 180 PM₁₀ candidates Grimm1 and Grimm2 participated in the first comparison. Candidates Grimm1-3 and Grimm1-4 participated in the second comparison at the Smestad roadside site (RSW2). The Grimm2 candidate broke down during startup at the urban background site, winter season (UBW), and was replaced by Grimm2a for the remaining comparison. Both are called Grimm2.

Figure 17 shows candidate (CM) and reference (RM) data from all sites and seasons. Both CMs measured concentrations below the RM at the roadside site. There was a tendency to measure above the RM at the urban background site and also at Smestad, the second roadside site (RSW2).




Figure 17. Grimm EDM 180 PM₁₀ candidates Grimm1 and Grimm2, data from all sites and seasons as indicated by vertical bars. Candidates Grimm1-3 and Grimm2-4 were operated at RSW2.

Figure 18 shows the relationship between the CMs and each CM and the RM. There is some spread in the CM vs RM data.




Figure 18. Grimm EDM 180 PM₁₀ candidates Grimm2 vs Grimm1, Grimm1 and Grimm2 vs reference method, first comparison. Red line indicates 1:1 relationship.

Table 25 in App. A shows the performance characteristics of the CM. The between-CM uncertainty $u_{b,CM}$ of all data was $0.94 \mu\text{g}/\text{m}^3$. For all data above $30 \mu\text{g}/\text{m}^3$ it was $1.78 \mu\text{g}/\text{m}^3$. This is below the $2.5 \mu\text{g}/\text{m}^3$ maximum criterion indicating good relationship between the CMs.

The expanded relative uncertainty of the uncorrected Grimm1 and Grimm2 data in the comparability test was 27.1 % and 25.3 % respectively. The uncertainties of both candidates were above the maximum uncertainty criterion of 25 % and failed the test. After calibration, the expanded relative uncertainty was reduced to 26.2 % and 24.2 % respectively and candidate Grimm1 failed the test. The expanded relative uncertainties of both candidates were less than 15 % at all sites and seasons except during winter season at the roadside site (RSW) where the uncertainties were 34.9 % and 34.3 % respectively. Calibration of all data using only intercept resulted in expanded relative uncertainties of

25.4 % and 24.2 % respectively and Grimm1 failed the test again. The candidate failed the expanded relative uncertainty test for values above $30 \mu\text{g}/\text{m}^3$.

Figure 19 indicates the slope of the calibration function for each site and season. The slope of the calibration function varied from 0.674, overestimating PM levels by 48 %, during summer season at the urban background site (UBS) to 1.394, underestimating PM levels by 28 %, during autumn season at the Hjortneskaia roadside site (RSA) (Grimm1). The intercept varied from $0.965 \mu\text{g}/\text{m}^3$ during autumn season at the roadside site to $4.684 \mu\text{g}/\text{m}^3$ during winter season at the same site. For all data, the slope and intercept were 1.026 and $1.519 \mu\text{g}/\text{m}^3$ respectively. Based on the results, if the slope for all data is applied in the field, the results may be reported from 26 % too low to 52 % too high depending on site and season.




Figure 19. Grimm EDM 180 PM_{10} candidates Grimm1 and Grimm2, season and site dependent slope of the calibration function. Red line indicates 1:1 relationship with reference. The candidate reports higher values than the reference sampler when slope is less than 1.00 (red line). Candidates Grimm1-3 and Grimm2-4 were operated at RSW2.

The roadside site Hjortneskaia is close to the harbour and sea salt may have had an effect on the measurements. A second comparison at the roadside site Smestad during winter season (RSW2) was organised to rule out the possible effect of sea salt. Both candidates Grimm1-3 and Grimm2-4 passed the uncertainty test at Smestad. The averages of pairs of Grimm1-3 and Grimm2-4 data were set to represent both Grimm1 and Grimm2 data at RSW2. When the averaged data series was added to the first comparison data the relative uncertainty decreased, but candidate Grimm1 still failed to pass the uncertainty test, also after calibration.

The slope of the calibration function changed from 1.161 at RSW to 0.783 at RSW2. The intercept of the calibration function was comparable changing from $4.7 \mu\text{g}/\text{m}^3$ at RSW to $4.5 \mu\text{g}/\text{m}^3$ at RSW2. It was not possible to explain the differences.

4.2.4 Comparison of TEOM 1405DF PM_{10} candidate

Only 1 TEOM 1405 DF candidate, called TEOMDF, participated in the first comparison. It was fresh from the factory. During service after the first measurement period (RSA) it was discovered that the tubes leading sampled air to the coolers were interchanged at the factory so that the PM_{10} fraction was measured by the $\text{PM}_{2.5}$ unit and vice versa. All data from RSA was discarded. Due to the repairs, there are only 19 samples from RSW.

Figure 20 shows candidate (CM) and reference (RM) data from all sites and seasons. The CMs measured concentrations below the reference method during the winter season (RSW and UBW), and close to the RM during the summer season at the urban background site (UBS).




Figure 20. TEOM 1405 DF PM_{10} candidate TEOMDF, data from all sites and seasons as indicated by vertical bars.

Figure 21 shows the relationship between the CM and the RM. There is little spread in the data except for one possible outlier. The CM reported results close to the RM.




Figure 21. TEOM 1405 DF PM_{10} candidate TEOMDF vs reference method, first comparison. Red line indicates 1:1 relationship.

Table 31 in App. A shows the performance characteristics of the CM. Because only one CM participated it was not possible to calculate the between-CM uncertainty.

The expanded relative uncertainty of the uncorrected TEOMDF data in the comparability test was 23.1 %. TEOMDF passed the uncertainty test, but there was a significant deviation in the slope of the regression. After calibration, the expanded relative uncertainty was reduced to 10.0 %. The candidate

passed the test for all data and each site and season. The expanded relative uncertainty was relatively high at 21.7 % at the roadside site (RSW), but considerably lower at 5.7 % at the urban background site (UBW and UBS). Calibration of all data using only slope resulted in an unchanged expanded relative uncertainty of 10.0 %. The candidate failed the expanded relative uncertainty test for values above $30 \mu\text{g}/\text{m}^3$, but the number of values was less than the required 40.

Figure 22 indicates the slope of the calibration function for each site and season. The slope of the calibration function varied from 0.963, overestimating PM levels by 4 %, during summer season at the urban background (UBS) site to 1.111, underestimating PM levels by 10 %, during winter season at the same site (UBW). The intercept of the calibration function varied from $0.078 \mu\text{g}/\text{m}^3$ during summer season at the urban background site to $1.971 \mu\text{g}/\text{m}^3$ during winter season at the roadside site (RSW). For all data, the slope and intercept were 1.126 and $-0.188 \mu\text{g}/\text{m}^3$, respectively. Based on the results, if the slope for all data is applied in the field the results may be reported 1 % to 17 % too high depending on site and season.




Figure 22. TEOM 1405 DF PM_{10} candidate TEOMDF, season and site dependent slope of the calibration function. Red line indicates 1:1 relationship with reference. The candidate reports higher values than the reference sampler when slope is less than 1.00 (red line).

4.2.5 Comparison of TEI FH62 IR PM_{10} candidate with SA 246b (USEPA) impactor

Only one TEI FH62 IR candidate with SA 246b (USEPA) impactor, called TEI3, participated in the first comparison. TEI3 participated only in the comparison at the roadside site (RSA and RSW). The candidate instrument participated without technical problems.

Figure 23 shows candidate (CM) and reference (RM) data from the roadside site. The CM measured concentrations above the RM method during both seasons.




Figure 23. TEI FH62 IR PM_{10} candidate TEI3, with SA 246b (USEPA) impactor, data from all sites and seasons. The sites and seasons are indicated between the vertical bars.

Figure 24 shows the relationship between the CM and the RM. TEI3 shows little spread in the data but overestimates the results.




Figure 24. TEI FH62 IR PM_{10} candidate TEI3, with SA 246b (USEPA) impactor vs reference method, first comparison. Red line indicates 1:1 relationship.

Table 33 in App. A shows the performance characteristics of the CM. Because only one CM participated it was not possible to calculate the between-CM uncertainty.

The expanded relative uncertainty of the uncorrected TEI3 data in the comparability test was 42.8 %. This is above the maximum allowed expanded relative uncertainty criterion of 25 %. After calibration, the expanded relative uncertainty was reduced to 9.5 % and the candidate passed the test for all data and both seasons at the roadside site. Calibration of all data using only slope resulted in an unchanged expanded relative uncertainty of 9.5 %. The candidate failed the expanded relative uncertainty test for values above $30 \mu\text{g}/\text{m}^3$, but the number of values was less than the required 40.

Figure 25 indicates the slope of the calibration function for each site and season. The slope of the calibration function varied from 0.783, overestimating PM levels by 28 %, during autumn season at the roadside site (RSA) to 0.829, overestimating PM levels by 21 %, during winter season at the same site. The intercept of the calibration function varied from 0.175 $\mu\text{g}/\text{m}^3$ during winter season at the roadside site (RSW) to 1.334 $\mu\text{g}/\text{m}^3$ during autumn season at the same site. For all data, the slope and intercept were 0.819 and 0.555 $\mu\text{g}/\text{m}^3$ respectively. Based on the results, if the slope for all data is applied in the field the results may be reported from 1 % too low to 5 % too high at the roadside site depending on the season.




Figure 25. TEI FH62 IR PM_{10} candidate TEI3 with SA 246b (USEPA) impactor, season and site dependent slope of the calibration function. Red line indicates 1:1 relationship with reference. The candidate reports higher values than the reference sampler when slope is less than 1.00 (red line).

4.2.6 Comparison of TEI FH62 IR PM_{10} candidate with EN12341 impactor

Only one TEI FH62 IR candidate with EN12341 impactor, called TEI3a, participated in the first comparison. TEI3a participated only in the comparison at the urban background site. The candidate participated without technical problems.

Figure 33 shows candidate (CM) and reference (RM) data from the urban background site. The CM measured concentrations close to the RM during winter season (UBW) and above the RM during summer season (UBS) at the urban background site.

Figure 33 shows PM_{10} candidate data from all sites and seasons.




Figure 26. TEI FH62 IR $\text{PM}_{2.5}$ candidate TEI3a, with EN12341 impactor, data from all sites and seasons. The sites and seasons are indicated between the vertical bars.

Figure 32 shows the relationship between the CM and the RM. There is little spread in the data and the CM reports results close to the reference method.




Figure 27. TEI FH62 IR PM_{10} candidate TEI3a, with EN12341 impactor vs reference method, first comparison. Red line indicates 1:1 relationship.

Table 35 in App. A shows the performance characteristics of the CM. Because only one CM participated, it was not possible to calculate the between-CM uncertainty.

The expanded relative uncertainty of the uncorrected TEI3a data in the comparability test was 10.5 %. This is below the maximum allowed expanded relative uncertainty criterion of 25 %. The slope was significant and required calibration. After calibration, the expanded relative uncertainty was reduced to 6.9 % and TEI3a passed the test for all data and both seasons at the urban background site. Calibration of all data using only slope resulted in an almost unchanged expanded relative uncertainty of 6.8 %. The candidate failed the expanded relative uncertainty test for values above 30 $\mu\text{g}/\text{m}^3$ but the number of values was less than the required 40.

Figure 25 indicates the slope of the calibration function for each site and season. The slope of the calibration function varied from 0.861, overestimating PM levels by 14 %, during summer season at the urban background site (UBS) to 0.989, overestimating PM levels by 1 %, during winter season (UBW) at the same site. The intercept of the calibration function varied from $-0.072 \mu\text{g}/\text{m}^3$ during winter season at the urban background site to $0.433 \mu\text{g}/\text{m}^3$ during summer season at the same site, both close to zero. For all data, the slope and intercept were 0.961 and -0.201 respectively. Based on the results, if the slope for all data is applied in the field the results may be reported from 3 % too low to 12 % too high at the urban background site depending on the season.




Figure 28. TEI FH62 IR PM_{10} candidate TEI3a with EN12341 impactor, season and site dependent slope of the calibration function. Red line indicates 1:1 relationship with reference. The candidate reports higher values than the reference sampler when slope is less than 1.00 (red line).

4.3 Results from $\text{PM}_{2.5}$ comparison

The performance characteristics of the $\text{PM}_{2.5}$ candidates based on data from both comparisons are summarized in Table 10. Numbers in red indicate significant deviations from performance criteria. Numbers in blue indicate a deviation in slope and/or intercept that makes calibration necessary. All paired candidates had satisfactory between-candidates uncertainties except for FH62IR where the second candidate (TEI2) was unstable throughout the whole measurement campaign and all its results were rejected. All candidates, except Fidas 200 and TEOM 1405 DF, failed to pass the comparability test for $\text{PM}_{2.5}$ with expanded relative uncertainties higher than 25 %. All candidates except Fidas 200 had significant deviation in the slope, while only TEOM 1400 AB had significant deviation in the offset. After calibration, all candidates passed the test for expanded relative uncertainty below 25 %. The expanded relative uncertainty after calibration was also calculated using the calibration function with the intercept a set to 0.0 (zero). The expanded relative uncertainties of all candidates except TEOM 1400 AB remained unchanged due to their intercepts being close to zero, indicating that the calibration function may be applied with the intercept a set to zero.

Table 10. Summary of performance characteristics of the PM_{2.5} candidates. Numbers in red indicate significant deviations from performance criteria. Numbers in blue indicate a deviation in slope and/or intercept that makes calibration necessary. "x" indicates there was only one candidate participating and between CM uncertainty could not be calculated.

Test PM _{2.5}	Criteria	Fidas200	EDM180	TEOM DF	FH62IR-1	TEOM
Between CM uncertainty	$u_{bs,CM} < 2.5 \mu\text{g}/\text{m}^3$	0.12	0.65	x	4.44	0.91
Comparability						
Number of values		215	230	111	180	183
Data capture	> 90 %	91.5%	97.9 %	58.4 %	94.7 %	96.3 %
Slope, b	$0.98 < b < 1.02$	0.9914	1.1077	0.9568	1.1948	0.6891
Intercept, a	$-1 < a < 1$	-0.6957	-0.3480	0.4113	-0.0020	2.6607
Expanded rel. uncertainty	25 %	18.9 %	26.3 %	10.6 %	46.2 %	45.1 %
Calibrated data, RM = a+b*CM						
Slope, b		1.009	0.903	1.045	0.837	1.451
Intercept, a		0.702	0.314	-0.43	0.002	-3.861
Expanded rel. uncertainty	25 %	19.0 %	17.2 %	10.2 %	22.7 %	12.2 %
Expanded rel. uncertainty, a=0	25 %	19.4 %	17.3 %	10.6 %	22.4 %	30.3 %

Legend: Fidas200: Palas Fidas 200 (Fidas1), EDM180: Grimm EDM 180 (Grimm1)

TEOM DF: TEOM 1405 DF (TEOMDF), FH62IR: TEI FH62IR-1 (TEI1), TEOM: TEOM 1400 AB (TEOM2)

TEOM DF failed the data coverage criterion because it did not participate in the whole comparison campaign.

Both Fidas 200 candidates passed the expanded relative uncertainty test for PM_{2.5} after calibration in the first comparison, but they failed to pass the uncertainty test at the Hjortneskaia roadside site during winter season. A second comparison at the roadside site Smestad during winter season was organised to rule out the possible effect of sea salt. Both candidates passed the uncertainty test at Smestad. The expanded relative uncertainty of all data after calibration did not change.

Figure 29 shows the site and season dependency of the slope of the calibration function for all candidates during the comparison. All candidates except TEOM 1405 DF (TEOMDF) participated at all sites. Fidas 200 (Fidas1) had the highest spread in slope values. The candidate underestimated the results by 44 % during autumn season at the Hjortneskaia roadside site (RSA) and overestimated the results by 26 % during winter season at the Smestad roadside site (RSW2). Still it had a slope close to 1 based on data from all sites and seasons, illustrating the challenges of how to apply the calibration factors. FH62-IR (TEI1) overestimated the results by almost 50 % at the Hjortneskaia roadside site during autumn season. TEOM 1400 AB (TEOM2) underestimated the results at all sites with a maximum of 69 % at the urban background site during winter season.




Figure 29. Site and season dependent slope of the calibration function, $PM_{2.5}$ candidates. Red line indicates 1:1 relationship with reference. Raw candidate data is multiplied by the slope and the intercept is added to the result to get calibrated data. The candidate reports values higher than the reference sampler when slope is less than 1.00 (red line). See Table 10 for legend explanation.

Figure 30 shows the site and season dependency of the intercept of the calibration function for all candidates. Most candidates had an intercept within or close to $0 \pm 1 \mu\text{g}/\text{m}^3$. Both Fidas 200 (Fidas1) and EDM 180 (Grimm1) had intercepts close to $2 \mu\text{g}/\text{m}^3$ at the Smestad roadside site during winter season (RSW2). TEOM 1400 AB (TEOM2) had intercepts between $-3.1 \mu\text{g}/\text{m}^3$ and $-4.8 \mu\text{g}/\text{m}^3$ at all sites.




Figure 30. Site and season dependent intercept of the calibration function, $PM_{2.5}$ candidates. Red line indicates zero offset from reference. Raw candidate data is multiplied by the slope and the intercept is added to the result to get calibrated data. The intercept indicates a constant deviation from the reference values. TEOM 1400 AB (TEOM2) had intercepts between $-3.1 \mu\text{g}/\text{m}^3$ and $-4.8 \mu\text{g}/\text{m}^3$ at all sites. See Table 10 for legend explanation.

4.3.1 Results from Leckel SEQ47/50 PM_{2.5} reference samplers

The Leckel SEQ47/50 PM_{2.5} reference samplers RM1 and RM3 participated in the first comparison. RM1 was in operation for 15 days from start at the roadside site during the autumn season before it was removed for service. The sampler was repaired and deployed again at the start of the urban background winter comparison. Only one reference sampler, called RM3-5, participated in the second roadside winter measurement campaign (RSW2). The weighing laboratory experienced problems with high humidity in the weighing room during the urban background summer comparison. This led to 3 pauses in the comparison.

The average of paired CM data was used in the comparison. Valid data from only one sampler represented the average when data from the other sampler was not available. This happened during 81 of 190 days in the first comparison where the reference samplers run in parallel.

Figure 31 shows PM_{2.5} reference data from all sites and seasons. The RMs compared well.




Figure 31. PM_{2.5} reference data from all sites and seasons as indicated by vertical bars. RM1 was out of operation during most of RSA and all of RSW. Only one reference sampler RM3-5 participated in the second roadside winter (RSW2) measurement campaign.

Figure 32 shows the relationship between the two reference samplers. There is little spread in the data.




Figure 32. Leckel SEQ47/50 $PM_{2.5}$ references, RM3 vs RM1, first comparison. Red line indicates 1:1 relationship.

Table 37 in App. B shows the performance characteristics of the RM. Suitability of data was calculated from averages of RM data (paired and not paired). Only 12 % of the averaged samples were above the $17 \mu\text{g}/\text{m}^3$ criterion for suitability of data. This is below the 20 % minimum criterion and the criterion is not fulfilled. This indicates that the ambient concentration of $PM_{2.5}$ was low during the comparison.

The between-RM uncertainty $u_{\text{bs, RM}}$ of all data was $0.42 \mu\text{g}/\text{m}^3$. For all data above $18 \mu\text{g}/\text{m}^3$ it was $0.53 \mu\text{g}/\text{m}^3$. This is well below the $2.0 \mu\text{g}/\text{m}^3$ maximum criterion indicating good relationship between the references although the latter uncertainty was based only 3 data pairs.

Because of the low between-RM uncertainty it was decided to let the averages of the paired reference values represent the reference values in the data analysis. Valid data from only one sampler represents the average when data from the other sampler was not available.

4.3.2 Comparison of Fidas 200 $PM_{2.5}$ candidate

The Palas Fidas 200 $PM_{2.5}$ candidates Fidas1 and Fidas2 participated in the first comparison. Candidates Fidas1-3 and Fidas1-4 participated in the second comparison at the Smestad roadside site (RSW2). All candidate instruments participated without technical problems.

Figure 33 shows candidate (CM) and reference (RM) data from all sites and seasons. The CMs measured lower concentrations than the RM at the roadside site (RSA and RSW) but compare well with the RM at the urban background site (UBW and UBS). The CMs measure higher concentrations than the RM during the second comparison at Smestad roadside site RSW2.




Figure 33. Fidas 200 $PM_{2.5}$ candidate data from all sites and seasons as indicated by vertical bars. Candidates Fidas1-3 and Fidas2-4 were in operation at RSW2.

Figure 34 shows the relationship between the CMs and each CM and the RM. There is some spread in the data. There is some spread in the CM vs RM data.




Figure 34. Fidas 200 $PM_{2.5}$ candidates Fidas2 vs Fidas1, Fidas1 and Fidas2 vs reference method, first comparison. Red line indicates 1:1 relationship.

Table 38 in App. B shows the performance characteristics of the CM. The between-CM uncertainty $u_{bs,CM}$ of all data was $0.12 \mu\text{g}/\text{m}^3$. For all data above $18 \mu\text{g}/\text{m}^3$ it was $0.25 \mu\text{g}/\text{m}^3$. This is below the $2.5 \mu\text{g}/\text{m}^3$ maximum criterion indicating good relationship between the candidates.

The expanded relative uncertainty of the uncorrected Fidas1 and Fidas2 data in the comparability test was 27.1 % and 27.8 % respectively. This is above the maximum criterion of 25 % and both candidates failed the test. After calibration, the expanded relative uncertainty was reduced to 19.2 % and 18.3 % respectively and both candidates passed the test for all data. Both candidates failed at the roadside site test during winter season (RSW) with expanded relative uncertainties of 27.3 % and 25.3 % respectively. Calibration of all data using only intercept resulted in almost unchanged expanded relative uncertainties of 19.4 % and 18.7 % respectively. The candidate failed the expanded relative uncertainty test for values above $18 \mu\text{g}/\text{m}^3$, but the number of values was less than the required 40.

Figure 35 indicates the slope of the calibration function for each site and season. The slope of the calibration function varied from 0.744, overestimating PM levels by 36 %, during winter season at the Smestad road side site (RSW2) to 1.443, underestimating PM levels by 30 %, during autumn season at the Hjortneskaia roadside site (RSA) (Fidas1). The intercept of the calibration function varied from 0.140 $\mu\text{g}/\text{m}^3$ during winter season at the urban background site (UBW) to 2.075 $\mu\text{g}/\text{m}^3$ during winter season at the Smestad roadside site (RSW2). For all data, the slope and intercept were 1.009 and 0.702 $\mu\text{g}/\text{m}^3$ respectively. Based on the results, if the slope for all data is applied in the field the results may be reported from 30 % too low to 21 % too high depending on site and season.




Figure 35. Palas Fidas 200 PM_{2.5} candidates Fidas1 and Fidas2, season and site dependent slope of the calibration function. Red line indicates 1:1 relationship with reference. The candidate reports higher values than the reference sampler when slope is less than 1.00 (red line). Candidates Fidas1-3 and Fidas2-4 were in operation at RSW2.

The roadside site Hjortneskaia is close to the harbour and sea salt may have had an effect on the measurements. A second comparison at the roadside site Smestad during winter season (RSW2) was organised to rule out the possible effect of sea salt. Both candidates Fidas1-3 and Fidas2-4 passed the uncertainty test at Smestad. There was a difference in the candidates' response at Smestad. The slope of the calibration function changed from 1.083 at RSW to 0.744 at RSW2. The intercept of the calibration function was comparable changing from 1.1 $\mu\text{g}/\text{m}^3$ at RSW to 2.1 $\mu\text{g}/\text{m}^3$ at RSW2. It was not possible to explain the differences.

The averages of pairs of Fidas1-3 and Fidas2-4 data were set to represent both Fidas 1 and Fidas2 data at RSW2. When the averaged data series was added to the first comparison data the relative uncertainty after calibration did not change.

4.3.3 Comparison of Grimm 180 PM_{2.5} candidate

The Grimm EDM 180 PM_{2.5} candidates Grimm1 and Grimm2 participated in the first comparison. Candidates Grimm1-3 and Grimm1-4 participated in the second comparison at the Smestad roadside site (RSW2). The Grimm2 candidate broke down during startup at the urban background site, winter season (UBW), and was replaced by Grimm2a for the remaining comparison. Both are called Grimm2.

Figure 36 shows candidate (CM) and reference (RM) data from all sites and seasons. The CMs measured concentrations around the RM at the roadside site (RSA and RSW) and above the RM at the urban background site (UBW and UBS) and during the second comparison at Smestad roadside site (RSW2).




Figure 36. Grimm EDM 180 $PM_{2.5}$ candidate data from all sites and seasons as indicated by vertical bars. Candidates Grimm1-3 and Grimm2-4 were operated at RSW2.

Figure 37 shows the relationship between the CMs and each CM and the RM. Two slopes are observed in the Grimm2 vs Grimm1 chart due to the change in candidate from Grimm2 to Grimm2a at the urban background site, winter season (UBW), see also Figure 38. The response of Grimm1 was lower than Grimm2 and slightly higher than Grimm2a. There is some spread in the CM vs RM data.




Figure 37. Grimm EDM 180 $PM_{2.5}$ candidates Grimm2 vs Grimm1, Grimm1 and Grimm2 vs reference method, first comparison. Red line indicates 1:1 relationship.

Table 44 in App. B shows the performance characteristics of the CM. The between-CM uncertainty $u_{bs,CM}$ of all data was $0.65 \mu\text{g}/\text{m}^3$. For all data above $18 \mu\text{g}/\text{m}^3$ it was $1.43 \mu\text{g}/\text{m}^3$. This is below the $2.5 \mu\text{g}/\text{m}^3$ maximum criterion indicating good relationship between the candidates.

The expanded relative uncertainty of the uncorrected Grimm1 and Grimm2 data in the comparability test was 26.3 % and 33.8 % respectively. Both candidates were above the maximum uncertainty criterion of 25 % and failed the test. After calibration the expanded relative uncertainty was reduced to 17.2 % and 15.5 % respectively and both candidates passed the test for all data and each site and

season. The expanded relative uncertainty at the roadside site during winter season was 24.4 % and 23.6 % respectively and just below the 25 % limit. Calibration of all data using only slope resulted in almost unchanged expanded relative uncertainties of 17.3 % and 15.9 % respectively. The candidate failed the expanded relative uncertainty test for values above $18 \mu\text{g}/\text{m}^3$, because the number of values was less than the required 40.

Figure 38 indicates the slope of the calibration function for each site and season. The slope of the calibration function varied from 0.727, overestimating PM levels by 38 %, during summer season at the urban background site (UBS) to 1.148, underestimating PM levels by 13 %, during autumn season at the Hjortneskaia roadside site (RSA) (Grimm1). The intercept of the calibration function varied from $-0.222 \mu\text{g}/\text{m}^3$ during winter season at the urban background site (UBW) to $1.977 \mu\text{g}/\text{m}^3$ during winter season at the Smestad roadside site (RSW2). For all data, the slope and intercept were 0.903 and $0.314 \mu\text{g}/\text{m}^3$, respectively. Based on the results, if the slope for all data is applied in the field the results may be reported from 21 % too low to 24 % too high, depending on site and season.




Figure 38. Grimm EDM 180 PM_{2.5} candidates Grimm1 and Grimm2, season and site dependent slope of the calibration function. Red line indicates 1:1 relationship with reference. The candidate reports higher values than the reference sampler when slope is less than 1.00 (red line).

Grimm EDM 180 participated in the second comparison at the roadside site Smestad (RSW2) to investigate why both CMs failed to pass the uncertainty test for PM₁₀ at the Hjortneskaia roadside site during winter season (RSW), see 4.2.3.

For PM_{2.5}, the CMs passed the test at RSW with expanded relative uncertainties close to 24 %. The expanded relative uncertainty of the candidates at RSW2 were 20.7 % and 19.5 %. The slope of the calibration function changed from approximately 0.90 at RSW to 0.76 at RSW2. The intercept of the calibration function changed from around $1.1 \mu\text{g}/\text{m}^3$ at RSW to $1.9 \mu\text{g}/\text{m}^3$ at RSW2.

4.3.4 Comparison of TEOM 1405 DF PM_{2.5} candidate

Only one TEOM 1405 DF candidate, called TEOMDF, participated in the first comparison. It was fresh from the factory. During service after the first measurement period (RSA) it was discovered that the tubes leading sampled air to the coolers were interchanged by the factory so that the PM₁₀ fraction was measured by the PM_{2.5} unit and vice versa. All data from RSA was discarded. Due to the repairs, there are only 19 samples from RSW.

Figure 39 shows candidate (CM) and reference (RM) data from all sites and seasons. The CMs measured concentrations close to the RM at the roadside site (RSW) and urban background site (UBW) during winter season, and above the reference method at the urban background site during summer season (UBS).




Figure 39. TEOM 1405 DF $PM_{2.5}$ candidate data from all sites and seasons as indicated by vertical bars.

Figure 40 shows the relationship between the CM and the RM. There is little spread in the data and the CM reports results close to the RM.




Figure 40. TEOM 1405 DF $PM_{2.5}$ candidate TEOMDF vs reference method, first comparison. Red line indicates 1:1 relationship. Based on data from RSW (partly), UBW and UBS.

Table 50 in App. B shows the performance characteristics of the CM. Because only one CM participated it was not possible to calculate the between-CM uncertainty.

The expanded relative uncertainty of the uncorrected TEOMDF data in the comparability test was 10.6 %. TEOMDF passed the uncertainty test, but there was a significant deviation in the slope of the

regression. After calibration, the expanded relative uncertainty was reduced to 10.2 % which is about the same as for uncorrected data. The candidate passed the test for all data and each site and season. The expanded relative uncertainty was 13.8 % at the roadside site (RSW) and 8.3 % during both seasons at the urban background site (USW and USB). Calibration of all data using only slope resulted in an almost unchanged expanded relative uncertainty of 10.6 %. The number of values was too low to calculate the expanded relative uncertainty test for values above $18 \mu\text{g}/\text{m}^3$.

Figure 41 indicates the slope of the calibration function for each site and season. The slope of the calibration function varied from 0.812, overestimating PM levels by 23 %, during summer season at the urban background site (UBS) to 1.077, underestimating PM levels by 7 %, during winter season at the roadside site (RSA). The intercept of the calibration function varied from $-0.286 \mu\text{g}/\text{m}^3$ during winter season at the roadside site to $0.932 \mu\text{g}/\text{m}^3$ during winter season at the urban background site (UBW), both within $0 \pm 1 \mu\text{g}/\text{m}^3$. For all data, the slope and intercept were 1.045 and $-0.43 \mu\text{g}/\text{m}^3$ respectively. Based on the results, if the slope for all data is applied in the field the results may be reported from 3 % too low to 29 % too high depending on site and season.




Figure 41. TEOM 1405 DF $\text{PM}_{2.5}$ candidate TEOMDF, season and site dependent slope of the calibration function. Red line indicates 1:1 relationship with reference. The candidate reports higher values than the reference sampler when slope is less than 1.00 (red line).

4.3.5 Comparison of TEI FH62 IR $\text{PM}_{2.5}$ candidate with EN12341 impactor

The TEI FH62 IR candidates TEI1 and TEI2 with EN12341 impactor participated in the first comparison. The TEI1 candidate participated without technical problems. The TEI2 candidate was unstable throughout the whole measurement campaign and all its data was rejected.

Figure 43 shows candidate (CM) and reference (RM) data from all sites and seasons. Candidate TEI1 measured concentrations above the reference method at the roadside site during autumn season (RSA), close to the reference method at both sites during winter season (RSW and UBW) and again above the reference method at the urban background site during summer season (UBS).




Figure 42. TEI FH62 IR $PM_{2.5}$ candidate data from all sites and seasons as indicated by vertical bars. TEI2 data is not shown because of too much noise.

Figure 43 shows the relationship between the CMs and each CM and the RM. TEI1 shows some spread in the data and overestimates the results while TEI2 shows excessive noise over the whole measurement range.




Figure 43. TEI FH62 IR $PM_{2.5}$ candidates TEI2 vs TEI1, TEI1 and TEI2 vs reference method, first comparison. Red line indicates 1:1 relationship. There is excessive noise in TEI2.

Table 52 in App. B shows the performance characteristics of the CMs. The between CM uncertainty $u_{bs,CM}$ of all data was $4.44 \mu\text{g}/\text{m}^3$. For all data above $18 \mu\text{g}/\text{m}^3$ it was $6.50 \mu\text{g}/\text{m}^3$. This is above the $2.5 \mu\text{g}/\text{m}^3$ maximum criterion indicating poor relationship between the candidates. Due to the instability of the TEI2 analyser it was not evaluated further.

The expanded relative uncertainty of the uncorrected TEI1 data in the comparability test was 46.2 %. This is above the maximum allowed expanded relative uncertainty criterion of 25 %. After calibration, the expanded relative uncertainty was reduced to 22.7 %. The candidate TEI1 passed the test for all data and each site and season except for the roadside site during autumn season (RSA). The expanded relative uncertainty at the roadside site during autumn season was 24.9 % and just below the 25 %

limit. TEI1 performed better at the urban background site where the expanded relative uncertainties were 11.1 % and 12.6 % during winter (UBW) and summer seasons (UBS) respectively. Calibration of all data using only slope resulted in an almost unchanged expanded relative uncertainty of 22.4 %. The candidate failed the expanded relative uncertainty test for values above $18 \mu\text{g}/\text{m}^3$, because the number of values was less than the required 40.

Figure 44 indicates the slope of the calibration function for each site and season. The slope of the calibration function varied from 0.538, overestimating PM levels by 86 %, during autumn season at the roadside site (RSA) to 1.118, underestimating PM levels by 11 %, during winter season at the urban background site (UBW). The intercept of the calibration function varied from $-1.064 \mu\text{g}/\text{m}^3$ during winter season at the urban background site to $1.161 \mu\text{g}/\text{m}^3$ during autumn season at the roadside site. For all data the slope and intercept were 0.837 and $0.002 \mu\text{g}/\text{m}^3$, respectively. Based on the results, if the slope for all data is applied in the field the results may be reported from -25 % too low to 56 % too high depending on site and season.




Figure 44. TEI FH62 IR PM_{2.5} candidate TEI1, season and site dependent slope of the calibration function. Red line indicates 1:1 relationship with reference. The candidate reports higher values than the reference sampler when slope is less than 1.00 (red line). TEI2 data was rejected due to unstable readings.

4.3.6 Comparison of TEOM 1400 AB PM_{2.5} candidate

The TEOM 1400 AB candidates TEOM1 and TEOM2 participated in the first comparison. Both candidates participated without technical problems.

Figure 45 shows PM_{2.5} candidate (CM) and reference (RM) data from all sites and seasons. The CMs measured concentrations below the RM at the roadside site and around the reference method at the urban background site (UBW and UBS). The CM usually measured higher than the RMs at low levels indicating an offset in the CM values.




Figure 45. TEOM 1400 AB $PM_{2.5}$ candidate data from all sites and seasons as indicated by vertical bars.

Figure 46 shows the relationship between the CMs and each CM and the RM. There is some spread in the data. The spread seems to be constant independent of level. Both CMs have an offset of about $4 \mu\text{g}/\text{m}^3$. Results above approximately $10 \mu\text{g}/\text{m}^3$ are reported lower than the reference method.




Figure 46. TEOM 1400 AB $PM_{2.5}$ candidates TEOM2 vs TEOM1, TEOM1 and TEOM2 vs reference method, first comparison. Red line indicates 1:1 relationship.

Table 55 in App. B shows the performance characteristics of the CM. The between-CM uncertainty $u_{\text{bs,CM}}$ of all data was $0.91 \mu\text{g}/\text{m}^3$. For all data above $18 \mu\text{g}/\text{m}^3$ it was $0.45 \mu\text{g}/\text{m}^3$. This is below the $2.5 \mu\text{g}/\text{m}^3$ maximum criterion indicating good relationship between the candidates.

The expanded relative uncertainty of the uncorrected TEOM1 and TEOM2 data in the comparability test was 46.2 % and 45.1 % respectively. The results were above the maximum uncertainty criterion of 25 % and both TEOMs failed the test. After calibration, the expanded relative uncertainty was reduced to 16.0 % and 12.2 % respectively. Both candidates passed the test for all data. TEOM1 passed the uncertainty test for all sites and seasons except for the summer season at the urban background site (UBS). TEOM2 passed the uncertainty test for all sites and seasons. Calibration of all data using only intercept or only slope did not improve the expanded relative uncertainties. The candidate failed the

expanded relative uncertainty test for values above $18 \mu\text{g}/\text{m}^3$, because the number of values was less than the required 40.

Figure 47 indicates the slope of the calibration function for each site and season. The slope of the calibration function varied from 1.238, underestimating PM levels by 19 %, during summer season at the urban background site (UBS) to 1.686, underestimating PM levels by 41 %, during winter season at the same site (UBW). The intercept of the calibration function varied from $-4.765 \mu\text{g}/\text{m}^3$ during winter season at the urban background site to $-3.111 \mu\text{g}/\text{m}^3$ during summer season at the same site. TEOM 1400 AB was operated with the factory set offset of $3 \mu\text{g}/\text{m}^3$ which may account for some of the intercept. For all data, the slope and intercept were 1.451 and $-3.861 \mu\text{g}/\text{m}^3$ respectively (TEOM2). Both candidates had approximately constant slope values during both seasons at the roadside site (RSA and RSW). Based on the results, if the slope for all data is applied in the field the results may be reported from 14 % too low to 17 % too high depending on site and season, not taking the intercept into account.




Figure 47. TEOM 1400 AB PM_{2.5} candidates TEOM1 and TEOM2, season and site dependent slope of the calibration function. Red line indicates 1:1 relationship with reference. The candidate reports higher values than the reference sampler when slope is less than 1.00 (red line).

5 Discussion and conclusion, PM comparison

5.1 PM₁₀ candidates

Of all PM₁₀ candidates, only Fidas 200 had a satisfactory data coverage. One of the EDM 180 candidates broke down during startup at the urban background site and had to be replaced. When including the replacement instrument, EDM 180 also had satisfactory data coverage. The three remaining candidates participated with only one instrument. TEOM 1405 DF was shipped from the factory with sample tubes interchanged and most of the data from the Hjortneskaia roadside site had to be rejected. The FH62 IR instrument ran half the comparison with USEPA inlet and the other half with EN12341 inlet. Repeatability in the field and slope of the calibration function representing more than one site type can only be claimed for Fidas 200 and EDM 180.

Both Fidas 200 and EDM 180 had between-candidate uncertainties well below the $2.5 \mu\text{g}/\text{m}^3$ criterion indicating good repeatability in the field. EDM 180 and FH62 IR with USEPA inlet failed to pass the comparability test for PM₁₀ with expanded relative uncertainties higher than 25 %. FH62 IR with EN12341 inlet had the lowest expanded relative uncertainty of all candidates (10.5 %). It is interesting to note that the candidate had much lower expanded relative uncertainty than the same instrument with USEPA inlet (42.8 %). This is probably caused by the inlet type or the change from roadside site to

urban background site. After calibration, all candidates except EDM 180 passed the test for expanded relative uncertainty below 25 %. TEOM 1405 DF and FH62 IR had lower expanded uncertainties than the others with FH62 IR using EN12341 inlet having the lowest (6.8 %). The optical methods Fidas 200 and EDM 180 had more spread in their data than the other methods.

The candidates had calibration functions that were site and season dependent compared to the reference method. All candidates measured higher than the reference method during summer season at the urban background site. EDM 180 had the highest seasonal variation in the calibration function varying from measuring 28 % too low during autumn season at the roadside to measuring 48 % too high during summer season at the urban background site. FH62 IR using USEPA inlet had the lowest variation measuring from 21 % to high during winter season at the roadside site to 28 % to high during summer season at the same site.

EDM 180 failed to pass the comparability test at the Hjortneskaia roadside site during winter season (RSW). A second comparison at the roadside site Smestad during winter season (RSW2) was organised to rule out the possible effect of sea salt. The candidates now passed the test. The slope of the calibration function changed from 1.161 at RSW to 0.783 at RSW2. The change in response from measuring lower than the reference method to measuring higher indicates a difference in site characteristics. It was not possible to quantify the difference. The intercept increased a little at Smestad.

5.2 PM_{2.5} candidates

Of all PM_{2.5} candidates, only Fidas 200 and TEOM 1400 AB had a satisfactory data coverage. One of the EDM 180 candidates broke down during startup at the urban background site and had to be replaced. When including the replacement instrument, EDM 180 also had satisfactory data coverage. The second FH62 IR candidate suffered from noisy values throughout the whole comparison, and all its data was rejected. TEOM 1405 DF was shipped from the factory with sample tubes interchanged and most of the data from the Hjortneskaia roadside site had to be rejected. Repeatability in the field and slope of the calibration function representing more than one site type can only be claimed for Fidas 200, EDM 180 and TEOM 1400 AB.

Both Fidas 200, EDM 180 and TEOM 1400 AB had between-candidate uncertainties well below the 2.5 µg/m³ criterion indicating good repeatability in the field. EDM 180, FH62 IR (EN12341 inlet) and TEOM 1400 AB failed to pass the comparability test for PM_{2.5} with expanded relative uncertainties higher than 25 %. TEOM 1405 DF had the lowest expanded relative uncertainty of all candidates (10.6 %). After calibration, all candidates passed the test for expanded relative uncertainty below 25 %. TEOM 1405 DF had the lowest expanded uncertainty (10.6 %). The optical methods Fidas 200 and EDM 180, and the beta-gauge method FH62 IR had more spread in their data than the other methods.

The candidates had calibration functions that were site and season dependent compared to the reference method. All candidates, except FH62 IR, measured higher than the reference method during summer season at the urban background site. FH62 IR had the highest seasonal variation in the calibration function varying from measuring 11 % too low during winter season at the urban background site to measuring 86 % too high during autumn season at the roadside site. TEOM 1400 AB on the other hand measured 48 % too low during winter season at the urban background site.

Fidas 200 failed to pass the comparability test at the Hjortneskaia roadside site during winter season (RSW). A second comparison at the roadside site Smestad during winter season (RSW2) was organised to rule out the possible effect of sea salt. The candidates now passed the test. The slope of the calibration function changed from 1.083 at RSW to 0.744 at RSW2. The change in response from measuring lower than the reference method to measuring higher indicates a difference in site characteristics. It was not possible to quantify the difference. The intercept remained almost unchanged.

6 Possible system for Norway to calibrate AMS PM data

The results from the PM comparison carried out in Oslo in 2015/2016 and 2018 (described in the chapters above) suggest that PM monitoring methods may be affected by parameters depending on season and on location. So far, the data basis is too limited to derive relationships between external parameters (meteorology, sources, etc.) and discrepancies between AMS and reference methods, which could allow correcting the data. Both climatic conditions and PM source contributions vary strongly over Norway – comparison results derived from measurements in Oslo will not be representative for all monitoring stations in Norway. Experience from other European countries shows that the discrepancies between methods vary both with season and with measurement site. Since the discrepancies can be quite significant, it is important to find a reliable method to correct² the PM data, which is reported to the EEA, in order to obtain a homogeneous data set which is independent of the monitoring method. Correct PM levels are required to derive daily average concentrations, annual average concentrations and number of exceedances, which may trigger the need for mitigation measures.

6.1 Background

EU's *air quality directive* (2008/50/EC, CAFÉ-directive, [2]) requires the member states (Norway included through the EEA agreement) to assess the air quality when certain criteria are fulfilled (exceedance of assessment thresholds, directive 2008/50/EC, Annex V). Requirements for assessing air quality in Norway are laid down in Forurensningsforskriften § 7-8. The requirement to assess air quality and the requirement to apply the **reference method** are established in the air quality directive. Measurement method requirements shall assure that the measurements have the required quality and are comparable. The reference method³ for sampling and analysis of PM₁₀ and PM_{2.5} is defined in the CAFÉ directive (Annex VI, section A, 4 and A, 5) and described in "NS-EN 12341:2014 Ambient air – Standard gravimetric measurement method for the determination of the PM₁₀ or PM_{2,5} mass concentration of suspended particulate matter". The standard requires sampling of PM on filter, followed by gravimetric analysis. This measurement method is not suitable for supplying online information to the public because the laboratory analysis delays the reporting of the results and the averaging time is 24 hours. According to the CAFÉ directive (Annex VI, section B, 1), the member states may use any other method which the member state can demonstrate displays a **consistent relationship to the reference method**, i.e., the relative expanded uncertainty is less than or equal to 25 % (data quality objectives). A measurement method that fulfils this requirement is regarded an *equivalent method*. The member states may, according to Annex VI, section B, 2 in the CAFÉ directive, be required by the European Commission to submit a report demonstrating the equivalence of the measurement method. Annex VI, section B, 3 in the CAFÉ directive refers to the *Guide to the demonstration of equivalence* (GDE) on the method for demonstrating equivalence, also see Section 6.1.1 below. There are a number of commercially available automated measuring systems (AMS) for measuring suspended particulates in ambient air in near real time. All measurement systems in operation in Norway have been tested successfully for equivalence by specialised test laboratories, such as TÜV in Germany. The test method is based on comparison by having reference sampler and automatic analyser measuring in parallel.

² The term "correction" has been used historically, but is replaced by the term "calibration" in the context of demonstrating equivalence of candidate methods for monitoring PM (GDE, 2010).

³ Reference method: measurement method which, by convention, gives the accepted reference value of the measured compound (EN 12341).

6.1.1 Equivalence testing

Testing for equivalence of a measurement method for measurement of local air quality according to “Forurensningsforskriften”/air quality directive includes three levels as described in EN 16450:

- *Type testing* (NS-EN 16450:2017, chapter 7) is performed once by an accredited test laboratory, on contract from the analyser (AMS) manufacturer. The type approval includes an evaluation of the performance characteristics of the AMS based on a series of tests in the laboratory and in the field. Two analysers of the same pattern are tested in parallel, both in the laboratory and in the field. The AMS is accepted as an equivalent method if both AMS units fulfil the performance criteria and data quality objectives.
- *Suitability test* (NS-EN 16450:2017, chapter 8.2) is performed once by the body responsible for the field operation before a type-approved AMS is put into operation. The test includes one AMS unit measuring in parallel with the reference method at one or more locations representative for conditions in the measurement network. The test follows the procedure in NS-EN 16450, chapter 7.5 “Field test procedures” and is limited to an evaluation of (a) the calibration function and (b) the expanded relative uncertainty of the AMS.
- *Ongoing verification of suitability* of the AMS (NS-EN 16450:2017, chapter 8.6; and Section 6.1.2 below) aims at ensuring the ongoing quality of the measurements obtained using the AMS by periodically checking the validity of the equivalence test and, if necessary, establish new calibration functions.

The method to test equivalence for an AMS is described in EN 16450, chapter 7.5. When an analyser type is put into operation for the first time in the national measurement network, the test is carried out as specified in chapter 7.5, i.e., using two AMS units of the same type (model, hardware, firmware and software, configuration, and version) measuring parallel with two reference samplers.

6.1.2 Ongoing verification of equivalence

The method of **ongoing verification of equivalence** of automatic PM analysers, described in EN 16450:2017 chapter 8.6, is almost identical with the procedure in chapter 9.9 in GDE.

NS-EN 16450:2017 (chapter 8.6) requires ongoing verification of the AMS at a number⁴ of sites **every year** to ensure that the results from the equivalence test are still valid and if necessary establish a new calibration function (GDE, chapter 9.9.2, EN 16450, chapter 8.6.2). This is important because the type approval tests and suitability evaluation were carried out under a limited number of particulate compositions, which may not continue to be representative for the actual conditions (EN 16450, chapter 8.6.2).

The verification includes one AMS and one reference instrument measuring side by side (EN 16450, chapter 8.6.2). The test should include **80 valid data pairs**⁵ covering **a full year**, e.g., by sampling **every four days**, at one or more sites representative of the various conditions that are typical for the network.

The number of measurement stations to be tested every year depends on the relative expanded uncertainty, W_{AMS} , resulting from combining the data obtained in the type approval test, suitability test and previous verifications. Requirements for the minimum number of stations to be tested every year are given in Table 11 (see Table 6 in GDE and Table 5 in EN 16450).

⁴ The number of sites per PM size fraction is defined by the relative expanded uncertainty, W_{AMS} , of the AMS, see Table 11.

⁵ A valid data pair consists of a valid reference value and a valid AMS average, both covering the same 24-hour period.

Table 11. Minimum requirement for number of stations for ongoing verification of AMS (dependent on expanded relative uncertainty, W_{AMS} , of AMS).

W_{AMS}	$\leq 10\%$	$>10\%$ to $\leq 15\%$	$> 15\%$ to $\leq 20\%$	$> 20\%$ to $\leq 25\%$
% of sites for ongoing equivalence*	10 %	10 %	15 %	20 %
Number of sites for ongoing equivalence*	2	3	4	5

*The smaller of the two resulting numbers may be applied. The minimum number of ongoing equivalence test sites is 2 for each type of AMS.

Example – Relative expanded uncertainty, $W_{TEOM\ DF}$, of TEOM 1405 DF from $PM_{2.5}$ comparison in 2015-2016 was 10.6 %. This is between 10 % and 15 %. TEOM 1405 DF is applied at 11 sites in Norway, i.e., 10 % of the number of sites equals 1.1. According to Table 11, ongoing verification should be performed at 1 site (resulting from upper row in Table 11) or 3 sites (resulting from lower row in Table 11). The smaller of the two resulting numbers may be applied, i.e., 1 site. However, the minimum number of ongoing equivalence test sites for each type of AMS is 2, thus ongoing equivalence tests of TEOM 1405 DF should be performed at 2 sites.

One of the test sites may be the site of the initial suitability test. The other sites should be different from that and should change from year to year to increase the coverage of the monitoring network. The sites shall be representative of all conditions where the AMS are operated (EN 16450, chapter 8.6.2), i.e., different environments should be tested year by year.

The test results should be evaluated every year using data accumulated over the previous 3-year period using the procedure in EN 16450, chapter 7.5.7. If the expanded relative uncertainty falls into a different category, the number of required test sites changes accordingly for the coming year (EN 16450, chapter 8.6.3).

AMS with a calculated **uncertainty larger than 25 %** need to be recalibrated applying the method described in EN 16450, chapter 7.5.8.5 (correction of slope and/or intercept).

The results from the 2015-2018 intercomparisons and Table 11 are used to decide the initial number of stations needed for ongoing verification in Norway.

6.1.3 AMS in Norwegian air quality monitoring networks

Particulate matter (PM) is measured at 57⁶ measurement stations in Norway (see Table 12). At most locations, both PM_{10} and $PM_{2.5}$ are measured. At five of the locations, only PM_{10} is measured, so there is a total of 109 PM sampling points⁷. Most of the locations are roadside locations (41 sites), 14 are urban background locations and 2 are industry related locations. Automated measuring systems (AMS) are applied at all sites and measured PM concentrations are stored as hourly averages and provided to the public via web portals (e.g. <https://luftkvalitet.miljodirektoratet.no/> and <https://luftkvalitet.nilu.no>). Online near real time data is necessary, both to be able to inform the public and as a tool for local municipalities, e.g., when deciding on measures to reduce/avoid episodes of high concentrations of suspended particulate matter. The Norwegian standard NS-EN 16450:2017 (“Ambient air – Automated measuring systems for the measurement of the concentration of particulate matter (PM_{10} ; $PM_{2.5}$)”) sets the requirements for automated measuring systems and

⁶ This covers only measurement sites characterised as “roadside”, “urban background” and “industry related”. PM measurements at regional background sites (Birkenes, Hurdal, Kårvatn) use the reference method. The number of stations and instrument types used in this report reflect the status in April 2021 and can be subject to changes.

⁷ PM_{10} and $PM_{2.5}$ measured at the same site count as two separate sampling points (CAFÉ directive, Annex V, A).

describes the method for equivalence testing (see Sections 6.1.1 and 6.1.2). The standard is based on chapter 9 in the GDE (2010; |1|).

A total of 67 automated measuring systems⁸ (analysers) are in operation in Norway. 51 % of the analysers use the TEOM method. About two thirds of the TEOM analysers are the older TEOM 1400 A and TEOM 1400 AB models, while one third is the newer TEOM 1405 DF model (measuring both PM size fractions concurrently). 24 % of the analysers are Grimm EDM 180 and an almost equal share of 22 % are Palas Fidas 200 models. Both analysers measure PM₁₀ and PM_{2.5} concurrently. TEI FH62-IR is in operation at one station (Skøyen in Oslo), where two units (3 % of the analysers) are installed measuring PM₁₀ and PM_{2.5}, respectively. Table 12 lists measurement sites with PM analysers in Norway. Fixed measurements using the reference method are performed only at three regional background sites in Norway (Birkenes, Hurdal and Kårvatn).

All PM analysers in use in Norway have been successfully tested for equivalence by, e.g., TÜV and others. The analyser models TEOM (1400 A, 1400 AB and 1405 DF), Grimm EDM 180 and Palas Fidas 200 are in use all over Norway, see Figure 48. Even though all AMS used in Norway have been approved for equivalence in Norway, comparison measurements carried out in 2015, 2016 and 2018 in Oslo indicate the need of calibration for most instrument types in use (see Chapter 4).

⁸ *Automated measuring system* is the term used in EN 16450. Here, we mainly use the term *analyser*.

Table 12. AMS in use at measurement sites in Norway. Status April 2021.

Measurement station	Station type	PM ₁₀	PM _{2.5}	AMS method PM ₁₀	AMS method PM _{2.5}	
Tromsø – Hansjordnesbukta	traffic	PM10	PM2.5	TEOM 1400 AB	TEOM 1400 AB	
Tromsø – Rambergan	urban background	PM10		TEOM 1400 AB		
Harstad – Seljestad RV83	traffic	PM10	PM2.5	EDM 180		
Narvik – Sentrum	traffic	PM10	PM2.5	Fidas 200		
Bodø – Olav V gate	traffic	PM10	PM2.5	Fidas 200		
Mo i Rana – Moheia Vest	industrial site	PM10	PM2.5	TEOM 1400 A	TEOM 1400 A	
Levanger – Kirkegata	traffic	PM10	PM2.5	TEOM 1400 AB	TEOM 1400 AB	
Trondheim – E6 Tiller	traffic	PM10	PM2.5	TEOM 1405 DF		
Trondheim – Elgeseter 1	traffic	PM10	PM2.5	TEOM 1405 DF		
Trondheim – Omkjøringsvegen	traffic	PM10	PM2.5	TEOM 1400 AB	TEOM 1400 AB	
Trondheim – Torvet	urban background	PM10	PM2.5	EDM 180		
Trondheim – Åsveien skole	traffic	PM10	PM2.5	TEOM 1405 DF		
Ålesund – Grimmerhaugen	urban background	PM10		TEOM 1400 AB		
Ålesund – Karl Eriksens plass	traffic	PM10		TEOM 1400 AB		
Bergen – Danmarks plass	traffic	PM10	PM2.5	Fidas 200		
Bergen – Klosterhaugen	urban background	PM10	PM2.5	Fidas 200		
Bergen – Rolland, Åsane	urban background	PM10	PM2.5	Fidas 200		
Bergen – Loddefjord	traffic	PM10	PM2.5	Fidas 200		
Bergen – Rådalen	traffic	PM10	PM2.5	Fidas 200		
Stavanger – Kannik	traffic	PM10	PM2.5	EDM 180		
Stavanger – Schancheholen	traffic	PM10	PM2.5	EDM 180		
Stavanger – Vålend	urban background	PM10	PM2.5	Fidas 200		
Lillehammer – Bankplassen	traffic	PM10	PM2.5	TEOM 1400 A	TEOM 1400 A	
Lillehammer – Barnehage	urban background	PM10	PM2.5	TEOM 1400 A	TEOM 1400 A	
Brumunddal – Ringsakervegen	traffic	PM10	PM2.5	EDM 180		
Gjøvik – Minnesundvegen	traffic	PM10	PM2.5	TEOM 1405 DF FDMS		
Hamar – Vangsveien	traffic	PM10	PM2.5	TEOM 1405 DF		
Elverum – Leiret	traffic	PM10	PM2.5	Fidas 200		
Oslo – Alnabru	traffic	PM10	PM2.5	TEOM 1405 DF		
Oslo – Bryn skole	urban background	PM10	PM2.5	EDM 180		
Oslo – Bygdøy Alle	traffic	PM10	PM2.5	EDM 180		
Oslo – E6 Alna Senter	traffic	PM10	PM2.5	EDM 180		
Oslo – Hjortnes	traffic	PM10	PM2.5	TEOM 1400 AB	TEOM 1400 AB	
Oslo – Kirkeveien	traffic	PM10	PM2.5	TEOM 1405 DF		
Oslo – Manglerud	traffic	PM10	PM2.5	TEOM 1400 AB	TEOM 1400 AB	
Oslo – RV4 Aker Sykehus	traffic	PM10	PM2.5	TEOM 1405 DF		
Oslo – Skøyen	urban background	PM10	PM2.5	FH 62 I-R	FH 62 I-R	
Oslo – Smestad	traffic	PM10	PM2.5	TEOM 1405 DF		
Oslo – Sofienbergparken	urban background	PM10	PM2.5	TEOM 1405 DF		
Bærum – Bekkestua	traffic	PM10	PM2.5	EDM 180		
Bærum – Eilif Dues Vei	traffic	PM10	PM2.5	EDM 180		
Drammen – Backeparken	urban background	PM10	PM2.5	Fidas 200		
Drammen – Bangeløkka	traffic	PM10		TEOM 1400 A		
Drammen – Vårveien	traffic	PM10	PM2.5	Fidas 200		
Lørenskog – Solheim	traffic	PM10	PM2.5	EDM 180		
Lillestrøm – Vigernes	traffic	PM10	PM2.5	TEOM 1405 DF		
Moss – Kransen	traffic	PM10	PM2.5	EDM 180		
Sarpsborg – Alvim	traffic	PM10	PM2.5	EDM 180		
Fredrikstad – Nygaardsgata	urban background	PM10	PM2.5	Fidas 200S		
Fredrikstad – St. Croix	traffic	PM10	PM2.5	EDM 180		
Tønsberg – Nedre Langgate	traffic	PM10	PM2.5	Fidas 200		
Grenland – Furulund	industrial site	PM10	PM2.5	Fidas 200		
Grenland – Knarrdalstranda	urban background	PM10	PM2.5	Fidas 200		
Grenland – Lensmannsdalen	traffic	PM10	PM2.5	TEOM 1400 AB	TEOM 1400 AB	
Grenland – Sverresgate	traffic	PM10		TEOM 1400 AB		
Kristiansand – Bjørndalssletta	traffic	PM10	PM2.5	EDM 180		
Kristiansand – Stener Heyerdahl	urban background	PM10	PM2.5	EDM 180		
Number of sampling points		57		52	Number of instrument	67
traffic		41		38	traffic	48
urban background		14		12	urban background	16
industrial site		2		2	industrial site	3



Figure 48. Geographical distribution of PM analyser models in Norway: Green: TEOM, red: Grimm EDM 180, blue: Palas Fidas 200, yellow: TEI FH 62 I-R. See also App. C.

In most cities having more than two monitoring stations, there is a mix of analyser models. The exception is Bergen, where Fidas 200 is in use at all 5 monitoring stations. At the two stations in Fredrikstad, different analyser models are used (this is not resolved in Figure 48). FH 62 I-R is in use in Oslo only (two instruments at the same station).

Comparison campaigns for PM₁₀ and PM_{2.5} analysers against the reference method were performed in 2015/2016 and 2018 at three locations in Oslo (two roadside sites and one urban background site). The comparisons were performed during different seasons. The results from the comparisons are presented in Chapter 4 of this report. All PM analyser types in use in Norway were included in the comparison exercise: TEOM 1400 AB, TEOM 1405 DF, Grimm EDM 180, Palas Fidas 200 and TEI FH 62 I-R. TEOM 1405 DF, Grimm EDM 180 and Palas Fidas 200 measure PM₁₀ and PM_{2.5} concurrently (see Table 12). When using TEOM 1400 and TEI FH 62-IR, two analysers are required to measure both PM size fractions simultaneously.

Norway has so far not performed yearly verification of the PM analysers at the monitoring sites as required by EN 16450 (see also Section 6.1.3). Results from comparison campaigns and observations in other countries suggest the need for testing equivalence periodically at various sites in the national measurement network. Significant differences of measurement performance were observed in the present comparison between the different analyser types used in Norway. Observations from comparison campaigns at different station types (e.g., roadside, industry related) in, e.g., Sweden, Austria and UK, show significant differences when measuring different types of aerosols. Automated PM analysers are optimised for measuring “city aerosol”, i.e., mainly related to traffic. Hence, it is important to test for equivalence at different locations, different seasons and, if necessary, correct measurement data by applying a calibration function (see Section 6.3.2).

The Excel workbook “Equivalence Tool, version 10” (Orthogonal regression and equivalence test utility), [7], developed by RIVM and used by AQUILA members was used in the first part of this report to compare the candidates with the reference method and to calculate calibration functions for the candidate methods. By applying orthogonal regression, the workbook calculates slope and intercept of the calibration function. The workbook runs a test for equivalence by calculating the *expanded relative uncertainty* for both raw and calibrated candidate data. The **uncertainty for PM₁₀ and PM_{2.5} shall be less than or equal to 25 %** as required by the CAFÉ directive (Annex I, Data quality objectives). There is a newer spreadsheet which may be used in the future, [8].

6.2 Status in Norway and some selected countries

In Norway, only PM₁₀ data from TEOM 1400 are corrected today. Based on a comparison carried out in Norway in 2001-2002, [9], TEOM 1400 data is multiplied by 1.1 and TEI FH 62 I-R data is multiplied by 1.0 (that is, no calibration). However, the PM comparison in 2015-2018 showed the necessity to correct data from optical methods as well.

Up to now, PM_{2.5} data from AMS have not been corrected in Norway. According to the comparison results in chapter 4.3, PM_{2.5} data may be underestimated or overestimated in different seasons and at different sites. This finding supports the need for calibration of PM data, in order to reflect the actual PM_{2.5} concentration in Norwegian cities.

PM comparisons have been performed in most European countries. How the results from the comparisons are applied, varies between the countries. A short overview highlighting procedures and observations for some countries – Sweden, Denmark, UK and Austria – is given in this section.

There is a lack of knowledge regarding how the performance of the measurement methods is affected by the different climatic conditions (and other local/regional factors) in the different parts of Norway. All comparisons so far have been organised in and around Oslo, where large variability was found, both

for location and season. It is likely that variability between locations and between seasons also is found in other parts of Norway. Comparisons in other European countries indicate both site and season dependent measurement performance.

Sweden

The Swedish monitoring network has over hundred PM measurement points, ca 60 % are traffic sites, 40 % are urban and regional background sites. NRL in Sweden runs a verification of suitability of AMS every year at one or several sites, moving up to 3 reference instruments (Derenda PNS 18T and Leckel SEQ 47/50) to other sites every year. Verification campaigns have been performed at several, mainly polluted, locations since 2012: Stockholm (2016, 2018, 2019, roadside), Sundsvall (2020), Västerås (2018, roadside), Umeå (2017), Norunda (2019/20, regional background). Reports from the ongoing verifications can be downloaded from the Swedish NRL's web page (<https://www.aces.su.se/reflab/rappporter/>). The verifications were mostly performed during winter and spring, when the highest PM concentrations are expected. A few comparison campaigns expanded into the summer season and a few were carried out in autumn. The Swedish reference laboratory indicated not to have enough capacity to achieve the 80 daily samples per year required by EN 16450.

Instrument types in use in the Swedish monitoring network are TEOM (older types: TEOM 1400 AB, TEOM 1400 A), Fidas 200, Grimm and SM200, distributed rather evenly over the country. Grimm EDM 180 instruments are not distributed over the entire country, but are mainly used in Stockholm, moreover in a little town and two in regional background.

For the comparisons, several instrument types, one of each type, are gathered at the same place to verify how the different instrument types measure in different environments. Comparisons are carried out at measurement sites of the monitoring network and other sites. During the comparisons, the instrument type used at the station is compared to the reference method and in some occasions also other instrument types are included to test their performance in the given environment. The variation in results shows that site-specific conditions have a major impact on the instrument performance. Different instruments perform differently in different environments.

Naturvårdsverket (Swedish Environmental Protection Agency) has decided⁹ that the intake head on instruments measuring PM₁₀ in Sweden must be the standardised American model (US inlet). In intercomparisons, the instrument performance is tested using different particle inlets (US EPA, EU, TSP). Also other conditions influencing the results have been tested in comparison campaigns: Naturvårdsverket's approval of TEOM 1400 AB for measuring PM₁₀ is based on data being corrected for losses of volatile particles (Volatile Correction Model, VCM). Performance tests with and without VCM correction show that the instrument's performance was better in the case of VCM correction, however not good enough to meet the data quality requirements. After calibration, the result was consistent both with and without VCM. Further calibration needs for Sweden's TEOM 1400 AB measurements need to be studied based on additional parallel measurements to obtain robust calibration factors.

The comparisons indicated that Fidas and SM200 results did not need to be calibrated. Certain instruments, e.g., Grimm, have problems and do not meet the requirements for equivalence with the reference method, not even after calibration. The Swedish reference laboratory is in dialogue with Grimm, discussing some theories regarding the found discrepancies. A big concern is that Grimm EDM 180 did not perform consistently, over- and underestimating PM₁₀-results at the same sites in different periods or over- and underestimating PM₁₀-results at different sites in the same period. It is now reviewed whether the instrument is suitable for continuous measurement of PM₁₀ in Sweden. Grimm PM₁₀ measurements remain uncorrected, Grimm PM_{2.5} measurements are corrected.

⁹ http://www.aces.su.se/reflab/wp-content/uploads/NV_beslut_PM10insug.pdf

Denmark

The Danish monitoring network comprises 11 measurement stations with PM sampling, consisting of five traffic stations, four urban background stations and two regional background stations. In Denmark, the reference method is used at all PM measurement locations (16 measurement points¹⁰ at the 11 sites), avoiding the need for ongoing verifications. At selected stations, PM is measured using an automated measuring system (TEOM), in addition, with a time resolution of 30 minutes, making it possible to resolve the diurnal variation of PM concentrations. These data measured with equivalent instruments are not reported and are solely used for real time information of the public. Only data obtained using the reference method are reported and used to control compliance with limit values.

Denmark introduced reference samplers at its PM measurement stations in 2012 as a consequence of problems experienced with beta gauge monitors used in the network before. The beta gauge monitors (Opsis SM200) also collected particles on filters which could be analysed, but showed problems regarding flow control and stability. The PM data were online accessible. Today's measurements, using reference samplers, proved to be more consistent. Ten years ago, differences in the range 25-30 % were observed between reference method and AMS. Now, the difference is smaller, the levels measured by the TEOM monitors correspond to those from the reference samplers or are within 10 % difference. Seasonally varying differences have not been studied. PM_{2.5} is measured at nine of the eleven sites using the reference method, at five sites (three traffic sites, one urban background site, one regional background site) both size fractions are measured.

In addition to the 16 measurement points using PM samplers, TEOM monitors are used at four of the measurement points (three at urban traffic stations, one at a regional background site) to report PM₁₀-data (also PM_{2.5}-data at one of the sites) to the public.

The Department of Environmental Science at Aarhus University has been responsible for the Danish air quality monitoring programme for over 30 years¹¹. They also hold the role of the National Reference Laboratory. The department participates in international intercomparisons in order to ensure and document measurements with high quality.

United Kingdom

All of the PM monitoring sites in UK have automated analysers. Similar to the situation in Norway, several different measurement techniques are used at the UK sites and are distributed all over the country. The instrument types used are approved for equivalence in the UK: Fidas (optical technique), Beta-Attenuation Monitor (BAM), Filter Dynamic Measurement System (FDMS) and Partisol (non-reference gravimetric sampler that collects daily samples onto a filter for subsequent weighing). There are ca. 76 measurement points for PM₁₀ and 78 measurement points for PM_{2.5} in the UK.

At two monitoring stations, permanent and continuous comparisons to the reference method are carried out. The stations are equipped with all AMS used in the UK (FDMS, Fidas, BAM) and one reference sampler per size fraction, measuring both size fractions (PM_{2.5} and PM₁₀) throughout the year. A lot of variability is observed, however it was chosen not to take further action regarding calibration of measurement data. The original results from type testing are used to calibrate the AMS.

For one of the sites, comparison data is available for a time span of 6-7 years. It is observed that the comparison result has never been the same as that from type testing. Fidas PM₁₀ data were in perfect agreement with the reference method under type testing, but not in the ongoing test. For BAM, the agreement is worse and the relationship is different for different environments (e.g., traffic sites,

¹⁰ <https://envs.au.dk/faglige-omraader/luftforurening-udledning-og-effekter/overvaagningsprogrammet/maalestationer/maalingepaastationerne/>

¹¹ <https://envs.au.dk/en/research-areas/air-pollution-emissions-and-effects/the-monitoring-program/>

industrial sites; close to steel works). The reason is supposed to be different composition of particles at different places. The relationships found between the different AMS and the reference method are not constant with time.

In UK, monitoring stations for comparisons are put at places with large range of (daily average) PM-concentrations.

Austria

In Austria, the individual federal states are responsible for performing ongoing equivalence tests of AMS in their networks, while the Environment Agency (Umweltbundesamt, UBA) is responsible for the network of 7 regional background stations¹². At the background measurement points, operated by UBA, measurements are permanently carried out with gravimetric and continuous methods in parallel, allowing a continuous test of the AMS with the gravimetric reference method.

Measurement stations in cities and close to industries are operated by the federal states. The data from all Austrian measurement stations are collected centrally by UBA and made available to the public. There are nine measurement networks in Austria (Burgenland, Kärnten, Niederösterreich, Oberösterreich, Salzburg, Steiermark, Tirol, Vorarlberg, Wien). Several different AMS types are used in Austrian networks and distributed over Austria: FH 62 I-R, Sharp 5030 (combines light scattering photometry and beta attenuation), TEOM FDMS 1400 (not 1405), MetOne BAM, Grimm EDM 180 (there is no Fidas in Austrian networks). Within most of the individual federal networks, one AMS type is used consistently for both PM₁₀ and PM_{2.5}, e.g., Sharp 5030 in Kärnten, Grimm EDM 180 in Oberösterreich, FH62I-R in Tirol. This may simplify the organisation of ongoing intercomparisons. In a few networks (e.g., Niederösterreich, Burgenland, Steiermark), at least two different AMS types are in use.

All measurement stations in Austria have a space reserved for a high volume sampler. Digital high volume samplers (Digital DHA-80) are used as reference samplers for testing of ongoing equivalence. The federal states independently carry out the comparisons/equivalence tests in their networks (there are two states which cooperate). The comparison exercises are carried out regularly at varying stations within each network. Critical sites, with high concentrations, are checked more often, background sites less often. It takes several years to carry out intercomparisons covering the most important environments. In each network, there is one sampler which always is located at the same station. Results from that site are extrapolated to other sites. In Vienna, which is a large city, it is critical with exceedances. At more than 50 % of the stations in Vienna, parallel measurements are carried out. Some of them sample every year, some change from year to year.

The Austrian PM monitoring network consists of ca. 130 measurement points for PM₁₀. In 2019, the gravimetric method was used in parallel with an AMS at ca. 31 measurement points (ca. 24 %) for ongoing equivalence tests. At ca. 93 of the stations, PM₁₀ was measured with equivalence-tested continuous methods (AMS) only. The five stations in Vorarlberg were only equipped with gravimetric samplers. At those measurement points with parallel gravimetric and continuous data, continuous (AMS) data is used to inform the public and is subject to equivalence tests, and gravimetric measurement data are used to assess compliance with respect to PM₁₀ limit values.

Calibration functions are derived from the ongoing equivalence tests. The calibration function derived from the previous year's data is used to calibrate the incoming data of the present year as a first approach. In the end of the year, the calibration function derived for the full year is applied to the raw data backward.

The Austrian measurement network operators and the Environment Agency organised an equivalence test for continuous PM₁₀ and PM_{2.5} monitors for the first time between December 2007 and August

¹² <https://www.umweltbundesamt.at/umweltthemen/luft/messnetz>

2008. In this exercise, the equivalence of the AMS types was assessed and calibration functions for the different AMS types and locations were determined. Since then, regular equivalence tests have been carried out at various measuring points.

From November 2017 to March 2018, the Environment Agency organised for the first time an intercomparison for the gravimetric determination of PM₁₀ and PM_{2.5}, respectively, in Steyregg in Oberösterreich, as proof of the competence of the Austrian measurement network operators. Further intercomparisons on the gravimetric determination of PM₁₀ took place between January and March 2019 and on the determination of PM_{2.5} between January and March 2020 in Graz.

Austria has a decentralised organisational structure. The individual federal states generally have more responsibility than administrative units in Norway, thus each state is responsible for performing ongoing equivalence tests. Some states have outsourced the comparison measurements to universities. The networks cover the extra cost for carrying out the ongoing equivalence tests. For Norway, it is proposed to organise the equivalence tests in a similar way as in Austria, however centrally coordinated through the reference laboratory.

Lack of knowledge

The intercomparison results obtained so far may only apply to Oslo, or only to the specific sites in Oslo where the comparisons were performed. No comparisons have been carried out in other cities or other parts of Norway. The calibration functions obtained from comparisons carried out in Oslo are not necessarily valid in other cities or at other measurement stations, especially if measurement stations are affected by different PM-sources (e.g., close to industries, near the coast, etc.).

The results from the comparison campaigns at Hjortneskaia, Sofienbergparken and Smestad suggest that the calibration functions may vary dependent on site and season. However, since the comparison was not repeated, e.g., the year after, at any of the sites, it is not possible to conclude that the calibration functions are consistent over time and location. Results from an intercomparison at one site in the UK over several years, for example, showed high variability in the calibration functions at that site for the same season in different years. Results from several sites in Austria, however, show a clear seasonal variation in the calibration functions.

Meteorological data (wind speed, temperature, relative humidity and pressure as 24 hour averages) are available from the measurement sites for the time periods of PM comparison, however there is not enough data to conclude on how meteorological parameters affect the different measurement methods. There was a large degree of correlation of meteorological parameters with the same parameters measured at Blindern, except from wind which is usually affected by local conditions. It is important to register meteorological data at measurement sites for ongoing verification of equivalence in order to identify possible relationships. For future verification experiments it should be evaluated if there is an official meteorological station (measuring wind speed, wind direction, relative humidity, precipitation) which is representative for the site or if it is necessary to measure meteorological parameters at the site. Usually, it is necessary to measure at the site, since wind is influenced locally.

Observations on PM measurements are available from other campaigns including analysers measuring parallel to reference samplers. Measurements in Oslo indicated that Grimm EDM 180 underestimates mass concentration for coarse particles (diameters larger than 2.5 µm). Such observations need to be analysed further. Also possible challenges for optical measurement methods measuring contributions from sea salt or close to metallurgical industries and during episodes need to be analysed further.

6.3 Suggested system for Norway

In the following subsections, it is suggested, how a future system in Norway for ongoing verification of PM-analysers may be set up. This involves both the decision of the number of sites to be tested for ongoing verification, the selection of these stations, the criteria how to select stations and how stations/networks may be grouped. Further, it is described how to calibrate the raw data and how the system for ongoing verification may be organised in Norway.

6.3.1 Selection of sites

NS-EN 16450 requires a number of AMS to be tested every year for verification of the calibration functions. The exact number of AMS depends on the measurement uncertainty, W_{AMS} , calculated for each PM fraction and instrument type based on the most recent intercomparison results, see Table 11.

Example – Fidas 200 comparison

In the PM_{10} intercomparison, the expanded relative uncertainties, $W_{Fidas\ PM_{10}}$, after calibration of Fidas1 and Fidas2 data were 16.4 % and 15.6 %, respectively. Being conservative, using the highest value of the two, the uncertainty is between 15 % and 20 %. The number of sites to be verified for equivalence is either 15 % of the sites or 4 sites. Fidas 200 is running at 15 measurement sites, i.e., 15 % equals 2 sites. According to Table 11, the smaller number (of 4 and 2) may be applied. The minimum number of sites running Fidas 200 analysers to be verified every for equivalence for PM_{10} is 2.

In the $PM_{2.5}$ intercomparison, the expanded relative uncertainties, $W_{Fidas\ PM_{2.5}}$, after calibration of Fidas1 and Fidas2 data were 19.2 % and 18.3 %, respectively. Using the same method as for PM_{10} , the minimum number of sites running Fidas 200 analysers for measuring $PM_{2.5}$ to be verified every year for equivalence is also 2.

Based on the results from the 2015-2018 intercomparisons, yearly verification of equivalence should be performed at 3 EDM 180 sites for PM_{10} (2 for $PM_{2.5}$), 2 TEOM 1400 sites for $PM_{2.5}$ (PM_{10} was not tested), 2 TEOM 1405 DF sites and zero FH62-IR sites. There is only one station running FH62-IR, requiring a different verification scheme. As there is a tendency to phase out FH62-IR in the Norwegian monitoring network, no equivalence test is planned for that instrument type.

Table 13. Number of sites in the entire Norwegian measurement network to be verified based on results from the 2015-2018 PM comparisons (valid for the first year of ongoing verification). Numbers have to be assessed every year.

	Fidas 200	EDM 180	TEOM 1400	TEOM 1405 DF	FH 62 I-R
PM₁₀					
Number of analysers	15	16	14	11	1
Measurement uncertainty W	16.4 %	26.2 %	-**	10.0 %	9.5 %
Number of sites for ongoing verification*	2 / 4	3 / 5	2	1*** / 2	0*** / 2
PM_{2.5}					
Number of analysers	15	16	9	11	1
Measurement uncertainty W	19.0 %	17.2 %	16.0 %	10.6 %	22.7 %
Number of sites for ongoing verification*	2 / 4	2 / 4	1*** / 4	1*** / 3	0*** / 5

*The first number is based on % of sites, the second number is based on number of sites in Table 11. (The smaller of the two numbers is applied).

**No PM₁₀ intercomparison was carried out for TEOM 1400

***The minimum number of on-going equivalence tests is 2 for each type of analyser.

In total, 9 PM₁₀ sites and 8 PM_{2.5} sites should be tested for equivalence in Norway in a first verification exercise. The number of sites for future ongoing verification will in each case be based on the results of the most recent verification tests. The required number of sites indicated above is, of course, a minimum number. Equivalence testing 9 out of 57 PM₁₀ sites and 8 out of 52 PM_{2.5} sites corresponds to 15.8 % and 15.4 % of all PM measurement sites, respectively. Using the minimum number of sites, it will take several years to acquire calibration functions for the entire network. In Austria, where the system for online verification of equivalence has been established over 10 years ago, yearly verification tests are carried out at ca. 24 % of all PM₁₀ sites.

The ongoing verification of PM₁₀ and PM_{2.5} analysers may be realised in Norway by dividing the national network into a number of regions within which reference samplers are shared. The definition of the suggested regions is given below. They are given by regional clusters of measurement stations as shown in Figure 49.

The Norwegian measurement network extends 1400 km between 58°N and 69°N. The distances between individual city networks can be quite large. It may be too costly for each individual city to rent or purchase its own reference sampler(s). It is recommended to split the measurement network into geographical regions – within each of the regions, the cities will share a number of reference samplers. The reference samplers will alternate between the measurement sites within the region on a yearly basis according to results summarised in Table 13 for the first period and according to updated comparison results in future periods (see chapter 6.3.4).

It was considered to let the regions equal the air quality zones defined in Forurensningsforskriften, Chapter 7, Annex 1 (https://lovdata.no/dokument/SF/forskrift/2004-06-01-931/KAPITTEL_3-

[1#KAPITTEL 3-1](#)), however, since these zones are rather large and cover very different climatic conditions, another regional division was considered more appropriate, see Figure 49. The city networks included in each of the clusters/regions also experience comparable climatic conditions. The individual regions contain a similar number of measurement stations (5-8 stations per region, however, 11 stations in Oslo). This ensures that equivalence testing will be carried out evenly over the entire Norwegian network.

- 1) Region “North” covers 6 measurement stations in Troms and Nordland (Tromsø, Harstad, Narvik, Bodø, Mo i Rana), forming the northernmost cluster of stations in Norway. The most-used instrument type in this region is TEOM 1400 (5 units), followed by Fidas 200 (2 units) and EDM 180 (1 unit).
- 2) Region “Trøndelag and Ålesund” covers 8 measurement stations in Trøndelag and Møre og Romsdal (Levanger, Trondheim, Ålesund) along the coast in the middle of Norway. The most used instrument type in this region is TEOM 1400 (6 units), followed by TEOM 1405 DF (3 units in Trondheim) and EDM 180 (1 unit).
- 3) Region “Bergen and Stavanger” covers 8 measurement stations in Bergen and Stavanger. In this region, exclusively instrument types applying optical detection principles are in use: Fidas 200 (6 units), EDM 180 (2 units).
- 4) Region “Innlandet” covers 6 monitoring stations located in a cluster of cities in Innlandet (Lillehammer, Brumunddal, Gjøvik, Hamar, Elverum). The most used instrument type in this region is TEOM 1400 (4 units), followed by TEOM 1405 DF (2 units), EDM 180 (1 unit) and Fidas 200 (1 unit).
- 5) Region “Oslo” concentrates on 11 measurement stations in the city of Oslo. The most used instrument type in this region is TEOM 1405 DF (5 units), followed by TEOM 1400 (4 units), EDM 180 (3 units) and FH62-IR (2 units).
- 6) Region “Viken” covers 7 measurement stations in Viken, surrounding Oslo (Bærum, Drammen, Lørenskog, Lillestrøm). The most used instrument type in this region is EDM 180 (3 units), followed by Fidas 200 (2 units), TEOM 1405 DF (1 unit) and TEOM 1400 (1 unit).
- 7) Region “Oslofjord” covers 5 measurement stations east and west of Oslo fjord (Moss, Sarpsborg, Fredrikstad, Tønsberg). In this region, exclusively instrument types applying optical detection principles are in use: EDM 180 (3 units), Fidas 200 (2 units).
- 8) Region “Grenland and Kristiansand” covers 6 measurement stations in Grenland and Kristiansand. The most used instrument type in this region is TEOM 1400 (3 units), followed by EDM 180 (2 units) and Fidas 200 (2 units).

Table 14 and Table 15 give an overview how the instrument types currently are distributed over Norway and the suggested regions. This distribution is used to find the most relevant sites to start ongoing verification with. Table 14 shows the *number of stations* running the different instrument types, the percentage of the instrument types per region, and how the instrument types are distributed over the regions. Table 15 shows the same for the *number of instrument units*. Since EDM 180, Fidas 200 and TEOM DF measure both size fractions with a single instrument, Table 14 and Table 15 only differ in the numbers shown for TEOM 1400. The tables give an indication at which stations to start equivalence testing. Since instruments are occasionally exchanged by another instrument type, this overview shows a snapshot valid in April 2021.




Figure 49. Map over all measurement stations in Norway where PM_{10} and/or $PM_{2.5}$ are measured with AMS. Suggested grouping of stations into regions sharing reference samplers.

Table 14. Number of **stations** running the instrument types EDM 180, Fidas 200, FH62-IR, TEOM 1400 and TEOM 1405 DF in the Norwegian measurement network in total and in the suggested regions. Numbers in parentheses show the percentage of the instrument types per region (red) and how the instrument types are distributed over the regions, in percent (blue).

Type	Total	Reg 1	Reg 2	Reg 3	Reg 4	Reg 5	Reg 6	Reg 7	Reg 8
EDM 180	16	1 (17%) (6%)	1 (13%) (6%)	2 (25%) (13%)	1 (17%) (6%)	3 (27%) (19%)	3 (43%) (19%)	3 (60%) (19%)	2 (33%) (13%)
Fidas 200	15	2 (33%) (13%)	0 (0%) (0%)	6 (75%) (40%)	1 (17%) (7%)	0 (0%) (0%)	2 (29%) (13%)	2 (40%) (13%)	2 (33%) (13%)
FH62-IR	1	0 (0%) (0%)	0 (0%) (0%)	0 (0%) (0%)	0 (0%) (0%)	1 (9%) (100%)	0 (0%) (0%)	0 (0%) (0%)	0 (0%) (0%)
TEOM 1400	14	3 (50%) (21%)	4 (50%) (29%)	0 (0%) (0%)	2 (33%) (14%)	2 (18%) (14%)	1 (14%) (7%)	0 (0%) (0%)	2 (33%) (14%)
TEOM DF	11	0 (0%) (0%)	3 (38%) (27%)	0 (0%) (0%)	2 (33%) (18%)	5 (45%) (45%)	1 (14%) (9%)	0 (0%) (0%)	0 (0%) (0%)
Total	57	6	8	8	6	11	7	5	6

Table 15. Number of **instrument units** of the 5 instrument types running in the Norwegian measurement network in total and in the suggested regions. Numbers in parentheses show the percentage of the instrument types per region (red) and how the instrument types are distributed over the regions, in percent (blue).

Type	Total	Reg 1	Reg 2	Reg 3	Reg 4	Reg 5	Reg 6	Reg 7	Reg 8
EDM 180	16 (24%)	1 (13%) (6%)	1 (10%) (6%)	2 (25%) (13%)	1 (13%) (6%)	3 (21%) (19%)	3 (43%) (19%)	3 (60%) (19%)	2 (29%) (13%)
Fidas 200	15 (22%)	2 (25%) (13%)	0 (0%) (0%)	6 (75%) (40%)	1 (13%) (7%)	0 (0%) (0%)	2 (29%) (13%)	2 (40%) (13%)	2 (29%) (13%)
FH62-IR	2 (3%)	0 (0%) (0%)	0 (0%) (0%)	0 (0%) (0%)	0 (0%) (0%)	2 (14%) (100%)	0 (0%) (0%)	0 (0%) (0%)	0 (0%) (0%)
TEOM 1400	23 (34%)	5 (63%) (22%)	6 (60%) (26%)	0 (0%) (0%)	4 (50%) (17%)	4 (29%) (17%)	1 (14%) (4%)	0 (0%) (0%)	3 (43%) (13%)
TEOM DF	11 (16%)	0 (0%) (0%)	3 (30%) (27%)	0 (0%) (0%)	2 (25%) (18%)	5 (36%) (45%)	1 (14%) (9%)	0 (0%) (0%)	0 (0%) (0%)
Total	67	8	10	8	8	14	7	5	7

Table 13 above shows that the first ongoing verification should be performed at at least 9 PM₁₀ measurement sites (the 8 PM_{2.5} measurement sites may be covered by selecting the same stations): 3 EDM 180, 2 Fidas 200, 0 FH62-IR, 2 TEOM 1400 and 2 TEOM 1405 DF, distributed as follows.

- Region (1) «North»: Tromsø – Hansjordnesbukta (TEOM 1400)
- Region (2) «Trøndelag and Ålesund»: Trondheim – Elgeseter (TEOM DF)
- Region (3) «Bergen and Stavanger»: Bergen – Danmarks plass (Fidas 200)
- Region (4) «Innlandet»: Elverum – Leiret (Fidas 200)
- Region (5) «Oslo»: Bygdøy Allé (EDM 180) and Smestad (TEOM DF)
- Region (6) «Viken»: Bærum – Eilif Dues Vei (EDM 180)
- Region (7) «Oslofjord»: Sarpsborg – Alvim (EDM 180)
- Region (8) «Grenland and Kristiansand»: Grenland – Lensmannsdalen (TEOM 1400)

The suggested distribution for the first ongoing verification is also shown in Table 16. The nine sampling sites are equally distributed over seven almost equally sized¹³ regions. In Oslo, which has most measurement stations, two comparisons are planned. The distribution is based on the percentage of instrument units (cp. Table 15) of the individual instrument types that are in use in the different regions and preferably road side sites are chosen.

Table 16. Suggested distribution of measurement sites for the first ongoing verification for both PM_{10} and $PM_{2.5}$, based on the number of sites required for PM_{10} . The grey shaded areas indicate that the instrument type is not in use in the particular region.

Type	Total	Reg 1	Reg 2	Reg 3	Reg 4	Reg 5	Reg 6	Reg 7	Reg 8
		Tromsø Hansj.b.	Trondh. Elgeseter	Bergen Danm.p.	Elverum Leiret	Bygdøy Allé + Smestad	Bærum Eilif DV	Sarpsborg Alvim	Grenland Lensm.d.
EDM 180	3					1	1	1	
Fidas 200	2		-	1	1				
FH62- IR	0*	-	-	-	-	0	-	-	-
TEOM 1400	2	1		-				-	1
TEOM DF	2	-	1	-		1		-	-
Total	9	1	1	1	1	2	1	1	1

The verification can initially be performed according to Table 16. The reference sampler/s is/are shifted to other stations in subsequent years. In this way, new calibration functions will be established for more and more stations in the network on a regular basis.

¹³ The regions are equally sized regarding the number of monitoring stations/instruments.

In order to study and understand discrepancies between the instrument types, meteorological parameters (wind speed, wind direction, temperature, relative humidity, rain) should be measured at the sites selected for ongoing verification of equivalence.

6.3.2 Calibration method

It is suggested to calibrate PM data in a similar way as performed in Austria (see Section 6.2):

The calibration function derived from the previous ongoing equivalence test at the same site is used to calibrate the incoming data of the present year as a first approach. In the end of the year, the calibration function derived for the full year is applied to the raw data backward. That means that raw data are first calibrated forward using the previous calibration. Finally, in the end of the verification period, the raw data are calibrated backward using the new calibration. However, acknowledging the calibration function's dependence on location in Norway it is recommended to apply an individual calibration function to each site.

It is assumed that the relationship between the measurement results of the analyser and the reference method can be described by a linear relationship of the form:

$$y_i = a + bx_i$$

where y_i is the result of the analyser for an individual 24 h period and x_i is the (average) result of the reference method for the same 24 h period (Chapter 7.5.8.5 in [4]), a is the intercept of the calibration function and b is the slope of the calibration function.

In the first year, no calibration will be carried out while the verification test is ongoing, since there is no input data available. After the year of sample collection has passed and the calibration functions for the entire year are available, PM analyser data is calibrated backward according to the calibration functions found. It is recommended to calibrate the data with the calibration function based on data from the entire year, but also to calculate the seasonal calibration functions, in order to study the seasonal variability of the calibration functions at the different sites.

It will take several years to carry out comparisons covering the most important environments. After a few years, stations will be revisited for ongoing verification tests, i.e., a calibration function from the previous verification test will be available. During the first few years, there will be stations calibrating PM data alongside with stations not calibrating PM data. If equivalence testing is carried out at more than the minimum required stations as resulted from the comparisons in 2015, 2016 and 2018, a full coverage of the entire network will be reached earlier.

The calibration functions are derived by the reference laboratory, based on raw data from the PM analysers and results from the reference samplers, using a standardised calculation scheme recommended by AQUILA.

It is recommended to start verification testing mainly at traffic sites as those are most dependent on correct measurement data in order to verify if limit values are met and to decide the need and extent of air quality measures. As suggested in Section 6.3.1, verification tests will be carried out at least at one station per region. This ensures that verification testing is carried out evenly over the whole of Norway. In order to establish individual calibration functions at each site, verification testing must be performed at all sites. When a new site is established or a new type of analyser (of other manufacturer or measurement principle) is installed at an existing site a verification testing must be performed at the site.

6.3.3 Organisation of ongoing verification in Norway

Referring to Section 6.3.1, the ongoing verification of PM-analysers may be realised in Norway by dividing the national network into a number of regions/regional clusters with similar climatic conditions.

There are several possibilities regarding the ownership of the reference samplers to be used for ongoing verification. The instruments cost between 3500 € and 13 500 €¹⁴, depending if manual single-filter samplers or automatic sequential samplers are chosen to be used.

- The cities within the regional clusters may share the acquisition costs for the reference sampler(s) to be used in their region. They will have to agree how to share the costs and maintain the instruments. This may be too complicated, both for the administrations of the existing city networks, and also for the case that new city networks are added to the national network (and thus individual regions).
- All reference samplers used in the Norwegian network may be owned by Miljødirektoratet or the Norwegian reference laboratory. Maintenance of the instruments is carried out by the Reference laboratory. Every year, the city networks that are included in the current verification measurements, pay a rental fee to the reference laboratory, which covers necessary maintenance of the instruments and a percentage of the acquisition costs in order to keep the instrument park updated. A number of particle samplers (ca. (10 - 12) * 2) will have to be purchased before ongoing verification can start. This number needs to include 2-3 backup instruments, so that defect instruments can be replaced as soon as possible without failing the requirement for data coverage. The number needs to cover a slightly higher number of samplers possibly required for future verification tests than the current 9 (for PM₁₀) and 8 (for PM_{2.5}) required in the first year.

Regarding the type of sampler, there are two choices – the manual (single-filter) sampler and the sequential sampler. It is recommended to use one sampler per size fraction.

- The single-filter sampler has a lower purchasing price (see table below). It requires manually changing the filter after/before each sample, i.e., every fourth day. This involves frequent station visits and may be irksome over an entire year.
- The sequential sampler initially costs almost 4 times more than the manual sampler, but offers the convenience of having a storage for 14 filters and a mechanism for automatically changing the filters. For sampling every fourth day, station visits are required every 56th day (every second month).

The filter analysis costs are in all cases covered by the individual cities participating in the current verification. An overview of estimated costs in connection with ongoing verification tests, both investments and operational costs, is given in Table 17. Costs (in NOK) are shown for 1 year of measurements at a station including PM₁₀ and PM_{2.5} measurements and requiring travel by air plane. An ongoing verification in short distance from the Reference laboratory's current location, e.g., in Oslo, reduces the yearly cost by ca. NOK 48 000.

¹⁴ The costs given are estimates based on prices for the instrument paid in recent years.

Table 17: Overview of estimated costs for implementation of ongoing verification. Yearly costs include costs for one year PM_{10} and $PM_{2.5}$ comparison at one station.

	Investment costs (NOK)	Yearly costs (NOK)
Sequential reference sampler	278 600 (purchasing 2 samplers, PM_{10} and $PM_{2.5}$)	94 000 (rental from reference laboratory, 2 samplers)
Manual (single-filter) reference sampler	73 000 (purchasing 2 samplers, PM_{10} and $PM_{2.5}$)	24 000 (rental from reference laboratory, 2 samplers)
Preparation, freight costs, Installation/ deinstallation of samplers at a site, training of personnel, 3 monthly inspection, yearly service*		132 000
Gravimetric analysis of filters		60 000 (7-8 filters per month and size fraction, total 170, PM_{10} and $PM_{2.5}$)
Data analysis and reporting		39 000 (3 working days)
Total 1 year, 2 sequential samplers (PM_{10} and $PM_{2.5}$)		325 000
Total 1 year, 2 manual samplers (PM_{10} and $PM_{2.5}$)		256 000

*Preparation/inspection costs given here involve freight costs and travel by air plane. For stations located at shorter distance from the Reference laboratory's location, the yearly costs are reduced by ca. NOK 48 000.

The purchasing costs for the particle samplers are a one-off cost. Once all samplers are bought, they are rented out to the participating cities, for a monthly rent depending on the sampler type. Other costs that occur in connection with the equivalence testing campaigns are installation and deinstallation of the reference samplers in the selected cities, i.e., before and after the campaign, and quarterly maintenance visits. This involves freight costs, travelling costs and working hours for preparation and installation/deinstallation to be paid by the cities to the reference laboratory. Training of local personnel in operating the sampler and handling filters may also be carried out in connection with the installation of the equipment.

The selection of measurement stations for ongoing verification of equivalence may be organised by NRL (or: suggested by NRL in agreement with Miljødirektoratet), based on the results of the most recent verification tests (see also Section 6.3.1). The preparation and analysis of reference sampler filters may be subcontracted by the city to any laboratory that has the necessary accreditation.

At most measurement sites, both PM_{10} and $PM_{2.5}$ are measured. Based on the intercomparison exercise (2015, 2016, 2018), at least 9 PM_{10} sites and at least 8 $PM_{2.5}$ sites should be tested for equivalence in the first comparison exercise.

- Using two manual reference samplers per site (one for each size fraction) requires attendance at the site every 4 days for changing filters.
- Using sequential reference samplers will lead to higher rental or purchasing cost, but requires less attendance (filter magazine is replaced every 56 days, compared to visiting the station every 4 days for replacing filter in a manual sampler). If both PM₁₀ and PM_{2.5} are measured at the site, which is recommended, two sequential samplers are needed per site.

Similar to the Austrian system, a place in or on top of each measurement shelter may be reserved for the reference instrument(s) to allow the ongoing verification to cover most of the national measurement network after a few years. For those stations with very constricted indoor space, solutions will be found to place the reference samplers next to the station.

PM monitoring stations in Norway are owned and operated either by the individual municipalities or by the local sections of the National Road Administration. Before each round of ongoing verification, the technical representatives from the cities / SVV included in the verification exercise will obtain training by NRL in operation of the reference sampler. It is most convenient and most effective to train local personnel in connection with installation of the samplers.

Timely before a verification testing starts at the selected stations, the samplers are installed at the stations. They will have passed an annual maintenance and flow calibration before, so that they are in optimum conditions. It may be recommendable to carry out verification measurements timely staggered at the selected stations. Doing so will avoid overloading technical personnel at the reference laboratory, maintaining, preparing and installing instruments at several sites at the same time. Even though, verification measurements will not last from new year to new year (calendar year), they will cover 12 months.

6.3.4 Ongoing verification after the first year

After the first year, comparison results will be available for individual stations in all regions. The test results should be evaluated every year to fine the basis for the next year's verification measurements.

The number of sites for the next year's ongoing verification measurements depends on the relative expanded uncertainty for each analyser type from the most recent verification campaign. For every analyser type, comparison results will be available from two or more units. In order to derive the number of equivalence tests necessary for the following year (according to Table 11), comparison results from the unit with higher relative expanded uncertainty are used.

If the relative expanded uncertainty falls into a different category, the number of required test sites changes accordingly for the coming year.

The resulting number of sites should be selected in a way that different environments are tested year by year and that the sites equally cover the suggested regions in Norway. In order to gradually cover the entire Norwegian network, new sites should be selected every year. It is recommended to prefer road sites in the beginning, which are most critical regarding exceedance of limit values.

It is beneficial to carry out verification measurements for both size fractions, PM₁₀ and PM_{2.5}, at the same site in order to study how analysers perform compared to the reference method for fine (diameter smaller than 2.5 µm) or coarse (diameter larger than 2.5 µm) particles. Calibration functions are derived for each individual station, in a first attempt, and applied to the raw data as described in chapter 6.3.2. Doing so will possibly help to find similarities between calibration functions for stations in similar environments.

References

- [1]. EC Working Group on Guidance for the Demonstration of Equivalence (2010) Guide to demonstration of equivalence of ambient air monitoring methods (GDE),
<https://ec.europa.eu/environment/air/quality/legislation/pdf/equivalence.pdf>
- [2]. European Parliament, Council of the European Union (2008) Directive 2008/50/EC on ambient air quality and cleaner air for Europe,
<https://eur-lex.europa.eu/legal-content/en/ALL/?uri=CELEX%3A32008L0050>
- [3]. CEN, European Committee for Standardization (2014) EN 12341:2014 "Standard gravimetric measurement method for the determination of the PM₁₀ or PM_{2,5} mass concentration of suspended particulate matter",
https://standards.cen.eu/dyn/www/f?p=204:110:0:::FSP_PROJECT,FSP_ORG_ID:29133,6245&cs=1DC6EB16DD302E384B46A7097AAC67CB5
- [4]. CEN, European Committee for Standardization (2017) EN 16450:2017 "Automated measuring systems for the measurement of the concentration of particulate matter (PM₁₀; PM_{2,5})",
https://standards.cen.eu/dyn/www/f?p=204:110:0:::FSP_PROJECT,FSP_ORG_ID:39906,6245&cs=1FE7843133ED AFC8D16B56744004D9499
- [5]. Jari Waldén, Tuomas Waldén, Sisko Laurila, Hannele Hakola (2017) Demonstration of the equivalence of PM_{2,5} and PM₁₀ measurement methods in Kuopio 2014-2015. FMI report 2017:1,
<https://helda.helsinki.fi/bitstream/handle/10138/173933/Equivalence%20report.pdf?sequence=1&isAllowed=y>
- [6]. Hans Areskoug (2012) Equivalence of PM₁₀ Instruments at a Road Traffic Site. A Study in Stockholm Spring 2012. Stockholm University, Department of Environmental Science and Analytical Chemistry, Atmospheric Science Unit. ACES report 4,
https://www.aces.su.se/reflab/wp-content/uploads/2016/11/ACES_Report_4.pdf
- [7]. Ruben Beijck, Theo Hafkenschied, David Harrison (2018) Equivalence Tool v2.10, Orthogonal regression and equivalence test utility. RIVM, Dutch Institute for Public Health and the Environment, Dep. Centre for Environment Monitoring,
<https://ec.europa.eu/environment/air/quality/legislation/pdf/Equivalence%20Tool%20v10.xlsm>
- [8]. David Harrison (2020) Equivalence Tool v3.1,
<https://ec.europa.eu/environment/air/quality/legislation/pdf/Equivalence%20Tool%20V3.1%2020720.xlsx>
- [9]. Leif Marsteen, Jan Schaug (2007) A PM₁₀ intercomparison exercise in Norway. NILU OR 41/2007,
<https://hdl.handle.net/11250/2718568>

Appendix A

PM10 performance characteristics tables

Table 18. Performance characteristics of the PM₁₀ reference samplers RM2 and RM4, first and second comparison. Between RM uncertainty is based on first comparison only. When one sampler did not produce a valid result, the result of the other sampler would represent the average of the two samplers. RM4-6 is the single PM₁₀ reference sampler in the second comparison.

Test PM ₁₀	Criteria	Ref. method	Pass/Fail
Suitability of data	20 % > 28 µg/m ³	23 %	Pass
Between RM uncertainty			
All data	$U_{bs, RM} < 2.0 \mu\text{g}/\text{m}^3$	0.69 µg/m ³	Pass
> 30 µg/m ³	$U_{bs, RM} < 2.0 \mu\text{g}/\text{m}^3$	0.92 µg/m ³	Pass
Number of paired samples			
All data		149	
> 30 µg/m ³		30	
Number of RM2 samples		190	
Number of RM4 samples		151	
Number of RM4-6 samples		48	
Number of RM averages			
First comparison		191	
First and second comparison		239	
Average			
First comparison		18.4 µg/m ³	
First and second comparison		20.4 µg/m ³	

Table 19. Performance characteristics of the Palas Fidas 200 PM₁₀ candidates Fidas1 and Fidas2, first and second comparison, RSA, RSW, UBW, UBS, RSW2. Between CM uncertainty is based on first comparison only. Comments in red indicate significant deviations from performance criteria. Comments in blue indicate that calibration is necessary.

Test PM ₁₀	Criteria	Fidas1	Fidas2	Comment
Between CM uncertainty				
All data	$U_{bs,CM} < 2.5 \mu\text{g}/\text{m}^3$	0.38 $\mu\text{g}/\text{m}^3$		Pass
> 30 $\mu\text{g}/\text{m}^3$	$U_{bs,CM} < 2.5 \mu\text{g}/\text{m}^3$	0.63 $\mu\text{g}/\text{m}^3$		Pass
Comparability				
All data				
Number of values		218	234	
Data capture	90 %	91.2 %	97.9 %	Pass
Average	Ref. 20.88 20.27	20.77	20.13	
Slope, b	$0.98 < b < 1.02$	1.1141	1.1190	Calibrate
Intercept, a	$-1 < a < 1$	-2.4923	-2.5496	Calibrate
Expanded relative uncertainty	25 %	21.8 %	21.7 %	Pass
Calibrated data, $RM = a + b * CM$				
All data				
Slope, b		0.898	0.894	
Intercept, a		2.237	2.279	
Expanded relative uncertainty	25 %	16.4 %	15.6 %	Pass
All data, use intercept only (b = 1)				
Expanded relative uncertainty	25 %	29.0 %	29.2 %	Fail
All data, use slope only (a = 0)				
Expanded relative uncertainty	25 %	18.9 %	18.3 %	Pass
All data, PM ₁₀ >30 $\mu\text{g}/\text{m}^3$				
Number of values	≥ 40	47	48	Pass
Slope, b		0.677	0.732	
Intercept, a		13.49	10.503	
Expanded rel. uncertainty	25 %	34.5 %	29.2 %	Fail
Hjortneskaia RSA, autumn				
Number of values	≥ 40	42	42	Pass
Slope, b		1.018	0.996	
Intercept, a		3.554	3.595	
Expanded rel. uncertainty	25 %	9.3 %	9.7%	Pass
Hjortneskaia RSW, winter				
Number of values	≥ 40	46	50	Pass
Slope, b		0.938	0.930	
Intercept, a		3.625	3.45	
Expanded rel. uncertainty	25 %	22.4 %	22.2 %	Pass
Sofienbergparken UBW, winter				
Number of values	≥ 40	49	50	Pass
Slope, b		1.047	1.065	
Intercept, a		0.318	0.195	
Expanded rel. uncertainty	25 %	6.2 %	6.9 %	Pass
Sofienbergparken UBS, summer				
Number of values	≥ 40	34	45	Pass
Slope, b		0.837	0.832	
Intercept, a		0.931	1.522	
Expanded rel. uncertainty	25 %	10.8 %	9.2 %	Pass
Smestad RSW2, winter				
Number of values	≥ 40	47	47	Pass
Slope, b		0.839	0.860	
Intercept, a		-0.510	0.684	
Expanded rel. uncertainty	25 %	16.5 %	16.5 %	Pass

Table 20. Regression analysis. Palas Fidas 200 PM₁₀ candidate Fidas1, first and second comparison, RSA, RSW, UBW, UBS, RSW2.

RAW DATA			RESULTS AFTER CALIBRATING		
Regression	0.898y + 2.237		N (Spring)	55	n
Regression (i=0)	0.969y		N (Summer)	14	n
N	218	n	N (Fall)	68	n
			N (Winter)	81	n
Outliers	8	n	Outliers	4	n
Outliers	3.7	%	Outliers	1.8	%
Mean CM	20.77	µg/m3	Mean CM	20.88	µg/m3
Mean RM	20.88	µg/m3	Mean RM	20.88	µg/m3
Number of RM > UAT	47	n	Number of CM > UAT	48	n
Number of RM > LV	10	n	Number of CM > LV	8	n
REGRESSION RESULTS (RAW)			REGRESSION RESULTS (CALIBRATED)		
Slope b	1.1141	significant	Slope b	0.9954	
Uncertainty of b	0.0215		Uncertainty of b	0.0193	
Intercept a	-2.4923	significant	Intercept a	0.0970	
Uncertainty of a	0.5389		Uncertainty of a	0.4837	
r ²	0.919		r ²	0.919	
Slope b forced through origin	1.031	significant			
Uncertainty of b (forced)	0.0121				
EQUIVALENCE TEST (RAW)			EQUIVALENCE TEST (CALIBRATED)		
Uncertainty of calibration	1.203	µg/m3	Calibration	(y+2.492) / 1.114	
Uncertainty of calibration (forced)	0.603	µg/m3	Uncertainty of calibration	1.203	µg/m3
Random term	4.4008	µg/m3	Random term	4.1086	µg/m3
Additional uncertainty (optional)	0.00	µg/m3	Additional uncertainty (optional)	0.00	µg/m3
Bias at LV	3.2115	µg/m3	Bias at LV	-0.1353	µg/m3
Combined uncertainty	5.4480	µg/m3	Combined uncertainty	4.1108	µg/m3
Expanded relative uncertainty	21.7921%	pass	Expanded relative uncertainty	16.4432%	pass
Ref sampler uncertainty	0.6900	µg/m3	Ref sampler uncertainty	0.6900	µg/m3
Limit value	50	µg/m3	Limit value	50	µg/m3

Table 21. Regression analysis. Palas Fidas 200 PM₁₀ candidate Fidas2, first and second comparison, RSA, RSW, UBW, UBS, RSW2.

RAW DATA			RESULTS AFTER CALIBRATING		
Regression	0.894y + 2.279		N (Spring)	55	n
Regression (i=0)	0.968y		N (Summer)	25	n
N	234	n	N (Fall)	73	n
			N (Winter)	81	n
Outliers	9	n	Outliers	7	n
Outliers	3.8	%	Outliers	3.0	%
Mean CM	20.13	µg/m3	Mean CM	20.27	µg/m3
Mean RM	20.27	µg/m3	Mean RM	20.27	µg/m3
Number of RM > UAT	48	n	Number of CM > UAT	49	n
Number of RM > LV	10	n	Number of CM > LV	9	n
REGRESSION RESULTS (RAW)			REGRESSION RESULTS (CALIBRATED)		
Slope b	1.1190	significant	Slope b	0.9955	
Uncertainty of b	0.0201		Uncertainty of b	0.0180	
Intercept a	-2.5496	significant	Intercept a	0.0909	
Uncertainty of a	0.4925		Uncertainty of a	0.4402	
r ²	0.924		r ²	0.924	
Slope b forced through origin	1.033	significant			
Uncertainty of b (forced)	0.0115				
EQUIVALENCE TEST (RAW)			EQUIVALENCE TEST (CALIBRATED)		
Uncertainty of calibration	1.120	µg/m3	Calibration	(y+2.550) / 1.119	
Uncertainty of calibration (forced)	0.575	µg/m3	Uncertainty of calibration	1.120	µg/m3
Random term	4.2152	µg/m3	Random term	3.9092	µg/m3
Additional uncertainty (optional)	0.00	µg/m3	Additional uncertainty (optional)	0.00	µg/m3
Bias at LV	3.3985	µg/m3	Bias at LV	-0.1334	µg/m3
Combined uncertainty	5.4146	µg/m3	Combined uncertainty	3.9115	µg/m3
Expanded relative uncertainty	21.6585%	pass	Expanded relative uncertainty	15.6458%	pass
Ref sampler uncertainty	0.6900	µg/m3	Ref sampler uncertainty	0.6900	µg/m3
Limit value	50	µg/m3	Limit value	50	µg/m3

Table 22. Performance characteristics of the Palas Fidas 200 PM₁₀ candidates Fidas1 and Fidas2, first comparison only, RSA, RSW, UBW, UBS. Comments in red indicate significant deviations from performance criteria. Comments in blue indicate that calibration is necessary.

Test PM ₁₀	Criteria	Fidas1	Fidas2	Comment
Between CM uncertainty				
All data	$U_{bs,CM} < 2.5 \mu\text{g}/\text{m}^3$	0.38 $\mu\text{g}/\text{m}^3$		Pass
> 30 $\mu\text{g}/\text{m}^3$	$U_{bs,CM} < 2.5 \mu\text{g}/\text{m}^3$	0.63 $\mu\text{g}/\text{m}^3$		Pass
Comparability				
All data				
Number of values		171	187	
Data capture	$\geq 90 \%$	89.5 %	97.9 %	Fail / Pass
Average	Ref. 18.99 18.38	17.52	16.99	
Slope, b	$0.98 < b < 1.02$	1.0035	1.015	Pass
Intercept, a	$-1 < a < 1$	-1.5394	-1.666	Calibrate
Expanded relative uncertainty	25 %	14.4 %	13.4 %	Pass
Calibrated data, $RM = a + b * CM$				
All data				
Slope, b		0.966	0.985	
Intercept, a		1.534	1.642	
Expanded relative uncertainty	25 %	14.0 %	13.3 %	Pass
All data, use intercept only (b = 1)				
Expanded relative uncertainty	25 %	13.5 %	13.3 %	Pass
All data, use slope only (a = 0)				
Expanded relative uncertainty	25 %	15.2 %	14.8 %	Pass
All data, PM ₁₀ >30 $\mu\text{g}/\text{m}^3$				
Number of values	≥ 40	29	30	Fail
Slope, b		0.654	0.745	
Intercept, a		16.885	12.281	
Expanded rel. uncertainty	25 %	42.2 %	33.5 %	Fail

Table 23. Regression analysis Palas Fidas 200 PM₁₀ candidate Fidas1, first comparison only, RSA, RSW, UBW, UBS.

RAW DATA			RESULTS AFTER CALIBRATING		
Regression	0.996y + 1.534		N (Spring)	34	n
Regression (i=0)	1.055y		N (Summer)	14	n
N	171	n	N (Fall)	68	n
			N (Winter)	55	n
Outliers	5	n	Outliers	4	n
Outliers	2.9	%	Outliers	2.3	%
Mean CM	17.52	µg/m3	Mean CM	18.99	µg/m3
Mean RM	18.99	µg/m3	Mean RM	18.99	µg/m3
Number of RM > UAT	29	n	Number of CM > UAT	25	n
Number of RM > LV	7	n	Number of CM > LV	7	n
REGRESSION RESULTS (RAW)			REGRESSION RESULTS (CALIBRATED)		
Slope b	1.0035		Slope b	0.9999	
Uncertainty of b	0.0200		Uncertainty of b	0.0199	
Intercept a	-1.5394	significant	Intercept a	0.0024	
Uncertainty of a	0.4584		Uncertainty of a	0.4568	
r ²	0.933		r ²	0.933	
Slope b forced through origin	0.948	significant			
Uncertainty of b (forced)	0.0113				
EQUIVALENCE TEST (RAW)			EQUIVALENCE TEST (CALIBRATED)		
Uncertainty of calibration	1.098	µg/m3	Calibration	(y+1.539) / 1.004	
Uncertainty of calibration (forced)	0.566	µg/m3	Uncertainty of calibration	1.098	µg/m3
Random term	3.3338	µg/m3	Random term	3.4981	µg/m3
Additional uncertainty (optional)	0.00	µg/m3	Additional uncertainty (optional)	0.00	µg/m3
Bias at LV	-1.3627	µg/m3	Bias at LV	-0.0038	µg/m3
Combined uncertainty	3.6016	µg/m3	Combined uncertainty	3.4981	µg/m3
Expanded relative uncertainty	14.4063%	pass	Expanded relative uncertainty	13.9923%	pass
Ref sampler uncertainty	0.6900	µg/m3	Ref sampler uncertainty	0.6900	µg/m3
Limit value	50	µg/m3	Limit value	50	µg/m3

Table 24. Regression analysis Palas Fidas 200 PM₁₀ candidate Fidas2, first comparison only, RSA, RSW, UBW, UBS.

RAW DATA			RESULTS AFTER CALIBRATING		
Regression	0.985y + 1.642		N (Spring)	34	n
Regression (i=0)	1.048y		N (Summer)	25	n
N	187	n	N (Fall)	73	n
			N (Winter)	55	n
Outliers	5	n	Outliers	5	n
Outliers	2.7	%	Outliers	2.7	%
Mean CM	16.99	µg/m3	Mean CM	18.38	µg/m3
Mean RM	18.38	µg/m3	Mean RM	18.38	µg/m3
Number of RM > UAT	30	n	Number of CM > UAT	27	n
Number of RM > LV	7	n	Number of CM > LV	6	n
REGRESSION RESULTS (RAW)			REGRESSION RESULTS (CALIBRATED)		
Slope b	1.0150		Slope b	0.9995	
Uncertainty of b	0.0187		Uncertainty of b	0.0184	
Intercept a	-1.6664	significant	Intercept a	0.0090	
Uncertainty of a	0.4185		Uncertainty of a	0.4123	
r ²	0.937		r ²	0.937	
Slope b forced through origin	0.954	significant			
Uncertainty of b (forced)	0.0108				
EQUIVALENCE TEST (RAW)			EQUIVALENCE TEST (CALIBRATED)		
Uncertainty of calibration	1.025	µg/m3	Calibration	(y+1.666) / 1.015	
Uncertainty of calibration (forced)	0.540	µg/m3	Uncertainty of calibration	1.025	µg/m3
Random term	3.2117	µg/m3	Random term	3.3232	µg/m3
Additional uncertainty (optional)	0.00	µg/m3	Additional uncertainty (optional)	0.00	µg/m3
Bias at LV	-0.9172	µg/m3	Bias at LV	-0.0155	µg/m3
Combined uncertainty	3.3401	µg/m3	Combined uncertainty	3.3233	µg/m3
Expanded relative uncertainty	13.3602%	pass	Expanded relative uncertainty	13.2931%	pass
Ref sampler uncertainty	0.6900	µg/m3	Ref sampler uncertainty	0.6900	µg/m3
Limit value	50	µg/m3	Limit value	50	µg/m3

Table 25. Performance characteristics of the Grimm EDM 180 PM₁₀ candidates Grimm1 and Grimm2, first and second comparison, RSA, RSW, UBW, UBS, RSW2. Between CM uncertainty is based on first comparison only. Comments in red indicate significant deviations from performance criteria. Comments in blue indicate that calibration is necessary.

Test PM ₁₀	Criteria	Grimm1	Grimm2	Comment
Between CM uncertainty				
All data > 30 µg/m ³	$U_{bs,CM} < 2.5 \mu\text{g}/\text{m}^3$ $U_{bs,CM} < 2.5 \mu\text{g}/\text{m}^3$	0.94 µg/m ³ 1.78 µg/m ³		Pass Pass
Comparability				
All data				
Number of values		234	227	
Data capture	≥ 90 %	97.9 %	95.0 %	Pass
Average	Ref. 20.36 20.67	18.36	18.68	
Slope, b	$0.98 < b < 1.02$	0.9743	0.9750	Calibrate
Intercept, a	$-1 < a < 1$	-1.4799	-1.4716	Calibrate
Expanded relative uncertainty	25 %	27.1 %	25.3 %	Fail
Calibrated data, $RM = a + b * CM$				
All data				
Slope, b		1.026	1.026	
Intercept, a		1.519	1.509	
Expanded relative uncertainty	25 %	26.2 %	24.2 %	Fail/pass
All data, use intercept only (b = 1)				
Expanded relative uncertainty	25 %	25.4 %	23.5 %	Fail/Pass
All data, use slope only (a = 0)				
Expanded relative uncertainty	25 %	26.7 %	24.7 %	Fail/Pass
All data, PM ₁₀ >30 µg/m ³				
Number of values	≥ 40	49	49	Pass
Slope, b		0.661	0.696	
Intercept, a		18.04	16.31	
Expanded rel. uncertainty	25 %	46.1 %	45.0 %	Fail
Hjortneskaia RSA, autumn				
Number of values	≥ 40	42	42	Pass
Slope, b		1.394	1.366	
Intercept, a		0.965	0.565	
Expanded rel. uncertainty	25 %	12.3 %	13.3 %	Pass
Hjortneskaia RSW, winter				
Number of values	≥ 40	52	52	Pass
Slope, b		1.161	1.161	
Intercept, a		4.684	3.448	
Expanded rel. uncertainty	25 %	34.9 %	34.3 %	Fail
Sofienbergparken UBW, winter				
Number of values	≥ 40	49	42	Pass
Slope, b		0.838	0.902	
Intercept, a		1.295	0.766	
Expanded rel. uncertainty	25 %	10.7 %	10.2 %	Pass
Sofienbergparken UBS, summer				
Number of values	≥ 40	47	47	Pass
Slope, b		0.674	0.729	
Intercept, a		1.962	1.802	
Expanded rel. uncertainty	25 %	10.5 %	8.5 %	Pass
Smestad RS2, winter				
Number of values	≥ 40	44	44	Pass
Slope, b		0.783	0.894	
Intercept, a		4.460	3.227	
Expanded rel. uncertainty	25 %	17.2 %	13.5 %	Pass

Table 26. Regression analysis Grimm EDM 180 PM₁₀ candidate Grimm1, RSA, RSW, UBW, UBS, first and second comparison, RSW2.

RAW DATA			RESULTS AFTER CALIBRATING		
Regression	1.026y + 1.519		N (Spring)	51	n
Regression (≠0)	1.083y		N (Summer)	27	n
N	234	n	N (Fall)	74	n
			N (Winter)	82	n
Outliers	7	n	Outliers	7	n
Outliers	3.0	%	Outliers	3.0	%
Mean CM	18.36	µg/m3	Mean CM	20.36	µg/m3
Mean RM	20.36	µg/m3	Mean RM	20.36	µg/m3
Number of RM > UAT	49	n	Number of CM > UAT	48	n
Number of RM > LV	10	n	Number of CM > LV	8	n
REGRESSION RESULTS (RAW)			REGRESSION RESULTS (CALIBRATED)		
Slope b	0.9743		Slope b	1.0030	
Uncertainty of b	0.0284		Uncertainty of b	0.0292	
Intercept a	-1.4799	significant	Intercept a	-0.0613	
Uncertainty of a	0.7013		Uncertainty of a	0.7198	
r ²	0.804		r ²	0.804	
Slope b forced through origin	0.924	significant			
Uncertainty of b (forced)	0.0180				
EQUIVALENCE TEST (RAW)			EQUIVALENCE TEST (CALIBRATED)		
Uncertainty of calibration	1.584	µg/m3	Calibration	(y+1.480) / 0.974	
Uncertainty of calibration (forced)	0.802	µg/m3	Uncertainty of calibration	1.584	µg/m3
Random term	6.1816	µg/m3	Random term	6.5507	µg/m3
Additional uncertainty (optional)	0.00	µg/m3	Additional uncertainty (optional)	0.00	µg/m3
Bias at LV	-2.7653	µg/m3	Bias at LV	0.0892	µg/m3
Combined uncertainty	6.7719	µg/m3	Combined uncertainty	6.5513	µg/m3
Expanded relative uncertainty	27.0876%	fail	Expanded relative uncertainty	26.2051%	fail
Ref sampler uncertainty	0.6900	µg/m3	Ref sampler uncertainty	0.6900	µg/m3
Limit value	50	µg/m3	Limit value	50	µg/m3

Table 27. Regression analysis Grimm EDM 180 PM₁₀ candidate Grimm2, RSA, RSW, UBW, UBS, first and second comparison, RSW2.

RAW DATA			RESULTS AFTER CALIBRATING		
Regression	1.026y + 1.509		N (Spring)	51	n
Regression (≠0)	1.08y		N (Summer)	27	n
N	227	n	N (Fall)	74	n
			N (Winter)	75	n
Outliers	8	n	Outliers	8	n
Outliers	3.5	%	Outliers	3.5	%
Mean CM	18.68	µg/m3	Mean CM	20.67	µg/m3
Mean RM	20.67	µg/m3	Mean RM	20.67	µg/m3
Number of RM > UAT	49	n	Number of CM > UAT	47	n
Number of RM > LV	10	n	Number of CM > LV	7	n
REGRESSION RESULTS (RAW)			REGRESSION RESULTS (CALIBRATED)		
Slope b	0.9750		Slope b	1.0024	
Uncertainty of b	0.0266		Uncertainty of b	0.0273	
Intercept a	-1.4716	significant	Intercept a	-0.0499	
Uncertainty of a	0.6644		Uncertainty of a	0.6815	
r ²	0.833		r ²	0.833	
Slope b forced through origin	0.926	significant			
Uncertainty of b (forced)	0.0149				
EQUIVALENCE TEST (RAW)			EQUIVALENCE TEST (CALIBRATED)		
Uncertainty of calibration	1.486	µg/m3	Calibration	(y+1.472) / 0.975	
Uncertainty of calibration (forced)	0.746	µg/m3	Uncertainty of calibration	1.486	µg/m3
Random term	5.7074	µg/m3	Random term	6.0483	µg/m3
Additional uncertainty (optional)	0.00	µg/m3	Additional uncertainty (optional)	0.00	µg/m3
Bias at LV	-2.7204	µg/m3	Bias at LV	0.0708	µg/m3
Combined uncertainty	6.3226	µg/m3	Combined uncertainty	6.0487	µg/m3
Expanded relative uncertainty	25.2904%	fail	Expanded relative uncertainty	24.1948%	pass
Ref sampler uncertainty	0.6900	µg/m3	Ref sampler uncertainty	0.6900	µg/m3
Limit value	50	µg/m3	Limit value	50	µg/m3

Table 28. Performance characteristics of the Grimm EDM 180 PM₁₀ candidates Grimm1 and Grimm2, first comparison only, RSA, RSW, UBW, UBS. Comments in red indicate significant deviations from performance criteria. Comments in blue indicate that calibration is necessary.

Test PM ₁₀	Criteria	Grimm1	Grimm2	Comment
Between CM uncertainty				
All data	$U_{bs,CM} < 2.5 \mu\text{g}/\text{m}^3$	0.94 $\mu\text{g}/\text{m}^3$		Pass
> 30 $\mu\text{g}/\text{m}^3$	$U_{bs,CM} < 2.5 \mu\text{g}/\text{m}^3$	1.78 $\mu\text{g}/\text{m}^3$		Pass
Comparability				
All data				
Number of values		190	183	
Data capture	$\geq 90 \%$	99.5 %	95.8 %	Pass
Average	Ref. 18.44 18.74	15.73	16.03	
Slope, b	$0.98 < b < 1.02$	0.8108	0.8160	Calibrate
Intercept, a	$-1 < a < 1$	0.7843	0.7372	Pass
Expanded relative uncertainty	25 %	41 %	39.1 %	Fail
Calibrated data, $RM = a + b * CM$				
All data				
Slope, b		1.233	1.226	
Intercept, a		-0.967	-0.903	
Expanded relative uncertainty	25 %	28.5 %	25.3 %	Fail
All data, use intercept only (b = 1)				
Expanded relative uncertainty	25 %	43.8 %	41.8 %	Fail
All data, use slope only (a = 0)				
Expanded relative uncertainty	25 %	29.1 %	25.8 %	Fail
All data, PM ₁₀ > 30 $\mu\text{g}/\text{m}^3$				
Number of values	≥ 40	31	31	Fail
Slope, b		0.799	0.904	
Intercept, a		16.603	12.306	
Expanded rel. uncertainty	25 %	55.5 %	54.8 %	Fail

Table 29. Regression analysis Grimm EDM 180 PM₁₀ candidate Grimm1, first comparison only, RSA, RSW, UBW, UBS.

RAW DATA			RESULTS AFTER CALIBRATING		
Regression	1.233y + -0.967		N (Spring)	34	n
Regression (i=0)	1.19y		N (Summer)	27	n
N	190	n	N (Fall)	74	n
			N (Winter)	55	n
Outliers	10	n	Outliers	16	n
Outliers	5.3	%	Outliers	8.4	%
Mean CM	15.73	µg/m3	Mean CM	18.44	µg/m3
Mean RM	18.44	µg/m3	Mean RM	18.44	µg/m3
Number of RM > UAT	31	n	Number of CM > UAT	33	n
Number of RM > LV	7	n	Number of CM > LV	6	n
REGRESSION RESULTS (RAW)			REGRESSION RESULTS (CALIBRATED)		
Slope b	0.8108	significant	Slope b	1.0323	
Uncertainty of b	0.0303		Uncertainty of b	0.0374	
Intercept a	0.7843		Intercept a	-0.5954	
Uncertainty of a	0.6812		Uncertainty of a	0.8401	
r ²	0.752		r ²	0.752	
Slope b forced through origin	0.841	significant			
Uncertainty of b (forced)	0.0179				
EQUIVALENCE TEST (RAW)			EQUIVALENCE TEST (CALIBRATED)		
Uncertainty of calibration	1.660	µg/m3	Calibration	(y-0.784) / 0.811	
Uncertainty of calibration (forced)	0.893	µg/m3	Uncertainty of calibration	1.660	µg/m3
Random term	5.4546	µg/m3	Random term	7.0441	µg/m3
Additional uncertainty (optional)	0.00	µg/m3	Additional uncertainty (optional)	0.00	µg/m3
Bias at LV	-8.6769	µg/m3	Bias at LV	1.0193	µg/m3
Combined uncertainty	10.2490	µg/m3	Combined uncertainty	7.1174	µg/m3
Expanded relative uncertainty	40.9959%	fail	Expanded relative uncertainty	28.4698%	fail
Ref sampler uncertainty	0.6900	µg/m3	Ref sampler uncertainty	0.6900	µg/m3
Limit value	50	µg/m3	Limit value	50	µg/m3

Table 30. Regression analysis Grimm EDM 180 PM₁₀ candidate Grimm2, first comparison only, RSA, RSW, UBW, UBS.

RAW DATA			RESULTS AFTER CALIBRATING		
Regression	1.226y + -0.903		N (Spring)	34	n
Regression (i=0)	1.186y		N (Summer)	27	n
N	183	n	N (Fall)	74	n
			N (Winter)	48	n
Outliers	7	n	Outliers	12	n
Outliers	3.8	%	Outliers	6.6	%
Mean CM	16.03	µg/m3	Mean CM	18.74	µg/m3
Mean RM	18.74	µg/m3	Mean RM	18.74	µg/m3
Number of RM > UAT	31	n	Number of CM > UAT	30	n
Number of RM > LV	7	n	Number of CM > LV	7	n
REGRESSION RESULTS (RAW)			REGRESSION RESULTS (CALIBRATED)		
Slope b	0.8160	significant	Slope b	1.0238	
Uncertainty of b	0.0276		Uncertainty of b	0.0338	
Intercept a	0.7372		Intercept a	-0.4462	
Uncertainty of a	0.6300		Uncertainty of a	0.7720	
r ²	0.801		r ²	0.801	
Slope b forced through origin	0.843	significant			
Uncertainty of b (forced)	0.0162				
EQUIVALENCE TEST (RAW)			EQUIVALENCE TEST (CALIBRATED)		
Uncertainty of calibration	1.517	µg/m3	Calibration	(y-0.737) / 0.816	
Uncertainty of calibration (forced)	0.808	µg/m3	Uncertainty of calibration	1.517	µg/m3
Random term	4.8991	µg/m3	Random term	6.2758	µg/m3
Additional uncertainty (optional)	0.00	µg/m3	Additional uncertainty (optional)	0.00	µg/m3
Bias at LV	-8.4641	µg/m3	Bias at LV	0.7442	µg/m3
Combined uncertainty	9.7797	µg/m3	Combined uncertainty	6.3198	µg/m3
Expanded relative uncertainty	39.1187%	fail	Expanded relative uncertainty	25.2791%	fail
Ref sampler uncertainty	0.6900	µg/m3	Ref sampler uncertainty	0.6900	µg/m3
Limit value	50	µg/m3	Limit value	50	µg/m3

Table 31. Performance characteristics of the TEOM 1405DF PM₁₀ candidate TEOMDF, based on data from RSW (partly) and urban background site UBW and UBS. Comments in red indicate significant deviations from performance criteria. Comments in blue indicate that calibration is necessary.

Test PM ₁₀	Criteria	TEOMDF	X	Comment
Between CM uncertainty				
All data > 30 µg/m ³	$U_{bs,CM} < 2.5 \mu\text{g}/\text{m}^3$ $U_{bs,CM} < 2.5 \mu\text{g}/\text{m}^3$			Only 1 candidate
Comparability				
All data				
Number of values		113		
Data capture	≥ 90 %	59.2 %		Fail
Average	Ref. 15.07	13.55		
Slope, b	$0.98 < b < 1.02$	0.8884		Calibrate
Intercept, a	$-1 < a < 1$	0.1670		Pass
Expanded relative uncertainty	25 %	23.1 %	%	Pass
Calibrated data, $RM = a + b * CM$				
All data				
Slope, b		1.126		
Intercept, a		-0.188		
Expanded relative uncertainty	25 %	10.0 %	%	Pass
All data, use intercept only (b = 1)				
Expanded relative uncertainty	25 %	23.8 %	%	Pass
All data, use slope only (a = 0)				
Expanded relative uncertainty	25 %	10.0 %	%	Pass
All data, PM ₁₀ > 30 µg/m ³				
Number of values	≥ 40	11		Fail
Slope, b		0.989		
Intercept, a		5.701		
Expanded rel. uncertainty	25 %	39.6 %	%	Fail
Hjortneskaia RSA, autumn				
Number of values	≥ 40			Not in operation
Slope, b				
Intercept, a				
Expanded rel. uncertainty	25 %	%	%	
Hjortneskaia RSW, winter				
Number of values	≥ 40	19		Fail
Slope, b		1.084		
Intercept, a		1.971		
Expanded rel. uncertainty	25 %	21.1 %	%	Pass
Sofienbergparken UBW, winter				
Number of values	≥ 40	47		Pass
Slope, b		1.111		
Intercept, a		1.039		
Expanded rel. uncertainty	25 %	5.7 %	%	Pass
Sofienbergparken UBS, summer				
Number of values	≥ 40	47		Pass
Slope, b		0.963		
Intercept, a		0.078		
Expanded rel. uncertainty	25 %	5.6 %	%	Pass

Table 32. Regression analysis TEOM 1405DF PM₁₀ candidate TEOMDF, first comparison, RSW, UBW, UBS.

RAW DATA			RESULTS AFTER CALIBRATING		
Regression	1.126y + -0.188		N (Spring)	34	n
Regression (j=0)	1.117y		N (Summer)	27	n
N	113	n	N (Fall)	0	n
			N (Winter)	52	n
Outliers	4	n	Outliers	4	n
Outliers	3.5	%	Outliers	3.5	%
Mean CM	13.55	µg/m3	Mean CM	15.07	µg/m3
Mean RM	15.07	µg/m3	Mean RM	15.07	µg/m3
Number of RM > UAT	11	n	Number of CM > UAT	9	n
Number of RM > LV	2	n	Number of CM > LV	1	n
REGRESSION RESULTS (RAW)			REGRESSION RESULTS (CALIBRATED)		
Slope b	0.8884	significant	Slope b	1.0027	
Uncertainty of b	0.0177		Uncertainty of b	0.0199	
Intercept a	0.1670		Intercept a	-0.0400	
Uncertainty of a	0.3337		Uncertainty of a	0.3756	
r ²	0.956		r ²	0.956	
Slope b forced through origin	0.896	significant			
Uncertainty of b (forced)	0.0107				
EQUIVALENCE TEST (RAW)			EQUIVALENCE TEST (CALIBRATED)		
Uncertainty of calibration	0.944	µg/m3	Calibration	(y-0.167) / 0.888	
Uncertainty of calibration (forced)	0.535	µg/m3	Uncertainty of calibration	0.944	µg/m3
Random term	2.0378	µg/m3	Random term	2.5086	µg/m3
Additional uncertainty (optional)	0.00	µg/m3	Additional uncertainty (optional)	0.00	µg/m3
Bias at LV	-5.4119	µg/m3	Bias at LV	0.0929	µg/m3
Combined uncertainty	5.7829	µg/m3	Combined uncertainty	2.5104	µg/m3
Expanded relative uncertainty	23.1314%	pass	Expanded relative uncertainty	10.0415%	pass
Ref sampler uncertainty	0.6900	µg/m3	Ref sampler uncertainty	0.6900	µg/m3
Limit value	50	µg/m3	Limit value	50	µg/m3

Table 33. Performance characteristics of TEI FH62 IR PM₁₀ candidate TEI3 with SA 246b (USEPA) impactor, based on data from RSA and RSW. Comments in red indicate significant deviations from performance criteria. Comments in blue indicate that calibration is necessary.

Test PM ₁₀	Criteria	TEI3	X	Comment
Between CM uncertainty				
All data > 30 µg/m ³	$U_{bs,CM} < 2.5 \mu\text{g}/\text{m}^3$ $U_{bs,CM} < 2.5 \mu\text{g}/\text{m}^3$			Only 1 candidate
Comparability				
All data				
Number of values		94		
Data capture	≥ 90 %	49.2 %		Fail
Average	Ref. 24.78	29.59		
Slope, b	$0.98 < b < 1.02$	1.2212		Calibrate
Intercept, a	$-1 < a < 1$	-0.6774	%	Pass
Expanded relative uncertainty	25 %	42.8 %		Fail
Calibrated data, $RM = a + b * CM$				
All data				
Slope, b		0.819		
Intercept, a		0.555		
Expanded relative uncertainty	25 %	9.5 %	%	Pass
All data, use intercept only (b = 1)				
Expanded relative uncertainty	25 %	45.5 %	%	Fail
All data, use slope only (a = 0)				
Expanded relative uncertainty	25 %	9.5 %	%	Pass
All data, PM ₁₀ > 30 µg/m ³				
Number of values	≥ 40	29		Fail
Slope, b		0.785		
Intercept, a		2.588		
Expanded rel. uncertainty	25 %	25.5 %	%	Fail
Hjortneskaia RSA, autumn				
Number of values	≥ 40	42		Pass
Slope, b		0.783		
Intercept, a		1.334		
Expanded rel. uncertainty	25 %	5.6 %	%	Pass
Hjortneskaia RSW, winter				
Number of values	≥ 40	52		Pass
Slope, b		0.829		
Intercept, a		0.175		
Expanded rel. uncertainty	25 %	13.1 %	%	Pass
Sofienbergparken UBW, winter				
Number of values	≥ 40			Not in operation
Slope, b				
Intercept, a				
Expanded rel. uncertainty	25 %	%	%	
Sofienbergparken UBS, summer				
Number of values	≥ 40			Not in operation
Slope, b				
Intercept, a				
Expanded rel. uncertainty	25 %	%	%	

Table 34. Regression analysis TEI FH62 IR PM₁₀ candidate TEI3 with SA 246b (USEPA) impactor, first comparison, RSA, RSW.

RAW DATA			RESULTS AFTER CALIBRATING		
Regression	0.819y + 0.555		N (Spring)	0	n
Regression (i=0)	0.833y		N (Summer)	0	n
N	94	n	N (Fall)	74	n
			N (Winter)	20	n
Outliers	3	n	Outliers	1	n
Outliers	3.2	%	Outliers	1.1	%
Mean CM	29.59	µg/m3	Mean CM	24.78	µg/m3
Mean RM	24.78	µg/m3	Mean RM	24.78	µg/m3
Number of RM > UAT	29	n	Number of CM > UAT	27	n
Number of RM > LV	7	n	Number of CM > LV	5	n
REGRESSION RESULTS (RAW)			REGRESSION RESULTS (CALIBRATED)		
Slope b	1.2212	significant	Slope b	0.9977	
Uncertainty of b	0.0194		Uncertainty of b	0.0158	
Intercept a	-0.6774		Intercept a	0.0577	
Uncertainty of a	0.5533		Uncertainty of a	0.4531	
r ²	0.977		r ²	0.977	
Slope b forced through origin	1.201	significant			
Uncertainty of b (forced)	0.0096				
EQUIVALENCE TEST (RAW)			EQUIVALENCE TEST (CALIBRATED)		
Uncertainty of calibration	1.115	µg/m3	Calibration	(y+0.677) / 1.221	
Uncertainty of calibration (forced)	0.480	µg/m3	Uncertainty of calibration	1.115	µg/m3
Random term	2.5959	µg/m3	Random term	2.3647	µg/m3
Additional uncertainty (optional)	0.00	µg/m3	Additional uncertainty (optional)	0.00	µg/m3
Bias at LV	10.3825	µg/m3	Bias at LV	-0.0587	µg/m3
Combined uncertainty	10.7021	µg/m3	Combined uncertainty	2.3655	µg/m3
Expanded relative uncertainty	42.8085%	fail	Expanded relative uncertainty	9.4618%	pass
Ref sampler uncertainty	0.6900	µg/m3	Ref sampler uncertainty	0.6900	µg/m3
Limit value	50	µg/m3	Limit value	50	µg/m3

Table 35. Performance characteristics of TEI FH62 IR PM₁₀ candidate TEI3a with EN12341 impactor, based on data from the UBW and UBS. Comments in red indicate significant deviations from performance criteria. Comments in blue indicate that calibration is necessary.

Test PM ₁₀	Criteria	TEI3a	X	Comment
Between CM uncertainty				
All data > 30 µg/m ³	$U_{bs,CM} < 2.5 \mu\text{g}/\text{m}^3$ $U_{bs,CM} < 2.5 \mu\text{g}/\text{m}^3$			Only 1 candidate
Comparability				
All data				
Number of values		96		
Data capture	≥ 90 %	50.3 %		Fail
Average	Ref. 12.25	12.96		
Slope, b	$0.98 < b < 1.02$	1.0407		Calibrate
Intercept, a	$-1 < a < 1$	0.2092		
Expanded relative uncertainty	25 %	10.5 %	%	Pass
Calibrated data, $RM = a + b * CM$				
All data				
Slope, b		0.961		
Intercept, a		-0.201		
Expanded relative uncertainty	25 %	6.9 %	%	Pass
All data, use intercept only (b = 1)				
Expanded relative uncertainty	25 %	9.9 %	%	
All data, use slope only (a = 0)				
Expanded relative uncertainty	25 %	6.8 %	%	
All data, PM ₁₀ > 30 µg/m ³				
Number of values	≥ 40	2		Fail
Slope, b				
Intercept, a				
Expanded rel. uncertainty	25 %	%	%	
Hjortneskaia RSA, autumn				
Number of values	≥ 40			Not in operation
Slope, b				
Intercept, a				
Expanded rel. uncertainty	25 %	%	%	
Hjortneskaia RSW, winter				
Number of values	≥ 40			Not in operation
Slope, b				
Intercept, a				
Expanded rel. uncertainty	25 %	%	%	
Sofienbergparken UBW, winter				
Number of values	≥ 40	49		Pass
Slope, b		0.989		
Intercept, a		-0.072		
Expanded rel. uncertainty	25 %	7.4 %	%	Pass
Sofienbergparken UBS, summer				
Number of values	≥ 40	47		Pass
Slope, b		0.861		
Intercept, a		0.433		
Expanded rel. uncertainty	25 %	8.0 %	%	Pass

Table 36. Regression analysis TEI FH62 IR PM₁₀ candidate TEI3a with EN12341 impactor, first comparison, UBW, UBS.

RAW DATA			RESULTS AFTER CALIBRATING		
Regression	0.961y + -0.201		N (Spring)	34	n
Regression (i=0)	0.949y		N (Summer)	27	n
N	96	n	N (Fall)	0	n
			N (Winter)	35	n
Outliers	4	n	Outliers	1	n
Outliers	4.2	%	Outliers	1.0	%
Mean CM	12.96	µg/m3	Mean CM	12.25	µg/m3
Mean RM	12.25	µg/m3	Mean RM	12.25	µg/m3
Number of RM > UAT	2	n	Number of CM > UAT	2	n
Number of RM > LV	0	n	Number of CM > LV	0	n
REGRESSION RESULTS (RAW)			REGRESSION RESULTS (CALIBRATED)		
Slope b	1.0407		Slope b	0.9992	
Uncertainty of b	0.0218		Uncertainty of b	0.0209	
Intercept a	0.2092		Intercept a	0.0104	
Uncertainty of a	0.3090		Uncertainty of a	0.2969	
r ²	0.959		r ²	0.959	
Slope b forced through origin	1.054	significant			
Uncertainty of b (forced)	0.0111				
EQUIVALENCE TEST (RAW)			EQUIVALENCE TEST (CALIBRATED)		
Uncertainty of calibration	1.132	µg/m3	Calibration	(y-0.209) / 1.041	
Uncertainty of calibration (forced)	0.554	µg/m3	Uncertainty of calibration	1.132	µg/m3
Random term	1.3731	µg/m3	Random term	1.7271	µg/m3
Additional uncertainty (optional)	0.00	µg/m3	Additional uncertainty (optional)	0.00	µg/m3
Bias at LV	2.2444	µg/m3	Bias at LV	-0.0320	µg/m3
Combined uncertainty	2.6311	µg/m3	Combined uncertainty	1.7274	µg/m3
Expanded relative uncertainty	10.5244%	pass	Expanded relative uncertainty	6.9094%	pass
Ref sampler uncertainty	0.6900	µg/m3	Ref sampler uncertainty	0.6900	µg/m3
Limit value	50	µg/m3	Limit value	50	µg/m3

Appendix B

PM_{2.5} performance characteristics tables

Table 37. Performance characteristics of the $PM_{2.5}$ reference method, first and second comparison. Between RM uncertainty is based on first comparison only. When one sampler did not produce a valid result the result of the other sampler would represent the average of the two samplers. RM3-5 is the single $PM_{2.5}$ reference sampler in the second comparison.

Test $PM_{2.5}$	Criteria	RM	Pass/Fail
Suitability of data	20 % > 17 $\mu\text{g}/\text{m}^3$	12 %	Fail
Between RM uncertainty			
All data	$U_{bs, RM} < 2.0 \mu\text{g}/\text{m}^3$	0.42 $\mu\text{g}/\text{m}^3$	Pass
> 18 $\mu\text{g}/\text{m}^3$	$U_{bs, RM} < 2.0 \mu\text{g}/\text{m}^3$	0.53 $\mu\text{g}/\text{m}^3$	Pass
Number of paired samples			
All data		109	
> 18 $\mu\text{g}/\text{m}^3$		3	
Number of RM1 samples		111	
Number of RM3 samples		189	
Number of RM3-5 samples		45	
Number of RM averages			
First comparison		190	
First and second comparison		235	
Average			
First comparison		9.0 $\mu\text{g}/\text{m}^3$	
First and second comparison		9.6 $\mu\text{g}/\text{m}^3$	

Table 38. Performance characteristics of the Fidas 200 PM_{2.5} candidates Fidas1 and Fidas2, first and second comparison, RSA, RSW, UBW, UBS, RSW2. Between CM uncertainty is based on first comparison only. Comments in red indicate significant deviations from performance criteria. Comments in blue indicate that calibration is necessary.

Test PM2.5	Criteria	Fidas1	Fidas2	Comment
Between CM uncertainty				
All data	$U_{bs,CM} < 2.5 \mu\text{g}/\text{m}^3$	0.12 $\mu\text{g}/\text{m}^3$		Pass
> 18 $\mu\text{g}/\text{m}^3$	$U_{bs,CM} < 2.5 \mu\text{g}/\text{m}^3$	0.20 $\mu\text{g}/\text{m}^3$		Pass
Comparability				
All data				
Number of values		215	230	
Data capture	$\geq 90 \%$	91.5 %	97.9 %	Pass
Average	Ref. 10.04 9.70	9.26	8.89	
Slope, b	$0.98 < b < 1.02$	0.9914	0.9841	Pass
Intercept, a	$-1 < a < 1$	-0.6957	-0.6635	Pass
Expanded relative uncertainty	25 %	18.9 %	18.7 %	Pass
Calibrated data, $RM = a + b * CM$				
All data				
Slope, b		1.009	1.016	
Intercept, a		0.702	0.674	
Expanded relative uncertainty	25 %	19.0 %	18.4 %	Pass
All data, use intercept only (b = 1)				
Expanded relative uncertainty	25 %	18.0 %	17.5 %	Pass
All data, use slope only (a = 0)				
Expanded relative uncertainty	25 %	19.4 %	18.7 %	Pass
All data, PM _{2.5} > 18 $\mu\text{g}/\text{m}^3$				
Number of values	≥ 40	26	26	Fail
Slope, b		0.477	0.480	
Intercept, a		12.78	12.763	
Expanded rel. uncertainty	25 %	69.2 %	69.1 %	Fail
Hjortneskaia RSA, autumn				
Number of values	≥ 40	42	42	Pass
Slope, b		1.443	1.439	
Intercept, a		0.652	0.674	
Expanded rel. uncertainty	25 %	15.5 %	15.2 %	Pass
Hjortneskaia RSW, winter				
Number of values	≥ 40	46	51	Pass
Slope, b		1.083	1.098	
Intercept, a		1.104	1.061	
Expanded rel. uncertainty	25 %	27.3 %	25.3 %	Fail
Sofienbergparken UBW, winter				
Number of values	≥ 40	49	49	Pass
Slope, b		0.957	0.952	
Intercept, a		0.14	0.155	
Expanded rel. uncertainty	25 %	7.5 %	7.5 %	Pass
Sofienbergparken UBS, summer				
Number of values	≥ 40	33	43	Fail/Pass
Slope, b		0.835	0.831	
Intercept, a		0.391	0.596	
Expanded rel. uncertainty	25 %	13.8 %	11.0 %	Pass
Smestad RSW2, winter				
Number of values	≥ 40	45	45	Pass
Slope, b		0.744	0.857	
Intercept, a		2.075	1.609	
Expanded rel. uncertainty	25 %	19.3 %	19.0 %	Pass

Table 39. Regression analysis Fidas 200 PM_{2.5} candidate Fidas1, first and second comparison, RSA, RSW, UBW, UBS, RSW2.

RAW DATA			RESULTS AFTER CALIBRATING		
Regression	1.009y + 0.702		N (Spring)	54	n
Regression (i=0)	1.063y		N (Summer)	14	n
N	215	n	N (Fall)	68	n
			N (Winter)	79	n
Outliers	8	n	Outliers	6	n
Outliers	3.7	%	Outliers	2.8	%
Mean CM	9.26	µg/m3	Mean CM	10.04	µg/m3
Mean RM	10.04	µg/m3	Mean RM	10.04	µg/m3
Number of RM > UAT	26	n	Number of CM > UAT	24	n
Number of RM > LV	1	n	Number of CM > LV	3	n
REGRESSION RESULTS (RAW)			REGRESSION RESULTS (CALIBRATED)		
Slope b	0.9914		Slope b	1.0009	
Uncertainty of b	0.0289		Uncertainty of b	0.0292	
Intercept a	-0.6957	significant	Intercept a	-0.0091	
Uncertainty of a	0.3418		Uncertainty of a	0.3447	
r ²	0.819		r ²	0.819	
Slope b forced through origin	0.940	significant			
Uncertainty of b (forced)	0.0152				
EQUIVALENCE TEST (RAW)			EQUIVALENCE TEST (CALIBRATED)		
Uncertainty of calibration	0.933	µg/m3	Calibration	(y+0.696) / 0.991	
Uncertainty of calibration (forced)	0.457	µg/m3	Uncertainty of calibration	0.933	µg/m3
Random term	2.6718	µg/m3	Random term	2.8536	µg/m3
Additional uncertainty (optional)	0.00	µg/m3	Additional uncertainty (optional)	0.00	µg/m3
Bias at LV	-0.9546	µg/m3	Bias at LV	0.0182	µg/m3
Combined uncertainty	2.8372	µg/m3	Combined uncertainty	2.8536	µg/m3
Expanded relative uncertainty	18.9144%	pass	Expanded relative uncertainty	19.0242%	pass
Ref sampler uncertainty	0.4200	µg/m3	Ref sampler uncertainty	0.4200	µg/m3
Limit value	30	µg/m3	Limit value	30	µg/m3

Table 40. Regression analysis Fidas 200 PM_{2.5} candidate Fidas2, first and second comparison, RSA, RSW, UBW, UBS, RSW2.

RAW DATA			RESULTS AFTER CALIBRATING		
Regression	1.016y + 0.674		N (Spring)	54	n
Regression (i=0)	1.07y		N (Summer)	24	n
N	230	n	N (Fall)	73	n
			N (Winter)	79	n
Outliers	12	n	Outliers	8	n
Outliers	5.2	%	Outliers	3.5	%
Mean CM	8.89	µg/m3	Mean CM	9.70	µg/m3
Mean RM	9.70	µg/m3	Mean RM	9.70	µg/m3
Number of RM > UAT	26	n	Number of CM > UAT	25	n
Number of RM > LV	1	n	Number of CM > LV	3	n
REGRESSION RESULTS (RAW)			REGRESSION RESULTS (CALIBRATED)		
Slope b	0.9841		Slope b	1.0016	
Uncertainty of b	0.0271		Uncertainty of b	0.0275	
Intercept a	-0.6635	significant	Intercept a	-0.0154	
Uncertainty of a	0.3118		Uncertainty of a	0.3168	
r ²	0.828		r ²	0.828	
Slope b forced through origin	0.935	significant			
Uncertainty of b (forced)	0.0145				
EQUIVALENCE TEST (RAW)			EQUIVALENCE TEST (CALIBRATED)		
Uncertainty of calibration	0.870	µg/m3	Calibration	(y+0.664) / 0.984	
Uncertainty of calibration (forced)	0.436	µg/m3	Uncertainty of calibration	0.870	µg/m3
Random term	2.5676	µg/m3	Random term	2.7533	µg/m3
Additional uncertainty (optional)	0.00	µg/m3	Additional uncertainty (optional)	0.00	µg/m3
Bias at LV	-1.1398	µg/m3	Bias at LV	0.0322	µg/m3
Combined uncertainty	2.8092	µg/m3	Combined uncertainty	2.7535	µg/m3
Expanded relative uncertainty	18.7281%	pass	Expanded relative uncertainty	18.3567%	pass
Ref sampler uncertainty	0.4200	µg/m3	Ref sampler uncertainty	0.4200	µg/m3
Limit value	30	µg/m3	Limit value	30	µg/m3

Table 41. Performance characteristics of the Fidas 200 PM_{2.5} candidates Fidas1 and Fidas2, first comparison only, RSA, RSW, UBW, UBS. Comments in red indicate significant deviations from performance criteria. Comments in blue indicate that calibration is necessary.

Test PM _{2.5}	Criteria	Fidas1	Fidas2	Comment
Between CM uncertainty				
All data	$u_{bs,CM} < 2.5 \mu\text{g}/\text{m}^3$	0.12 $\mu\text{g}/\text{m}^3$		Pass
> 18 $\mu\text{g}/\text{m}^3$	$u_{bs,CM} < 2.5 \mu\text{g}/\text{m}^3$	0.20 $\mu\text{g}/\text{m}^3$		Pass
Comparability				
All data				
Number of values		170	185	
Data capture	$\geq 90 \%$	89.5 %	97.4 %	Fail / Pass
Average	Ref. 9.47 9.10	8.28	7.89	
Slope, b	$0.98 < b < 1.02$	0.8997	0.8908	Calibrate
Intercept, a	$-1 < a < 1$	-0.2459	-0.2127	
Expanded relative uncertainty	25 %	27.1 %	27.8 %	Fail
Calibrated data, RM = a + b * CM				
All data				
Slope, b		1.112	1.123	
Intercept, a		0.273	0.239	
Expanded relative uncertainty	25 %	19.2 %	18.3 %	Pass
All data, use intercept only (b = 1)				
Expanded relative uncertainty	25 %	25.9 %	26.7 %	Fail
All data, use slope only (a = 0)				
Expanded relative uncertainty	25 %	19.1 %	18.1 %	Pass
All data, PM _{2.5} > 18 $\mu\text{g}/\text{m}^3$				
Number of values	≥ 40	20	20	Fail
Slope, b		0.511	0.519	
Intercept, a		13.038	12.956	
Expanded rel. uncertainty	25 %	68.6 %	69.0 %	Fail

Table 42. Regression analysis Fidas 200 PM_{2.5} candidate Fidas1, first comparison only, RSA, RSW, UBW, UBS.

RAW DATA			RESULTS AFTER CALIBRATING		
Regression	1.112y + 0.273		N (Spring)	33	n
Regression (i=0)	1.134y		N (Summer)	14	n
N	170	n	N (Fall)	68	n
			N (Winter)	55	n
Outliers	8	n	Outliers	9	n
Outliers	4.7	%	Outliers	5.3	%
Mean CM	8.28	µg/m3	Mean CM	9.47	µg/m3
Mean RM	9.47	µg/m3	Mean RM	9.47	µg/m3
Number of RM > UAT	20	n	Number of CM > UAT	14	n
Number of RM > LV	1	n	Number of CM > LV	4	n
REGRESSION RESULTS (RAW)			REGRESSION RESULTS (CALIBRATED)		
Slope b	0.8997	significant	Slope b	1.0104	
Uncertainty of b	0.0290		Uncertainty of b	0.0322	
Intercept a	-0.2459		Intercept a	-0.0987	
Uncertainty of a	0.3314		Uncertainty of a	0.3684	
r ²	0.829		r ²	0.829	
Slope b forced through origin	0.881	significant			
Uncertainty of b (forced)	0.0163				
EQUIVALENCE TEST (RAW)			EQUIVALENCE TEST (CALIBRATED)		
Uncertainty of calibration	0.930	µg/m3	Calibration	(y+0.246) / 0.900	
Uncertainty of calibration (forced)	0.488	µg/m3	Uncertainty of calibration	0.930	µg/m3
Random term	2.4302	µg/m3	Random term	2.8772	µg/m3
Additional uncertainty (optional)	0.00	µg/m3	Additional uncertainty (optional)	0.00	µg/m3
Bias at LV	-3.2558	µg/m3	Bias at LV	0.2138	µg/m3
Combined uncertainty	4.0628	µg/m3	Combined uncertainty	2.8851	µg/m3
Expanded relative uncertainty	27.0851%	fail	Expanded relative uncertainty	19.2343%	pass
Ref sampler uncertainty	0.4200	µg/m3	Ref sampler uncertainty	0.4200	µg/m3
Limit value	30	µg/m3	Limit value	30	µg/m3

Table 43. Regression analysis Fidas 200 PM_{2.5} candidate Fidas2, first comparison only, RSA, RSW, UBW, UBS.

RAW DATA			RESULTS AFTER CALIBRATING		
Regression	1.123y + 0.239		N (Spring)	33	n
Regression (i=0)	1.143y		N (Summer)	24	n
N	185	n	N (Fall)	73	n
			N (Winter)	55	n
Outliers	7	n	Outliers	10	n
Outliers	3.8	%	Outliers	5.4	%
Mean CM	7.89	µg/m3	Mean CM	9.10	µg/m3
Mean RM	9.10	µg/m3	Mean RM	9.10	µg/m3
Number of RM > UAT	20	n	Number of CM > UAT	14	n
Number of RM > LV	1	n	Number of CM > LV	4	n
REGRESSION RESULTS (RAW)			REGRESSION RESULTS (CALIBRATED)		
Slope b	0.8908	significant	Slope b	1.0106	
Uncertainty of b	0.0266		Uncertainty of b	0.0299	
Intercept a	-0.2127		Intercept a	-0.0963	
Uncertainty of a	0.2950		Uncertainty of a	0.3312	
r ²	0.840		r ²	0.840	
Slope b forced through origin	0.875	significant			
Uncertainty of b (forced)	0.0152				
EQUIVALENCE TEST (RAW)			EQUIVALENCE TEST (CALIBRATED)		
Uncertainty of calibration	0.852	µg/m3	Calibration	(y+0.213) / 0.891	
Uncertainty of calibration (forced)	0.457	µg/m3	Uncertainty of calibration	0.852	µg/m3
Random term	2.2904	µg/m3	Random term	2.7296	µg/m3
Additional uncertainty (optional)	0.00	µg/m3	Additional uncertainty (optional)	0.00	µg/m3
Bias at LV	-3.4890	µg/m3	Bias at LV	0.2211	µg/m3
Combined uncertainty	4.1736	µg/m3	Combined uncertainty	2.7385	µg/m3
Expanded relative uncertainty	27.8241%	fail	Expanded relative uncertainty	18.2569%	pass
Ref sampler uncertainty	0.4200	µg/m3	Ref sampler uncertainty	0.4200	µg/m3
Limit value	30	µg/m3	Limit value	30	µg/m3

Table 44. Performance characteristics of the Grimm 180 PM_{2.5} candidates Grimm1 and Grimm2, first and second comparison, RSA, RSW, UBW, UBS, RSW2. Between CM uncertainty is based on first comparison only. Comments in red indicate significant deviations from performance criteria. Comments in blue indicate that calibration is necessary.

Test PM _{2.5}	Criteria	Grimm1	Grimm2	Comment
Between CM uncertainty				
All data > 18 µg/m ³	$U_{bs,CM} < 2.5 \mu\text{g}/\text{m}^3$ $U_{bs,CM} < 2.5 \mu\text{g}/\text{m}^3$	0.65 µg/m ³ 1.43 µg/m ³		Pass Pass
Comparability				
All data				
Number of values		230	223	
Data capture	≥ 90 %	97.9 %	94.9 %	Pass
Average	Ref. 9.64 9.77	10.33	10.87	
Slope, b	0.98 < b < 1.02	1.1077	1.1626	Calibrate
Intercept, a	-1 < a < 1	-0.3480	-0.4919	Pass
Expanded relative uncertainty	25 %	26.3 %	33.8 %	Fail
Calibrated data, RM = a + b * CM				
All data				
Slope, b		0.903	0.860	
Intercept, a		0.314	0.423	
Expanded relative uncertainty	25 %	17.2 %	15.5 %	Pass
All data, use intercept only (b = 1)				
Expanded relative uncertainty	25 %	28.1 %	36.7 %	Fail
All data, use slope only (a = 0)				
Expanded relative uncertainty	25 %	17.3 %	15.9 %	Pass
All data, PM _{2.5} > 18 µg/m ³				
Number of values	≥ 40	27	27	Fail
Slope, b		0.461	0.462	
Intercept, a		11.645	11.031	
Expanded rel. uncertainty	25 %	67.7 %	63.2 %	
Hjortneskaia RSA, autumn				
Number of values	≥ 40	42	42	Pass
Slope, b		1.148	1.027	
Intercept, a		0.824	0.857	
Expanded rel. uncertainty	25 %	18.4 %	19.3 %	Pass
Hjortneskaia RSW, winter				
Number of values	≥ 40	52	52	Pass
Slope, b		0.910	0.839	
Intercept, a		1.115	1.074	
Expanded rel. uncertainty	25 %	24.4 %	23.6 %	Pass
Sofienbergparken UBW, winter				
Number of values	≥ 40	49	42	Pass
Slope, b		0.812	0.840	
Intercept, a		-0.222	-0.449	
Expanded rel. uncertainty	25 %	8.8 %	9.2 %	Pass
Sofienbergparken UBS, summer				
Number of values	≥ 40	46	46	Pass
Slope, b		0.727	0.742	
Intercept, a		0.239	0.131	
Expanded rel. uncertainty	25 %	12.9 %	11.9 %	Pass
Smestad RSW2, winter				
Number of values	≥ 40	41	41	Pass
Slope, b		0.748	0.758	
Intercept, a		1.977	1.947	
Expanded rel. uncertainty	25 %	20.7 %	19.5 %	Pass

Table 45. Regression analysis Grimm 180 PM_{2.5} candidate Grimm1, first and second comparison, RSA, RSW, UBW, UBS, RSW2.

RAW DATA			RESULTS AFTER CALIBRATING		
Regression	0.903y + 0.314		N (Spring)	50	n
Regression (i=0)	0.924y		N (Summer)	27	n
N	230	n	N (Fall)	74	n
			N (Winter)	79	n
Outliers	11	n	Outliers	3	n
Outliers	4.8	%	Outliers	1.3	%
Mean CM	10.33	µg/m3	Mean CM	9.64	µg/m3
Mean RM	9.64	µg/m3	Mean RM	9.64	µg/m3
Number of RM > UAT	27	n	Number of CM > UAT	23	n
Number of RM > LV	1	n	Number of CM > LV	3	n
REGRESSION RESULTS (RAW)			REGRESSION RESULTS (CALIBRATED)		
Slope b	1.1077	significant	Slope b	0.9915	
Uncertainty of b	0.0280		Uncertainty of b	0.0253	
Intercept a	-0.3480		Intercept a	0.0818	
Uncertainty of a	0.3221		Uncertainty of a	0.2908	
r ²	0.852		r ²	0.852	
Slope b forced through origin	1.082	significant			
Uncertainty of b (forced)	0.0153				
EQUIVALENCE TEST (RAW)			EQUIVALENCE TEST (CALIBRATED)		
Uncertainty of calibration	0.900	µg/m3	Calibration	(y+0.348) / 1.108	
Uncertainty of calibration (forced)	0.459	µg/m3	Uncertainty of calibration	0.900	µg/m3
Random term	2.6892	µg/m3	Random term	2.5727	µg/m3
Additional uncertainty (optional)	0.00	µg/m3	Additional uncertainty (optional)	0.00	µg/m3
Bias at LV	2.8817	µg/m3	Bias at LV	-0.1726	µg/m3
Combined uncertainty	3.9416	µg/m3	Combined uncertainty	2.5785	µg/m3
Expanded relative uncertainty	26.2770%	fail	Expanded relative uncertainty	17.1900%	pass
Ref sampler uncertainty	0.4200	µg/m3	Ref sampler uncertainty	0.4200	µg/m3
Limit value	30	µg/m3	Limit value	30	µg/m3

Table 46. Regression analysis Grimm 180 PM_{2.5} candidate Grimm2, first and second comparison, RSA, RSW, UBW, UBS, RSW2.

RAW DATA			RESULTS AFTER CALIBRATING		
Regression	0.86y + 0.423		N (Spring)	50	n
Regression (i=0)	0.888y		N (Summer)	27	n
N	223	n	N (Fall)	74	n
			N (Winter)	72	n
Outliers	13	n	Outliers	4	n
Outliers	5.8	%	Outliers	1.8	%
Mean CM	10.87	µg/m3	Mean CM	9.77	µg/m3
Mean RM	9.77	µg/m3	Mean RM	9.77	µg/m3
Number of RM > UAT	27	n	Number of CM > UAT	22	n
Number of RM > LV	1	n	Number of CM > LV	4	n
REGRESSION RESULTS (RAW)			REGRESSION RESULTS (CALIBRATED)		
Slope b	1.1626	significant	Slope b	0.9903	
Uncertainty of b	0.0267		Uncertainty of b	0.0230	
Intercept a	-0.4919		Intercept a	0.0952	
Uncertainty of a	0.3109		Uncertainty of a	0.2674	
r ²	0.881		r ²	0.881	
Slope b forced through origin	1.127	significant			
Uncertainty of b (forced)	0.0145				
EQUIVALENCE TEST (RAW)			EQUIVALENCE TEST (CALIBRATED)		
Uncertainty of calibration	0.860	µg/m3	Calibration	(y+0.492) / 1.163	
Uncertainty of calibration (forced)	0.435	µg/m3	Uncertainty of calibration	0.860	µg/m3
Random term	2.5334	µg/m3	Random term	2.3219	µg/m3
Additional uncertainty (optional)	0.00	µg/m3	Additional uncertainty (optional)	0.00	µg/m3
Bias at LV	4.3871	µg/m3	Bias at LV	-0.1970	µg/m3
Combined uncertainty	5.0660	µg/m3	Combined uncertainty	2.3303	µg/m3
Expanded relative uncertainty	33.7735%	fail	Expanded relative uncertainty	15.5351%	pass
Ref sampler uncertainty	0.4200	µg/m3	Ref sampler uncertainty	0.4200	µg/m3
Limit value	30	µg/m3	Limit value	30	µg/m3

Table 47. Performance characteristics of the Grimm 180 PM_{2.5} candidates Grimm1 and Grimm2, first comparison only, RSA, RSW, UBW, UBS. Comments in red indicate significant deviations from performance criteria. Comments in blue indicate that calibration is necessary.

Test PM _{2.5}	Criteria	Grimm1	Grimm2	Comment
Between CM uncertainty				
All data	$U_{bs,CM} < 2.5 \mu\text{g}/\text{m}^3$	0.65 $\mu\text{g}/\text{m}^3$		Pass
> 18 $\mu\text{g}/\text{m}^3$	$U_{bs,CM} < 2.5 \mu\text{g}/\text{m}^3$	1.43 $\mu\text{g}/\text{m}^3$		Pass
Comparability				
All data				
Number of values		189	182	
Data capture	$\geq 90 \%$	99.5 %	95.8 %	Pass
Average	Ref. 9.08 9.21	9.61	10.24	
Slope, b	$0.98 < b < 1.02$	1.0647	1.1398	Calibrate
Intercept, a	$-1 < a < 1$	-0.0548	-0.2623	Pass
Expanded relative uncertainty	25 %	21.8 %	31.2 %	Pass/Fail
Calibrated data, $RM = a + b * CM$				
All data				
Slope, b		0.939	0.877	
Intercept, a		0.052	0.230	
Expanded relative uncertainty	25 %	17.9 %	16.1 %	Pass
All data, use intercept only (b = 1)				
Expanded relative uncertainty	25 %	22.2 %	32.8 %	Pass/Fail
All data, use slope only (a = 0)				
Expanded relative uncertainty	25 %	17.8 %	16.1 %	Pass
All data, PM _{2.5} > 18 $\mu\text{g}/\text{m}^3$				
Number of values	≥ 40	21	21	Fail
Slope, b		0.452	0.443	
Intercept, a		12.472	11.952	
Expanded rel. uncertainty	25 %	75.8 %	72.8 %	Fail

Table 48. Regression analysis Grimm 180 PM_{2.5} candidate Grimm1, first comparison only, RSA, RSW, UBW, UBS.

RAW DATA			RESULTS AFTER CALIBRATING		
Regression	0.939y + 0.052		N (Spring)	33	n
Regression (i=0)	0.943y		N (Summer)	27	n
N	189	n	N (Fall)	74	n
			N (Winter)	55	n
Outliers	11	n	Outliers	5	n
Outliers	5.8	%	Outliers	2.6	%
Mean CM	9.61	µg/m3	Mean CM	9.08	µg/m3
Mean RM	9.08	µg/m3	Mean RM	9.08	µg/m3
Number of RM > UAT	21	n	Number of CM > UAT	16	n
Number of RM > LV	1	n	Number of CM > LV	4	n
REGRESSION RESULTS (RAW)			REGRESSION RESULTS (CALIBRATED)		
Slope b	1.0647	significant	Slope b	0.9946	
Uncertainty of b	0.0304		Uncertainty of b	0.0285	
Intercept a	-0.0548		Intercept a	0.0493	
Uncertainty of a	0.3363		Uncertainty of a	0.3159	
r ²	0.846		r ²	0.846	
Slope b forced through origin	1.061	significant			
Uncertainty of b (forced)	0.0176				
EQUIVALENCE TEST (RAW)			EQUIVALENCE TEST (CALIBRATED)		
Uncertainty of calibration	0.971	µg/m3	Calibration	(y+0.055) / 1.065	
Uncertainty of calibration (forced)	0.528	µg/m3	Uncertainty of calibration	0.971	µg/m3
Random term	2.6796	µg/m3	Random term	2.6869	µg/m3
Additional uncertainty (optional)	0.00	µg/m3	Additional uncertainty (optional)	0.00	µg/m3
Bias at LV	1.8862	µg/m3	Bias at LV	-0.1136	µg/m3
Combined uncertainty	3.2769	µg/m3	Combined uncertainty	2.6893	µg/m3
Expanded relative uncertainty	21.8460%	pass	Expanded relative uncertainty	17.9286%	pass
Ref sampler uncertainty	0.4200	µg/m3	Ref sampler uncertainty	0.4200	µg/m3
Limit value	30	µg/m3	Limit value	30	µg/m3

Table 49. Regression analysis Grimm 180 PM_{2.5} candidate Grimm2, first comparison only, RSA, RSW, UBW, UBS.

RAW DATA			RESULTS AFTER CALIBRATING		
Regression	0.877y + 0.23		N (Spring)	33	n
Regression (i=0)	0.893y		N (Summer)	27	n
N	182	n	N (Fall)	74	n
			N (Winter)	48	n
Outliers	12	n	Outliers	5	n
Outliers	6.6	%	Outliers	2.7	%
Mean CM	10.24	µg/m3	Mean CM	9.21	µg/m3
Mean RM	9.21	µg/m3	Mean RM	9.21	µg/m3
Number of RM > UAT	21	n	Number of CM > UAT	15	n
Number of RM > LV	1	n	Number of CM > LV	4	n
REGRESSION RESULTS (RAW)			REGRESSION RESULTS (CALIBRATED)		
Slope b	1.1398	significant	Slope b	0.9914	
Uncertainty of b	0.0292		Uncertainty of b	0.0257	
Intercept a	-0.2623		Intercept a	0.0791	
Uncertainty of a	0.3285		Uncertainty of a	0.2883	
r ²	0.880		r ²	0.880	
Slope b forced through origin	1.120	significant			
Uncertainty of b (forced)	0.0168				
EQUIVALENCE TEST (RAW)			EQUIVALENCE TEST (CALIBRATED)		
Uncertainty of calibration	0.937	µg/m3	Calibration	(y+0.262) / 1.140	
Uncertainty of calibration (forced)	0.503	µg/m3	Uncertainty of calibration	0.937	µg/m3
Random term	2.5530	µg/m3	Random term	2.4099	µg/m3
Additional uncertainty (optional)	0.00	µg/m3	Additional uncertainty (optional)	0.00	µg/m3
Bias at LV	3.9307	µg/m3	Bias at LV	-0.1785	µg/m3
Combined uncertainty	4.6870	µg/m3	Combined uncertainty	2.4165	µg/m3
Expanded relative uncertainty	31.2468%	fail	Expanded relative uncertainty	16.1099%	pass
Ref sampler uncertainty	0.4200	µg/m3	Ref sampler uncertainty	0.4200	µg/m3
Limit value	30	µg/m3	Limit value	30	µg/m3

Table 50. Performance characteristics of the TEOM 1405 DF PM_{2.5} candidate TEOMDF, based on data from RSW (only a few samples), UBW and UBS. Comments in red indicate significant deviations from performance criteria. Comments in blue indicate that calibration is necessary.

Test PM _{2.5}	Criteria	TEOMDF	X	Comment
Between CM uncertainty				
All data > 18 µg/m ³	$u_{bs,CM} < 2.5 \mu\text{g}/\text{m}^3$ $u_{bs,CM} < 2.5 \mu\text{g}/\text{m}^3$			Only 1 candidate
Comparability				
All data				
Number of values		111		
Data capture	≥ 90 %	58.4 %		Fail
Average	Ref. 7.75	7.82		
Slope, b	$0.98 < b < 1.02$	0.9568		Calibrate
Intercept, a	$-1 < a < 1$	0.4113		Pass
Expanded relative uncertainty	25 %	10.6 %	%	Pass
Calibrated data, $RM = a + b * CM$				
All data				
Slope, b		1.045		
Intercept, a		-0.43		
Expanded relative uncertainty	25 %	10.2 %	%	Pass
All data, use intercept only (b = 1)				
Expanded relative uncertainty	25 %	12.4 %	%	Pass
All data, use slope only (a = 0)				
Expanded relative uncertainty	25 %	10.6 %	%	Pass
All data, PM _{2.5} > 18 µg/m ³				
Number of values	≥ 40	10		Fail
Slope, b		0.924		
Intercept, a		3.026		
Expanded rel. uncertainty	25 %	%	%	N too low
Hjortneskaia RSA, autumn				
Number of values	≥ 40			Not in operation
Slope, b				
Intercept, a				
Expanded rel. uncertainty	25 %	%	%	
Hjortneskaia RSW, winter				
Number of values	≥ 40	18		Fail
Slope, b		1.077		
Intercept, a		-0.286		
Expanded rel. uncertainty	25 %	13.8 %	%	Pass
Sofienbergparken UBW, winter				
Number of values	≥ 40	47		Pass
Slope, b		0.964		
Intercept, a		0.932		
Expanded rel. uncertainty	25 %	8.3 %	%	Pass
Sofienbergparken UBS, summer				
Number of values	≥ 40	46		Pass
Slope, b		0.812		
Intercept, a		-0.033		
Expanded rel. uncertainty	25 %	8.3 %	%	Pass

Table 51. Regression analysis TEOM 1405 DF PM_{2.5} candidate TEOMDF, first comparison, RSW, UBW, UBS.

RAW DATA			RESULTS AFTER CALIBRATING		
Regression	1.045y + -0.43		N (Spring)	33	n
Regression (I=0)	1.01y		N (Summer)	27	n
N	111	n	N (Fall)	0	n
			N (Winter)	51	n
Outliers	2	n	Outliers	1	n
Outliers	1.8	%	Outliers	0.9	%
Mean CM	7.82	µg/m3	Mean CM	7.75	µg/m3
Mean RM	7.75	µg/m3	Mean RM	7.75	µg/m3
Number of RM > UAT	10	n	Number of CM > UAT	8	n
Number of RM > LV	0	n	Number of CM > LV	1	n
REGRESSION RESULTS (RAW)			REGRESSION RESULTS (CALIBRATED)		
Slope b	0.9568	significant	Slope b	1.0012	
Uncertainty of b	0.0211		Uncertainty of b	0.0221	
Intercept a	0.4113		Intercept a	-0.0095	
Uncertainty of a	0.2093		Uncertainty of a	0.2188	
r ²	0.947		r ²	0.947	
Slope b forced through origin	0.990	significant			
Uncertainty of b (forced)	0.0136				
EQUIVALENCE TEST (RAW)			EQUIVALENCE TEST (CALIBRATED)		
Uncertainty of calibration	0.668	µg/m3	Calibration	(y-0.411) / 0.957	
Uncertainty of calibration (forced)	0.408	µg/m3	Uncertainty of calibration	0.668	µg/m3
Random term	1.3174	µg/m3	Random term	1.5363	µg/m3
Additional uncertainty (optional)	0.00	µg/m3	Additional uncertainty (optional)	0.00	µg/m3
Bias at LV	-0.8848	µg/m3	Bias at LV	0.0272	µg/m3
Combined uncertainty	1.5870	µg/m3	Combined uncertainty	1.5365	µg/m3
Expanded relative uncertainty	10.5797%	pass	Expanded relative uncertainty	10.2437%	pass
Ref sampler uncertainty	0.4200	µg/m3	Ref sampler uncertainty	0.4200	µg/m3
Limit value	30	µg/m3	Limit value	30	µg/m3

Table 52. Performance characteristics of the TEI FH62 IR PM_{2.5} candidates TE1 and TE2, first comparison, RSA, RSW, UBW, UBS. Candidate TE2 data was rejected due to unstable instrument. Comments in red indicate significant deviations from performance criteria. Comments in blue indicate that calibration is necessary.

Test PM _{2.5}	Criteria	TEI1	TEI2	Comment
Between CM uncertainty				
All data	$U_{bs,CM} < 2.5 \mu\text{g}/\text{m}^3$	4.44 $\mu\text{g}/\text{m}^3$		Fail
> 18 $\mu\text{g}/\text{m}^3$	$U_{bs,CM} < 2.5 \mu\text{g}/\text{m}^3$	6.50 $\mu\text{g}/\text{m}^3$		Fail
Comparability				
All data				
Number of values		180	171	
Data capture	$\geq 90 \%$	94.7 %		Pass
Average	Ref. 9.36	11.18		
Slope, b	$0.98 < b < 1.02$	1.1948		Calibrate
Intercept, a	$-1 < a < 1$	-0.0020		
Expanded relative uncertainty	25 %	46.2 %		Fail
Calibrated data, $RM = a + b * CM$				
All data				
Slope, b		0.837		
Intercept, a		0.002		
Expanded relative uncertainty	25 %	22.7 %		Pass
All data, use intercept only (b = 1)				
Expanded relative uncertainty	25 %	46.4 %	%	Fail
All data, use slope only (a = 0)				
Expanded relative uncertainty	25 %	22.4 %	%	Pass
All data, PM _{2.5} > 18 $\mu\text{g}/\text{m}^3$				
Number of values	≥ 40	21		Fail
Slope, b		0.446		
Intercept, a		12.615		
Expanded rel. uncertainty	25 %	81.8 %		Fail
Hjortneskaia RSA, autumn				
Number of values	≥ 40	42		Pass
Slope, b		0.538		
Intercept, a		1.161		
Expanded rel. uncertainty	25 %	30.6 %		Fail
Hjortneskaia RSW, winter				
Number of values	≥ 40	52		Pass
Slope, b		0.938		
Intercept, a		0.306		
Expanded rel. uncertainty	25 %	24.9 %		Pass
Sofienbergparken UBW, winter				
Number of values	≥ 40	49		Pass
Slope, b		1.118		
Intercept, a		-1.064		
Expanded rel. uncertainty	25 %	11.1 %		Pass
Sofienbergparken UBS, summer				
Number of values	≥ 40	37		Fail
Slope, b		0.771		
Intercept, a		0.139		
Expanded rel. uncertainty	25 %	12.6 %		Pass

Table 53. Regression analysis TEI FH62 IR PM_{2.5} candidate TEI1, first comparison, RSA, RSW, UBW, UBS.

RAW DATA			RESULTS AFTER CALIBRATING		
Regression	$0.837y + 0.002$		N (Spring)	33	n
Regression (i=0)	$0.837y$		N (Summer)	18	n
N	180	n	N (Fall)	74	n
			N (Winter)	55	n
Outliers	8	n	Outliers	4	n
Outliers	4.4	%	Outliers	2.2	%
Mean CM	11.18	µg/m3	Mean CM	9.36	µg/m3
Mean RM	9.36	µg/m3	Mean RM	9.36	µg/m3
Number of RM > UAT	21	n	Number of CM > UAT	18	n
Number of RM > LV	1	n	Number of CM > LV	1	n
REGRESSION RESULTS (RAW)			REGRESSION RESULTS (CALIBRATED)		
Slope b	1.1948	significant	Slope b	0.9756	
Uncertainty of b	0.0421		Uncertainty of b	0.0352	
Intercept a	-0.0020		Intercept a	0.2283	
Uncertainty of a	0.4765		Uncertainty of a	0.3988	
r ²	0.769		r ²	0.769	
Slope b forced through origin	1.195	significant			
Uncertainty of b (forced)	0.0244				
EQUIVALENCE TEST (RAW)			EQUIVALENCE TEST (CALIBRATED)		
Uncertainty of calibration	1.349	µg/m3	Calibration	$(y+0.002) / 1.195$	
Uncertainty of calibration (forced)	0.731	µg/m3	Uncertainty of calibration	1.349	µg/m3
Random term	3.7375	µg/m3	Random term	3.3606	µg/m3
Additional uncertainty (optional)	0.00	µg/m3	Additional uncertainty (optional)	0.00	µg/m3
Bias at LV	5.8417	µg/m3	Bias at LV	-0.5037	µg/m3
Combined uncertainty	6.9350	µg/m3	Combined uncertainty	3.3981	µg/m3
Expanded relative uncertainty	46.2334%	fail	Expanded relative uncertainty	22.6539%	pass
Ref sampler uncertainty	0.4200	µg/m3	Ref sampler uncertainty	0.4200	µg/m3
Limit value	30	µg/m3	Limit value	30	µg/m3

Table 54. Regression analysis TEI FH62 IR PM_{2.5} candidate TEI2, first comparison, RSA, RSW, UBW, UBS. All data was rejected.

RAW DATA			RESULTS AFTER CALIBRATING		
Regression	$0.6y + 1$		N (Spring)	33	n
Regression (i=0)	$0.652y$		N (Summer)	14	n
N	171	n	N (Fall)	71	n
			N (Winter)	53	n
Outliers	3	n	Outliers	0	n
Outliers	1.8	%	Outliers	0.0	%
Mean CM	13.93	µg/m3	Mean CM	9.36	µg/m3
Mean RM	9.36	µg/m3	Mean RM	9.36	µg/m3
Number of RM > UAT	19	n	Number of CM > UAT	16	n
Number of RM > LV	1	n	Number of CM > LV	0	n
REGRESSION RESULTS (RAW)			REGRESSION RESULTS (CALIBRATED)		
Slope b	1.6670	significant	Slope b	0.8754	significant
Uncertainty of b	0.0718		Uncertainty of b	0.0431	
Intercept a	-1.6669	significant	Intercept a	1.1661	significant
Uncertainty of a	0.8094		Uncertainty of a	0.4855	
r ²	0.614		r ²	0.614	
Slope b forced through origin	1.535	significant			
Uncertainty of b (forced)	0.0414				
EQUIVALENCE TEST (RAW)			EQUIVALENCE TEST (CALIBRATED)		
Uncertainty of calibration	2.301	µg/m3	Calibration	$(y+1.667) / 1.667$	
Uncertainty of calibration (forced)	1.243	µg/m3	Uncertainty of calibration	2.301	µg/m3
Random term	6.6523	µg/m3	Random term	4.3356	µg/m3
Additional uncertainty (optional)	0.00	µg/m3	Additional uncertainty (optional)	0.00	µg/m3
Bias at LV	18.3440	µg/m3	Bias at LV	-2.5725	µg/m3
Combined uncertainty	19.5130	µg/m3	Combined uncertainty	5.0413	µg/m3
Expanded relative uncertainty	130.0866%	fail	Expanded relative uncertainty	33.6090%	fail
Ref sampler uncertainty	0.4200	µg/m3	Ref sampler uncertainty	0.4200	µg/m3
Limit value	30	µg/m3	Limit value	30	µg/m3

Table 55. Performance characteristics of the TEOM 1400 AB PM_{2.5} candidates TEOM1 and TEOM2, first comparison, RSA, RSW, UBW, UBS. Comments in red indicate significant deviations from performance criteria. Comments in blue indicate that calibration is necessary.

Test PM _{2.5}	Criteria	TEOM1	TEOM2	Comment
Between CM uncertainty				
All data	$U_{bs,CM} < 2.5 \mu\text{g}/\text{m}^3$	0.91 $\mu\text{g}/\text{m}^3$		Pass
> 18 $\mu\text{g}/\text{m}^3$	$U_{bs,CM} < 2.5 \mu\text{g}/\text{m}^3$	0.45 $\mu\text{g}/\text{m}^3$		Pass
Comparability				
All data				
Number of values		179	183	
Data capture	$\geq 90 \%$	94.2 %	96.3 %	Pass
Average	Ref. 9.05 8.96	9.35	8.84	
Slope, b	$0.98 < b < 1.02$	0.6621	0.6891	Calibrate
Intercept, a	$-1 < a < 1$	3.3622	2.6607	Calibrate
Expanded relative uncertainty	25 %	46.2 %	45.1 %	Fail
Calibrated data, $RM = a + b * CM$				
All data				
Slope, b		1.51	1.451	
Intercept, a		-5.078	-3.861	
Expanded relative uncertainty	25 %	16.0 %	12.2 %	Pass
All data, use intercept only (b = 1)				
Expanded relative uncertainty	25 %	68.3 %	62.3 %	Fail
All data, use slope only (a = 0)				
Expanded relative uncertainty	25 %	40.5 %	30.3 %	Fail
All data, PM _{2.5} > 18 $\mu\text{g}/\text{m}^3$				
Number of values	≥ 40	20	20	Fail
Slope, b		1.082	1.040	
Intercept, a		3.452	4.365	
Expanded rel. uncertainty	25 %	37.9 %	33.7 %	Fail
Hjortneskaia RSA, autumn				
Number of values	≥ 40	39	40	Fail/Pass
Slope, b		1.656	1.420	
Intercept, a		-6.435	-3.723	
Expanded rel. uncertainty	25 %	14.5 %	10.1 %	Pass
Hjortneskaia RSW, winter				
Number of values	≥ 40	49	55	Pass
Slope, b		1.563	1.410	
Intercept, a		-6.695	-3.438	
Expanded rel. uncertainty	25 %	16.6 %	13.7 %	Pass
Sofienbergparken UBW, winter				
Number of values	≥ 40	49	43	Pass
Slope, b		1.909	1.686	
Intercept, a		-7.773	-4.765	
Expanded rel. uncertainty	25 %	22.3 %	18.9 %	Pass
Sofienbergparken UBS, summer				
Number of values	≥ 40	42	45	Pass
Slope, b		1.382	1.238	
Intercept, a		-3.809	-3.111	
Expanded rel. uncertainty	25 %	33.9 %	11.9 %	Fail/Pass

Table 56. Regression analysis TEOM 1400 AB PM_{2.5} candidate TEOM1, first comparison, RSA, RSW, UBW, UBS.

RAW DATA			RESULTS AFTER CALIBRATING		
Regression	1.51y + -5.078		N (Spring)	30	n
Regression (=0)	1.077y		N (Summer)	26	n
N	179	n	N (Fall)	68	n
			N (Winter)	55	n
Outliers	6	n	Outliers	22	n
Outliers	3.4	%	Outliers	12.3	%
Mean CM	9.35	µg/m3	Mean CM	9.05	µg/m3
Mean RM	9.05	µg/m3	Mean RM	9.05	µg/m3
Number of RM > UAT	20	n	Number of CM > UAT	18	n
Number of RM > LV	1	n	Number of CM > LV	3	n
REGRESSION RESULTS (RAW)			REGRESSION RESULTS (CALIBRATED)		
Slope b	0.6621	significant	Slope b	1.0259	
Uncertainty of b	0.0176		Uncertainty of b	0.0265	
Intercept a	3.3622	significant	Intercept a	-0.2344	
Uncertainty of a	0.1945		Uncertainty of a	0.2937	
r ²	0.881		r ²	0.881	
Slope b forced through origin	0.929	significant			
Uncertainty of b (forced)	0.0173				
EQUIVALENCE TEST (RAW)			EQUIVALENCE TEST (CALIBRATED)		
Uncertainty of calibration	0.562	µg/m3	Calibration	(y-3.362) / 0.662	
Uncertainty of calibration (forced)	0.518	µg/m3	Uncertainty of calibration	0.562	µg/m3
Random term	1.4487	µg/m3	Random term	2.3320	µg/m3
Additional uncertainty (optional)	0.00	µg/m3	Additional uncertainty (optional)	0.00	µg/m3
Bias at LV	-6.7736	µg/m3	Bias at LV	0.5425	µg/m3
Combined uncertainty	6.9268	µg/m3	Combined uncertainty	2.3942	µg/m3
Expanded relative uncertainty	46.1788%	fail	Expanded relative uncertainty	15.9615%	pass
Ref sampler uncertainty	0.4200	µg/m3	Ref sampler uncertainty	0.4200	µg/m3
Limit value	30	µg/m3	Limit value	30	µg/m3

Table 57. Regression analysis TEOM 1400 AB PM_{2.5} candidate TEOM2, first comparison, RSA, RSW, UBW, UBS.

RAW DATA			RESULTS AFTER CALIBRATING		
Regression	1.451y + -3.861		N (Spring)	33	n
Regression (=0)	1.113y		N (Summer)	26	n
N	183	n	N (Fall)	69	n
			N (Winter)	55	n
Outliers	8	n	Outliers	18	n
Outliers	4.4	%	Outliers	9.8	%
Mean CM	8.84	µg/m3	Mean CM	8.96	µg/m3
Mean RM	8.96	µg/m3	Mean RM	8.96	µg/m3
Number of RM > UAT	20	n	Number of CM > UAT	16	n
Number of RM > LV	1	n	Number of CM > LV	3	n
REGRESSION RESULTS (RAW)			REGRESSION RESULTS (CALIBRATED)		
Slope b	0.6891	significant	Slope b	1.0144	
Uncertainty of b	0.0143		Uncertainty of b	0.0207	
Intercept a	2.6607	significant	Intercept a	-0.1294	
Uncertainty of a	0.1570		Uncertainty of a	0.2278	
r ²	0.924		r ²	0.924	
Slope b forced through origin	0.898	significant			
Uncertainty of b (forced)	0.0137				
EQUIVALENCE TEST (RAW)			EQUIVALENCE TEST (CALIBRATED)		
Uncertainty of calibration	0.457	µg/m3	Calibration	(y-2.661) / 0.689	
Uncertainty of calibration (forced)	0.412	µg/m3	Uncertainty of calibration	0.457	µg/m3
Random term	1.1582	µg/m3	Random term	1.8075	µg/m3
Additional uncertainty (optional)	0.00	µg/m3	Additional uncertainty (optional)	0.00	µg/m3
Bias at LV	-6.6664	µg/m3	Bias at LV	0.3038	µg/m3
Combined uncertainty	6.7662	µg/m3	Combined uncertainty	1.8328	µg/m3
Expanded relative uncertainty	45.1083%	fail	Expanded relative uncertainty	12.2187%	pass
Ref sampler uncertainty	0.4200	µg/m3	Ref sampler uncertainty	0.4200	µg/m3
Limit value	30	µg/m3	Limit value	30	µg/m3

Appendix C

Geographic distribution of analyser types over Norway




Figure 50. 23 TEOM 1400 units are located at 14 stations distributed over most of Norway.




Figure 51. 11 TEOM DF units are located at 11 stations distributed in Trondheim, Innlandet and Greater Oslo.




Figure 52. 16 Grimm EDM 180 units are located at 16 stations distributed over Southern Norway, and one in the North.




Figure 53. 15 Palas Fidas 200 units are located at 15 stations distributed over most of Norway.




Figure 54. 2 TEI FH 62 I-R units are located at 1 station in Oslo.

NILU – Norwegian Institute for Air Research

NILU – Norwegian Institute for Air Research is an independent, nonprofit institution established in 1969. Through its research NILU increases the understanding of climate change, of the composition of the atmosphere, of air quality and of hazardous substances. Based on its research, NILU markets integrated services and products within analyzing, monitoring and consulting. NILU is concerned with increasing public awareness about climate change and environmental pollution.

NILU's values: Integrity - Competence - Benefit to society

NILU's vision: Research for a clean atmosphere

NILU – Norwegian Institute for Air Research
P.O. Box 100, NO-2027 KJELLER, Norway

E-mail: nilu@nilu.no

<http://www.nilu.no>

ISBN: 978-82-425-3059-2
ISSN: 2464-3327